

Robust Individual Circadian Parameter Estimation for Biosignal-based Personalisation of Cancer Chronotherapy

B. Schott, J. Stegmaier, A. Arbaud, M. Reischl, R. Mikut and F. Lévi

Abstract—In cancer treatment, chemotherapy is administered according to a constant schedule. The chronotherapy approach, considering chronobiological drug delivery, adapts the chemotherapy profile to the circadian rhythms of the human organism. This reduces toxicity effects and at the same time enhances efficiency of chemotherapy. To personalize cancer treatment, chemotherapy profiles have to be further adapted to individual patients. Therefore, we present a new model to represent cycle phenomena in circadian rhythms. The model enables a more precise modelling of the underlying circadian rhythms. In comparison with the standard model, our model delivers better results in all defined quality indices. The new model can be used to adapt the chemotherapy profile efficiently to individual patients. The adaption to individual patients contributes to the aim of personalizing cancer therapy.

Index Terms—cancer, chronobiology, circadian rhythm

I. INTRODUCTION

Cancer is a major public health problem all over the world. It is the leading cause of death in economically developed countries and the second leading cause in developing countries. There is also a big economic impact of cancer through a huge amount of financial costs for both the person suffering from cancer and the society. The general aim is to increase the overall survival rate. Therefore, the treatment efficiency for patients suffering from cancer has to be improved [10].

For this purpose circadian drug delivery in chemotherapy is used. Here, the tolerability and efficiency of anticancer drugs are improved on the basis of the circadian timing system in humans [13]. Today, often the standard chemotherapy as well as the radiation dose are chosen individually according to the tumour type. The new approach, considering the circadian drug delivery, further adapts the chemotherapy profile to circadian rhythms (chronomodulation) of the human organism. This reduces toxicity effects of the chemotherapeutic drugs and at the same time enhances the efficiency. To access circadian rhythms [19] in human, previous work dealt with the requirements to measure a marker rhythm which characterizes the timing of

the internal circadian system [11]. Such a marker rhythm has to be periodic and easy to measure over a long period of time with non-invasive methods.

Up to now, the state-of-the-art was a fixed chronomodulated chemotherapy profile for all patients [16]. This fixed profile was extracted from experimental models of the rest-activity rhythm [7, 13]. To personalize profiles, information out of biosignals (e.g., rest-activity, temperature) was required to access the patient-specific circadian rhythms. Such biosignals containing cycle phenomena are often modelled using the Cosinor model [6, 17]. To model the rhythms, a cosine wave is fitted to the data using optimization algorithms. In general, the parameters of a cosine wave are used to describe the underlying rhythm [21]. However, the Cosinor model is not able to model different lengths of intervals within one period of a cycle phenomenon [19].

Therefore, the estimation of such rhythms containing different lengths of intervals within one period of a cycle phenomenon is not adequate using the Cosinor analysis [1]. This results in an increasing need for novel models considering the different lengths of intervals [2, 4]. This leads to a better description of individual differences in individual patients to further optimize the benefit of chronomodulated chemotherapy [7, 13]. The analysis of such models and the extraction of relevant information to characterize the patients is an important topic in research [5, 9, 12, 14].

To achieve this aim we developed a new method called Extended Cosinor model [20]. This model is able to represent cycle phenomena containing different lengths of intervals within one period. Therefore, a new segment-wise defined function is developed consisting of two cosine waves. With the different circular frequencies of the two cosine waves it is possible to model the different lengths of intervals. Also through the segment-wise definition of the model function it is possible to model the superior periodicity of the underlying rhythm.

This more detailed model of cycle phenomena in biosignals can be used to adapt the chronomodulated chemotherapy efficiently to individual patients. This contributes to the aim of personalizing cancer therapy.

II. METHODS

A. Extended Cosinor model

For the representation of the anti-periodicity within rhythms the Extended Cosinor model was developed. The model is able

B.S. is with the Institute for Applied Computer Science, Karlsruhe Institute of Technology, Karlsruhe, Germany e-mail: benjamin.schott@kit.edu.

J.S. and M.R. are with the Institute for Applied Computer Science, Karlsruhe Institute of Technology, Germany

A.A. was with the Institut National de la Santé et de la Recherche Médicale, Paris

R.M. is Professor at the Institute for Applied Computer Science, Karlsruhe Institute of Technology, Germany

F.L. is Professor at the Warwick Medical School, Warwick Systems Biology Center, England and was former Professor at the Institut National de la Santé et de la Recherche Médicale, Paris

to consider different lengths of two intervals (e.g. day and night, rest and activity) within one period. At the same time the model is able to consider the periodicity of the superior rhythm (e.g. 24-hour cycle). As a basis for the Extended Cosinor model, a new segment-wise defined model function is developed ($f_{\cos, \text{Extended}}$). This model function consists of two cosine half-waves. The circular frequencies of the two cosine waves are used to model the different lengths of the intervals within one period. See Figure 1 for an example of the newly developed model function, where two cosine waves were combined to a new function.




Fig. 1. Model function of the Extended Cosinor model: The two shifted cosine half-waves are shown in red ($f_{\cos1}$) and green ($f_{\cos2}$). Due to the fact that φ_1 is equal to zero in this case, the intersection is at π/ω_1

B. Using the Extended Cosinor as a model for rhythms

For modelling rhythms with the Extended Cosinor model, the model function $f_{\cos, \text{Extended}}$ has to be fitted to the original data extracted from the measured biosignals. For this purpose, a numerical optimization algorithm was used to estimate the parameter vector of the model function. Therefore, a quality index describing the goodness of the model was used to optimize these parameters. Furthermore, constraints are defined within the optimization algorithm. These constraints force the parameters of the model function to be set in a predefined interval. These intervals influence the shape of the function following characteristic patterns. Therefore, the constraints within the optimization algorithm are required for a better representation of the patterns within a rhythm (see Figure 2). All models and algorithms used were implemented using the Matlab toolbox Gait-CAD [15].

III. RESULTS

For the evaluation of our model we used two databases (Database A and Database B). Database A was created as part of a study in 2013 [8, 18]. Two biosignals, the temperature and the rest-activity signal, were recorded using wireless sensors (sampling rate of $\frac{1}{min}$). Both biosignals were monitored for 4




Fig. 2. Different methods to fit a model function to the biosignal (in this case a rest-activity rhythm). Here the actimetry in the zero crossing mode (ZC) is shown over the span of a whole day.

days in 10 patients suffering from cancer (for more details see [20]). Database B was created in connection with the PICADO project in 2012 [3, 8]. Here also the temperature and the rest-activity signal were recorded in 34 healthy people over a period of 4 days (sampling rate of $\frac{5}{min}$). For evaluation purposes we used the rest-activity signal in both databases as an example of a biosignal. The Cosinor model, as the standard model for representing rest-activity rhythms is compared with the new Extended Cosinor model using the quality indices listed in Table I.

TABLE I
QUALITY INDICES FOR EVALUATION

Quality index	Description	Abbreviation	Values for a good model
Q_1	Sum squared errors	SSE	Lower
Q_2	R^2	R squared	Higher
Q_3	$adjustedR^2$	Adjusted R squared	Higher
Q_4	RMSE	Root mean squared error	Lower
Q_5	r_{xy}	Correlation coefficient	Higher

The results of modelling all rest-activity rhythms in Database A are listed in Table II. Here, the bold black values state the best results for each quality measure. As shown in the results the Extended Cosinor model is superior to the Cosinor models in all quality indices. An example of one rest-activity rhythm modelled with the Cosinor and Extended Cosinor model is shown in Figure 3.

TABLE II
EVALUATION OF DIFFERENT METHODS TO REPRESENT THE REST-ACTIVITY TIME SERIES IN DATABASE A. HERE BOLD VALUES REPRESENT BETTER RESULTS.

	Cosinor model	Extended Cosinor model
Q_1 Mean \pm STD	$1.25e+07 \pm 3.9e+06$	$1.2e+07 \pm 4e+06$
Q_2 Mean \pm STD	0.173 ± 0.122	0.201 ± 0.144
Q_3 Mean \pm STD	0.173 ± 0.122	0.198 ± 0.145
Q_4 Mean \pm STD	92.4 ± 13.8	90.6 ± 13.7
Q_5 Mean \pm STD	0.429 ± 0.172	0.445 ± 0.164

For Database B, Table III lists the results of modelling the rest-activity rhythm. Here, also the Extended Cosinor model was superior to the Cosinor model regarding all quality indices. Moreover, Figure 4 shows one example for modelling a rest-activity rhythm in Database B.

TABLE III

EVALUATION OF DIFFERENT METHODS FOR THE REST-ACTIVITY TIME SERIES IN DATABASE B. HERE BOLD VALUES REPRESENT BETTER RESULTS.

	Cosinor model	Extended Cosinor model
Q_1 Mean \pm STD	$2.82e+06 \pm 2.92e+06$	$2.67e+06 \pm 2.65e+06$
Q_2 Mean \pm STD	0.339 ± 0.136	0.369 ± 0.135
Q_3 Mean \pm STD	0.338 ± 0.136	0.367 ± 0.135
Q_4 Mean \pm STD	37.6 ± 23.7	36.7 ± 23
Q_5 Mean \pm STD	0.573 ± 0.124	0.599 ± 0.115

The developed Extended Cosinor model provides better representation of rhythms containing intervals with different lengths in comparison to the Cosinor model. This advantage is based on the fact that the Extended Cosinor model in contrast to the Cosinor model is able to model different lengths of intervals within one cycle period. Furthermore, the underlying model function of the Extended Cosinor model provides more parameters in comparison with the Cosinor model. The higher number of parameters can be used for further accessing of the rhythms.




Fig. 3. Example of the Cosinor and Extended Cosinor model for one rest-activity rhythm in Database A. Here it can be clearly observed, that the Extended Cosinor model (blue dashed line) models the data better in comparison with the Cosinor mode (green solid line).

IV. CONCLUSION

In this paper a new model (Extended Cosinor) representing cycle phenomena in rhythms is presented. The new model is evaluated against the standard Cosinor model using rest-activity rhythms as one exemplary biosignal. The Extended Cosinor model provides a better representation of rhythms containing intervals with different lengths within one cycle period.

The new model can be used to access biorhythms of individual patients. In the context of chronomodulated chemotherapy, this can serve as a basis for further personalization of




Fig. 4. Example of the Cosinor and Extended Cosinor model for one rest-activity rhythm in Database B. Here it can be observed, that the Extended Cosinor model (blue dashed line) models the data better in comparison with the Cosinor mode (green solid line).

chemotherapy according to the biorhythms of the patients. However, there are some limitations so far. If the biorhythms are too disrupted, the quality of modeling is not as good as modelling normal rhythms. Therefore, future research has to handle the problem of adapting the constraints for model generation to cope with such phenomena and further adapt these constraints to every individual rhythm. It is suggested that adapting the constraints improves the model quality for individual patients and therefore makes a further steps towards the personalization of chemotherapy.

ACKNOWLEDGMENT

The authors would like to thank the Institut National de la Santé et de la Recherche Médicale (INSERM) for providing the datasets.

REFERENCES

- [1] A. Ballesta, J. Clairambault, S. Dulong, and F. Lévi. A Systems Biomedicine Approach for Chronotherapeutics Optimization: Focus on the Anticancer Drug Irinotecan. In A. Onofrio, P. Cerrai, and A. Gandolfi, editors, *New Challenges for Cancer Systems Biomedicine*, SIMAI Springer Series, 301–327. Springer Milan, 2012.
- [2] F. Billy, J. Clairambault, O. Fercoq, T. Lorenzi, A. Lorz, and B. Perthame. Modelling Targets for Anticancer Drug Control Optimisation in Physiologically Structured Cell Population Models. In *ICNAAM 2012 - 10th International Conference of Numerical Analysis and Applied Mathematics*, 1479: 1323–1326, Kos, Grèce, Sept. 2012. Simos, Thodoros, American Institute of Physics.
- [3] A. Cis. Développement d'une Infrastructure Technologique Innovante et Polyvalente pour la Domomédecine. <http://www.utt.fr/fr/toute-l-actualite/domomedecine.html>.

- [4] J. Clairambault, S. Gaubert, and T. Lepoutre. Circadian Rhythm and Cell Population Growth. *Mathematical and Computer Modelling*, 53: 1558 – 1567, 2011.
- [5] F. Gachon and D. Firsov. The Role of Circadian Timing System on Drug Metabolism and Detoxification. *Expert Opinion on Drug Metabolism & toxicology*, 7(2):147–158, 2011.
- [6] P. Innominato, M. Mormont, T. Rich, J. Waterhouse, F. Lévi, and G. Bjarnason. Circadian Disruption, Fatigue, and Anorexia Clustering in Advanced Cancer Patients: Implications for Innovative Therapeutic Approaches. *Integrative Cancer Therapies*, 8(4):361–370, 2009.
- [7] P. Innominato, F. Lévi, and G. Bjarnason. Chronotherapy and the Molecular Clock: Clinical Implications in Oncology. *Advanced Drug Delivery Reviews*, 62:979 – 1001, 2010.
- [8] INSERM. Intérêt de l’actimétrie dans le Suivi des Maladies Chroniques. 2013.
- [9] F. Lévi. Chronotherapeutics: The Relevance of Timing in Cancer Therapy. *Cancer Causes & Control*, 17(4):611–621, 2006.
- [10] F. Lévi, E. Filipinski, V. King, X. Li, T. Granda, M. Mormont, X. Liu, B. Claustrat, and H. Hastings. Host Circadian Clock as a Control Point in Tumor Progression. *Journal of the National Cancer Institute*, 94(9):690–697, 2002.
- [11] F. Lévi, C. Focan, A. Karaboue, V. Valette, D. Focan-Henrard, B. Baron, F. Kreutz, and S. Giacchetti. Implications of Circadian Clocks for the Rhythmic Delivery of Cancer Therapeutics. *Advanced Drug Delivery Reviews*, 59:1015 – 1035, 2007.
- [12] F. Lévi and A. Okyar. Circadian Clocks and Drug Delivery Systems: Impact and Opportunities in Chronotherapeutics. *Expert Opinion on Drug Delivery*, 8(12):1535–1541, 2011.
- [13] F. Lévi, A. Okyar, S. Dulong, P. Innominato, and C. J. Circadian Timing in Cancer Treatments. *Annu. Rev. Pharmacol. Toxicol.*, 50:377–421, 2010.
- [14] R. Mikut. *Data Mining in der Medizin und Medizintechnik*, Universitätsverlag Karlsruhe, 2008.
- [15] R. Mikut, O. Burmeister, and M. Reischl. The Open Source Matlab Toolbox Gait-CAD and its Application to Bioelectric Signal Processing. In *Proc., DGBMT-Workshop Biosignalverarbeitung, Potsdam*, 2008.
- [16] M. Mormont and F. Lévi. Cancer Chronotherapy: Principles, Applications, and Perspectives. *Cancer*, 97(1):155–169, 2003.
- [17] M. Mormont, J. Waterhouse, P. Bleuzen, S. Giacchetti, A. Jami, A. Bogdan, J. Lellouch, J. Misset, Y. Touitou, and F. Levi. Marked 24-h Rest-activity Rhythms are Associated with Better Quality of Life, Better Response, and Longer Survival in Patients with Metastatic Colorectal Cancer and Good Performance Status. *Clinical Cancer Research*, 6:3038, 2000.
- [18] V. Roche, A. Mohamad-Djafari, P. Innominato, A. Karaboue, and L. F. Gorbach, A. Thoracic Surface Temperature Rhythms as Circadian Biomarkers for Cancer Chronotherapy. *Chronobiology International*, 31(3):409–420, 2014.
- [19] J. Sarabia, M. Rol, P. Mendiola, and J. Madrid. Circadian Rhythm of Wrist Temperature in Normal-living Subjects: A Candidate of new Index of the Circadian System. *Physiology Behavior*, 95(4):570 – 580, 2008.
- [20] B. Schott. Development of an Automated Framework for the Biosignal-based Personalisation of Cancer Chronotherapy. Master’s thesis, Karlsruhe Institute of Technology, 2014.
- [21] C. Scully, A. Karaboue, J. Liu, WM. AMD Meyer, P. Innominato, K. Chon, A. Gorbach, and F. Lévi. Skin Surface Temperature Rhythms as Potential Circadian Biomarkers for Personalized Chronotherapeutics in Cancer Patients. *Interface Focus*, 1(1):48–60, 2011.