Application of single walled carbon nanotubes for heating agent in photothermal therapy

Syahril Siregar^{a,*}, Israr Ul Haq^a, Ryo Nagaoka^a, Yoshifumi Saijo^a

^a Graduate School of Biomedical Engineering, Tohoku University, Sendai 980-8579, Japan

Abstract

We present the theoretical investigation of the single walled carbon nanotubes (SWNTs) as the heating agent of photothermal therapy. In our model, the SWNT is modeled by rigid tube surrounded by cancer cells. In this model, we neglect the angle dependence of temperature and assume that the length of SWNT is much longer than the radius of tube. We calculated the temperature rise of the SWNT and its surrounding cancer cells during the laser heating by solving one-dimensional heat conduction equation in steady state condition. We found that the maximum temperature is located at the interface between SWNT and cancer cells. This maximum temperature is proportional to the square of SWNTs diameter and diameter of SWNTs depends on their chirality. These results extend our understanding of the temperature distribution in SWNT during the laser heating process and provide the suggested specification of SWNT for the improvement the photothermal therapy in the future. Keywords: Photothermal therapy, single-walled carbon nanotubes, cancer, laser heating, heat conduction equation 2016 MSC: 00-01, 99-00

1. Introduction

Over the past decade, the interaction between laser and nanoparticle has received outstanding attention, because of the promised potential application

Preprint submitted to Journal of Physics in Medicine and Biology

^{*}Corresponding author

Email address: siregar.syahril.p8@dc.tohoku.ac.jp (Syahril Siregar)

for the future cancer treatment [1, 2, 3, 4, 5]. The cancer treatment using this interaction is called the photothermal therapy (PTT).

PTT is recently investigated for cancer treatment, because of its safety as a consequence of using non-ionized radiation from laser. The PTT process is given as follows: (i) the heating agent is injected into the center of the cancer cell, (ii) the laser is irradiated directed to the cancer cells, (iii) the heating agent will absorb the laser energy, finally (iv) the temperature of heating agent and its surrounding will be significantly increased. The rise in the temperature of cancer cells causes the cancer cells to become hyperthermia and damaged. The cancer cells become damaged due to the destitute of blood supply, when its temperature reaches at least 41-47 0 C for several minutes [6, 7].

The heating agent is very important in the PTT, because irradiating the tumor cells without injecting the heating agent will not damage the cancer cells [8]. The heating agents commonly used in PTT are gold nanorod, and gold nanosphere [6, 9, 10, 11].

Even though, gold nanoparticles have been proved as potential heating agents of PTT from some previous works, gold nanoparticles have limitation for heating agents of PTT. The absorption spectra of gold nanoparticles lies in the first near infra red (NIR) window (600 - 900 nm), furthermore, the penetration of light into heating agents in the first NIR window is less efficient compared with the heating agents having absorption spectra in the second NIR window (1000 -1350 nm) [12, 13].

Carbon based materials, such as graphene, graphene oxide and SWNTs are potential candidate for heating agents material of PTT, since carbon based materials have very strong absorption in the first and second NIR windows [8, 14, 15, 16, 17].

There are some merits of using SNWTs as heating agents of PTT, such as the diameter of SWNTs is relatively smaller than diameter of most gold nanoparticles, thus we can use SWNTs for cancer cells which located in the specific place. Moreover, the thermal conductivity of SWNT is ten times higher than gold nanoparticles. Therefore, based on its physical properties, SWNTs is promised potential candidates for PTT heating agents.

Previous work about the experiment of PTT using SWNTs as heating agent, concluded that the SWNTs significantly enhanced the destruction of tumor cells without destructing the normal cells [15].

The theoretical simulations of PTT using laser and carbon nanotubes have been carried out from previous work. Feng Gong *et al.* investigated the rise in temperature of SWNTs during the laser heating process by using Monte Carlo approach [18]. They made a cluster model, with cancer cells located in the center and SWNTs distributed in the surroundings of cancer cells, then investigated the effect of SWNTs orientation distribution relative to the laser beam, such as random, perpendicular, and parallel distribution [18].

Toshiyuki Nakamiya *et al.* investigated the thermal analysis of multi-walled carbon nanotubes (MWCNTs) during pulsed laser heating by using finite element method [19].

In the present work, we made very simple model to calculate the rise in temperature of the heating agents and cancer cells by solving one-dimensional heat conduction equation within second order differential equation. Based on our model, we can provide the suggested characteristic of SWNT to improve the PTT for future cancer treatment.

2. Properties of SWNT

The electronic and optical properties of SWNTs are depends on the their chirality. Chirality is described by chiral vector (n,m) which shows how the SWNT formed from the graphene sheet. In Fig. 1 (a), the example of (8,4) SWNT formed from the graphene sheet is shown.

The electronic properties of SWNTs is classified into two groups, metallic and semiconductor. SWNTs can be classified as metallic if n - m = 3i and semiconductor if $n - m \neq 3i$, with *i* is integer [20, 21]. Even so, in this work, we do not consider the electronics properties of SWNTs, based on the chirality we only take the information about diameter by using the function as shown in



Figure 1: (Color online) (a) Geometry of (8,4) SWNT formed from graphene sheet with the graphene unit vectors a_1 and a_2 . If the O is joined to P, and Q is joined to Q', the tube can be constructed. (b) UV-vis-NIR spectrum of several chiralities of SWNT. There are two strong peaks with the stronger peak appear in the 900-1300 nm regions, data is adapted from reference [22].

Tab. 1.

In case of PTT, the information about the UV-vis-NIR spectra of heating agents is necessary, because it demonstrates suggested wavelength of laser. In Fig. 1 (b) we showed the UV-vis-NIR spectra adapted from reference [22]. Generally, two peaks exist in UV-vis-NIR spectra for each chirality, the stronger peak appear in the 900 nm - 1300 nm (around the second NIR window) and the weaker peak appear in the 500 nm - 800 nm (around the first NIR window).

Although the thermal conductivity is sensitive to the temperature, in the

Table 1: Physical parameters of tissue, SWNT and laser.		
Physical parameters		
Thermal conductivity of tissue	k_t	0.567 W/mK [24]
Thermal conductivity of SWNTs	k_c	3000 - 3500 W/mK [25]
Initial temperature	T_{∞}	$37^o \mathrm{C}$
Reflectivity	R	0.1
Absorption coefficient of SWNTs	α	$2.5 \times 10^9 \text{ m}^{-1}$
Laser intensity	I_0	$1 \times 10^{6} \text{ W/cm}^{2} \text{ [26]}$
Diameter of SWNT	$d_{\rm SWNT}$	$(n^2 + m^2 + nm)^{1/2} 0.0783 \text{ nm} [27]$
Radius of SWNT	a	$d_{\rm SWNT}/2$
The farthest considered distance	b	1000 a

present work, the thermal conductivity of SWNTs and cancer cells are assumed to be constant to the temperature. The absorption coefficient of single molecules SWNTs was not measured nor calculated from previous work. Thus, we tune the absorption coefficient of single molecule SWNTs, so that the (6,5) SWNT gives the maximum temperature value around 57°C, the similar temperature from experiment in the previous work [15, 23], and we use that value to calculate the temperature rise of another chirality.

3. Model

In this work, SWNT is modeled by rigid tube with radius *a* surrounded by the cancer cells as shown in Fig. 2. In order to simplify the problem, we assume the length of tube is much longer than radius of tube, so we can neglect the dependence on the length of tube. We also neglect the angle dependence of temperature. Therefore, the temperature rise during the laser heating only depends on radial distance.



Figure 2: (Color online) Schematic model of SWNT and its surrounding (cancer cells) in PTT. The temperature in the outer cell with distance b from center is assumed to be temperature of body 37° C.

The PTT heating process is described by heat conduction equation given as

$$\rho_c c_c \frac{\partial T}{\partial t} = \frac{k_c}{r} \frac{\partial}{\partial r} \left(r \frac{\partial T}{\partial r} \right) + q, \text{ for } 0 \le r \le a, \tag{1}$$

$$\rho_t c_t \frac{\partial T}{\partial t} = \frac{k_t}{r} \frac{\partial}{\partial r} \left(r \frac{\partial T}{\partial r} \right), \quad \text{for } r > a, \tag{2}$$

where $\rho_{c,t}$ is the density of SWNT / cancer cells, $k_{c,t}$ is the thermal conductivity of SWNT/cancer cells, T is temperature, and q is the heat source which represents the laser heating process. r is radial distance measured from the center of the tube. The heat source q can be expressed as follow

$$q = (1 - R)I_0 \alpha \exp(-\alpha z), \tag{3}$$

where I_0 is the laser intensity, α is absorption coefficient of SWNT, R is the reflectivity, and z is the depth, measured from the interface to the center of the tube.

We consider only steady state condition to simplify the problem, since many of interactions might be happen, if we consider the unsteady state condition, such as phonon-phonon, electron-phonon, and electron-heating interactions. The steady state condition of the heat conduction equation in eqs. (1) and (2), which are the time independent equations given by,

$$k_c \frac{1}{r} \frac{d}{dr} \left(r \frac{dT}{dr} \right) + q = 0, \quad \text{for } 0 \le r \le a, \tag{4}$$

$$k_t \frac{d}{dr} \left(r \frac{dT}{dr} \right) = 0, \quad \text{for } r > a,$$
 (5)

with the boundary conditions

$$\frac{dT}{dr} = 0, \text{ at } r = 0, \tag{6}$$

$$T(a_{-}) = T(a_{+}), \tag{7}$$

$$k_c \left(\frac{dT}{dr}\right)_{a_-} = k_t \left(\frac{dT}{dr}\right)_{a_+},\tag{8}$$

$$T = T_{\infty} \text{ at } r \to \infty.$$
 (9)

The temperature in the center of the tube should be finite, as described in eq. (6). Eqs (7) and (8) explain that the temperature at the boundary should be continuous. The Farthest considered distance (1000 *a*) measured from the center of tube is assumed has the same temperature with the cells temperature 37 °C (T_{∞}) and does not affected by laser heating, as can be seen in eq. (9).

4. Analytical Results

By doing some mathematical operations with considering the boundary condition in eqs. (6)-(9), the solutions of eqs. (4) and (5) can be obtained. The temperature as a function radial position is given by,

$$T(r) = \begin{cases} \frac{q}{4k_c} \left(a^2 - r^2\right) + \frac{qa^2}{2k_t} \ln\left(\frac{b}{a}\right) + T_{\infty} & \text{for } 0 \le r \le a, \\ \frac{qa^2}{2k_t} \ln\left(\frac{b}{r}\right) + T_{\infty} & \text{for } r > a. \end{cases}$$
(10)

The q is appear even in the r > a, but the value of z in eq. (3) is zero for this region. The theoretical solution in eq. (10) shows that the temperature in the interface is proportional to the square of SWNT radius. This result is in contrast with previous work that concluded the smaller diameter have better heating efficiency [18]. In the present work, the larger diameter gives more depth for light to penetrate into the center of the tube, as a consequence, the tube will absorbs more energy from the light and convert it to the thermal energy. According to the plot in Fig. (3)(b), the maximum temperature exist in the interface between tube and the cancer cells.



Figure 3: (Color online) (a) The calculated temperature rise as a function of radial distance measured from the center of (6,5) SWNT, (b) temperature in the interface between the (6,5) SWNT and the cancer cells, (c) the calculated interface temperature as a function of SWNT diameter for several chirallities.

Based on theoretical result in eq. (10), we plot the temperature as a function of radial distance for (6,5) SWNT, as shown in Fig. 3 (a) and (b). We also plot the maximum temperature for several chiralities as shown in Fig. 3 (c) with selected chiralities following the previous work about the UV-vis-NIR spectra [22].

5. Conclusion

Based on the theoretical calculation of laser heating process, we have shown that SWNTs is potential candidates for PTT. The maximum temperature during the laser heating process exist precisely in the interface between SWNT and cancer cells. This maximum temperature is proportional to the diameter of SWNTs. We believe that, our simple model can be used for further theoretical calculation of laser heating process of SWNTs by considering the transient effect.

Acknowledgement

S.S. acknowledge the ImPACT project for research assistantship. I.U.H. is supported by MEXT scholarship.

References

References

- C. T. Avedisian, R. E. Cavicchi, P. L. McEuen, X. Zhou, Nanoparticles for cancer treatment, Annals of the New York Academy of Sciences 1161 (1) (2009) 62–73.
- [2] D. O'Neal, L. R. Hirsch, N. J. Halas, J. Payne, J. L. West, Photo-thermal tumor ablation in mice using near infrared-absorbing nanoparticles, Cancer Letters 209 (2) (2004) 171 - 176. doi:http://dx.doi.org/10.1016/j.canlet.2004.02.004. URL http://www.sciencedirect.com/science/article/pii/S0304383504001442
- [3] D. S. Pissuwan, М. Valenzuela, М. В. Cortie, Therapeutic possibilities of plasmonically heated gold nanoparticles, 24Trends in Biotechnology (2)(2006)6267. _ doi:http://dx.doi.org/10.1016/j.tibtech.2005.12.004. URL http://www.sciencedirect.com/science/article/pii/S0167779905003276
- [4] R. R. Letfullin, C. Joenathan, T. F. George, V. P. Zharov, Laser-induced explosion of gold nanoparticles: potential role for nanophotothermolysis of cancer, Nanomedicine 1 (4) (2006) 473–480.

- [5] C. M. Pitsillides, E. K. Joe, X. Wei, R. R. Anderson, C. P. Lin, Selective cell targeting with light-absorbing microparticles and nanoparticles, Biophysical journal 84 (6) (2003) 4023–4032.
- [6] X. Huang, P. K. Jain, I. H. El-Sayed, M. A. El-Sayed, Plasmonic photothermal therapy (pptt) using gold nanoparticles, Lasers in medical science 23 (3) (2008) 217–228.
- [7] L. O. Svaasand, C. J. Gomer, E. Morinelli, On the physical rationale of laser induced hyperthermia, Lasers in Medical Science 5 (2) (1990) 121–128.
- [8] H. K. Moon, S. H. Lee, H. C. Choi, In vivo near-infrared mediated tumor destruction by photothermal effect of carbon nanotubes, ACS nano 3 (11) (2009) 3707–3713.
- [9] X. Huang, I. H. El-Sayed, W. Qian, M. A. El-Sayed, Cancer cell imaging and photothermal therapy in the near-infrared region by using gold nanorods, Journal of the American Chemical Society 128 (6) (2006) 2115– 2120.
- [10] E. B. Dickerson, E. C. Dreaden, X. Huang, I. H. El-Sayed, H. Chu, S. Pushpanketh, J. F. McDonald, M. A. El-Sayed, Gold nanorod assisted nearinfrared plasmonic photothermal therapy (pptt) of squamous cell carcinoma in mice, Cancer letters 269 (1) (2008) 57–66.
- [11] L. Tong, Q. Wei, A. Wei, J.-X. Cheng, Gold nanorods as contrast agents for biological imaging: optical properties, surface conjugation and photothermal effects, Photochemistry and photobiology 85 (1) (2009) 21–32.
- [12] M.-F. Tsai, S.-H. G. Chang, F.-Y. Cheng, V. Shanmugam, Y.-S. Cheng, C.-H. Su, C.-S. Yeh, Au nanorod design as light-absorber in the first and second biological near-infrared windows for in vivo photothermal therapy, ACS nano 7 (6) (2013) 5330–5342.
- [13] A. M. Smith, M. C. Mancini, S. Nie, Second window for in vivo imaging, Nature nanotechnology 4 (11) (2009) 710.

- [14] K. Yang, S. Zhang, G. Zhang, X. Sun, S.-T. Lee, Z. Liu, Graphene in mice: ultrahigh in vivo tumor uptake and efficient photothermal therapy, Nano letters 10 (9) (2010) 3318–3323.
- [15] F. Zhou, D. Xing, Z. Ou, B. Wu, D. E. Resasco, W. R. Chen, Cancer photothermal therapy in the near-infrared region by using single-walled carbon nanotubes, Journal of Biomedical Optics 14 (2) (2009) 021009-021009-7. doi:10.1117/1.3078803. URL http://dx.doi.org/10.1117/1.3078803
- [16] J. T. Robinson, S. M. Tabakman, Y. Liang, H. Wang, H. Sanchez Casalongue, D. Vinh, H. Dai, Ultrasmall reduced graphene oxide with high near-infrared absorbance for photothermal therapy, Journal of the American Chemical Society 133 (17) (2011) 6825–6831.
- [17] K. Yang, J. Wan, S. Zhang, B. Tian, Y. Zhang, Z. Liu, The influence of surface chemistry and size of nanoscale graphene oxide on photothermal therapy of cancer using ultra-low laser power, Biomaterials 33 (7) (2012) 2206 2214. doi:http://dx.doi.org/10.1016/j.biomaterials.2011.11.064.
- [18] F. Gong, Z. Hongyan, D. V. Papavassiliou, K. Bui, C. Lim, H. M. Duong, Mesoscopic modeling of cancer photothermal therapy using single-walled carbon nanotubes and near infrared radiation: insights through an offlattice monte carlo approach, Nanotechnology 25 (20) (2014) 205101.
- [19] T. Nakamiya, T. Ueda, T. Ikegami, K. Ebihara, R. Tsuda, Thermal analysis of carbon nanotube film irradiated by a pulsed laser, Current Applied Physics 8 (3) (2008) 400–403.
- [20] N. Hamada, S.-i. Sawada, A. Oshiyama, New one-dimensional conductors: graphitic microtubules, Physical Review Letters 68 (10) (1992) 1579.
- [21] R. Saito, M. Fujita, G. Dresselhaus, M. S. Dresselhaus, Electronic structure of chiral graphene tubules, Applied Physics Letters

60 (18) (1992) 2204-2206. doi:http://dx.doi.org/10.1063/1.107080. URL http://scitation.aip.org/content/aip/journal/apl/60/18/10.1063/1.107080

- [22] X. Tu, S. Manohar, A. Jagota, M. Zheng, Dna sequence motifs for structure-specific recognition and separation of carbon nanotubes, Nature 460 (7252) (2009) 250–253.
- [23] A. L. Antaris, J. T. Robinson, O. K. Yaghi, G. Hong, S. Diao, R. Luong, H. Dai, Ultra-low doses of chirality sorted (6, 5) carbon nanotubes for simultaneous tumor imaging and photothermal therapy, ACS nano 7 (4) (2013) 3644–3652.
- [24] T. E. Cooper, G. J. Trezek, A probe technique for determining the thermal conductivity of tissue, Journal of Heat Transfer 94 (2) (1972) 133–140.
- [25] E. Pop, D. Mann, Q. Wang, K. Goodson, H. Dai, Thermal conductance of an individual single-wall carbon nanotube above room temperature, Nano letters 6 (1) (2006) 96–100.
- [26] Z. Qin, J. C. Bischof, Thermophysical and biological responses of gold nanoparticle laser heating, Chemical Society Reviews 41 (3) (2012) 1191– 1217.
- [27] R. Saito, G. Dresselhaus, M. S. Dresselhaus, et al., Physical properties of carbon nanotubes, Vol. 35, World Scientific, 1998.