Development and Evaluation of Video Recordings for the OLSA Matrix Sentence Test

Gerard Llorach^{†1,2,3} gerard.llorach.to@uni-oldenburg.de

> Frederike Kirschner^{2,3} frederike.kirschner@unioldenburg.de

¹Hörzentrum Oldenburg GmbH Oldenburg, Germany Giso Grimm^{2,3} g.grimm@uni-oldenburg.de

Melanie A. Zokoll^{1,2} m.zokoll@hoerzentrumoldenburg.de

² Cluster of Excellence Hearing4All CvO Universität Oldenburg Oldenburg, Germany Kirsten C. Wagener^{1,2} k.wagener@hoerzentrumoldenburg.de

Volker Hohmann^{1,2,3} volker.hohmann@uni-oldenburg.de

³ Auditorische Signalverarbeitung CvO Universität Oldenburg Oldenburg, Germany

ABSTRACT

The matrix sentence test (MST) is an established multi-lingual method for testing speech intelligibility. Most versions of this test are designed with audio-only stimuli. However, visual cues play an important role in speech intelligibility, mostly making it easier to understand speech by speechreading. In this work, we present and evaluate an audiovisual MST, using videos dubbed with the Oldenburger female MST (OLSA). 28 normal-hearing participants completed test and retest sessions with conditions including audio and visual modalities, speech in quiet and noise, and open and closed-set response formats. The levels yielding 80% sentence intelligibility were measured adaptively for the different conditions. In quiet, the audiovisual benefit over audio only was 7 dB in sound pressure level (SPL). In noise, the audiovisual benefit was 5 dB in signal-to-noise ratio (SNR). Speechreading scores ranged from 0% to 84% speech reception in visual-only sentences, with an average of 50% across participants. This large variability in speechreading abilities was reflected in the audiovisual speech reception thresholds (SRTs), which had a larger standard deviation than the audio-only SRTs. Training and learning effects in audiovisual sentences were found: SRTs improved by approximately 3 dB SNR after 5 trials. Participants retained their best scores on a separate retest session and further improved their SRTs by approx. -1.5 dB.

KEYWORDS

OLSA, matrix sentence test, audiovisual, speechreading, speechreading, speech intelligibility

1 Introduction

Speech audiometry is an essential element in audiology [1][2]. It assesses the ability to understand speech acoustically, which is crucial for human communication. The matrix sentence test

(MST) is a well-established method for measuring speech audiometry, and it exists in several languages [3]. MSTs use sentences of 5 words with a "noun - verb - number - adjective - object" structure. There are 10 possible words for each word category (e.g., 10 nouns, 10 verbs, etc.); these are combined to create unpredictable, semantically correct sentences.

Although speech can be understood through sounds only, it is a multimodal process. Being able to see the speaker provides additional cues such as lip movements, which make speech much easier to understand [4]. Audiovisual speech perception has been mentioned as a predictor of real-world hearing disability [5] but it is usually not considered in audiometry. In particular, it is crucial for evaluating severely impaired listeners, where speech communication can be supplemented with visual speech. The MST is intended as a speech test for severely impaired listeners, therefore an audiovisual version is an important extension for its applicability. Nevertheless, audiovisual (or auditory-visual) MSTs have only been developed in Malay, New Zealander English, and Dutch [6][7][8].

The ability to speechread (most commonly known as lipreading) plays a key role in audiovisual speech tests. In particular, audiovisual MSTs are highly affected by speechreading ability. In the Malay MST [6] young, normal-hearing participants scored from 25% to 85% speech reception just by speechreading, i.e., in the visual-only condition. Even when using virtual characters in the German MST [10], a cochlear-implanted participant could understand up to 73% of the visual-only speech. Such visual-only scores indicate that participants are able to understand speech without any acoustic cues. This means that there is a floor effect in the audiovisual MSTs: even if speech is completely masked by noise and not heard, participants achieve their visual-only speech reception threshold.

Recording and validating an MST is quite an extensive undertaking: selection of the phonetically balanced speech material, recording of the speech, cutting and processing of the sound files, making each word equally intelligible to the others, evaluation, and validation [3]. When recording an audiovisual MST, further considerations must be taken into account, such as the head movements and facial expressions of the speaker [6]. Because audio-only speech tests already exist and have been used extensively, it is reasonable to reuse the audio material in audiovisual tests; this also ensures validity across studies. One approach that has been proposed uses virtual characters with lipsynchronization together with existing audio-only speech tests [9, 10, 39]. The advantage of virtual characters is that they can be reused in different configurations with relatively little effort [11]. Another possible approach is to create video recordings dubbed with existing audio for speech tests. The video recording usually provides better quality and is more realistic than a virtual character. Nevertheless, asynchronies between the audio and the video have to be kept below 100 ms in order to pass unnoticed and not affect speech intelligibility [12][13].

One of the advantages of MSTs is that the sentences are unpredictable and there are too many word combinations to be memorized, so consecutive trials can be carried out. Nevertheless, the simple sentence structure and the limited number of words enable participants to learn and improve their results. This training effect has already been shown in audio-only MSTs [14][15] and is particularly noticeable in the first trial, where differences in SRTs of about 1 dB are expected. After 2-4 trials, there is usually an absolute improvement of 2 dB, and the training effects in the following trials are quite small. In audiovisual MSTs, it is expected that participants further improve their SRTs by becoming familiar with the speaker and the visual material [16].

Another factor to take into account is the response format of the MST. After hearing a sentence, participants either repeat what they heard (open-set response format) or select the answers from all possible words (closed-set response format). In the open-set format, a researcher must be present in order to assess whether the answer is correct, while in the closed-set format, participants can do the test by themselves. The closed-set format may give participants an advantage, since they are provided with a list of all possible words; in fact, SRTs have been found to be lower with closed-set type in some MSTs [17][18], although not for German and other languages [3]. Whether such effects appear in audiovisual MSTs has not yet been investigated.

In this work we created an audiovisual version of the female German MST (AV-OLSAf). We recorded videos with a female speaker, dubbed them with the original sentences of the female speaker [15][19] and evaluated the material. Our first contribution is the methodology for producing the dubbed videos and getting the best synchronized video recordings. Our second contribution is the evaluation of the AV-OLSAf with normal-hearing listeners in different conditions: we show the audiovisual training effects in the open-set and closed-set responses; we discuss the speechreading scores and the effects of speechreading in the audiovisual SRTs; and we compare between the audio-only and audiovisual SRTs in noise and in quiet conditions. To conclude, we discuss the implications and recommendations for using the AV-OLSAf.

2 Method

2.1 Recording the Video Material

Although in theory there are 100,000 possible sentences (5 word categories with 10 words per category), the female OLSA uses only 150 predetermined sentences. This relatively small number of sentences permitted us to record videos of the spoken sentences in a single afternoon. We were able to recruit the same speaker that recorded the audio-only version of the German female MST (OLSA) [15][19]. During the recording session, the speaker had to speak the sentences simultaneously while hearing them through an earphone on the right ear. Each sentence was played five times consecutively. Three short "beep" signals were given before each repetition started. The first repetition was used as a reference: the speaker was to listen only in order to know what sentence was coming. In the remaining 4 repetitions, she was to speak simultaneously while hearing the sentence.

The videos of the female speaker were recorded in the studio of the Media Technology and Production of the CvO University of Oldenburg. The available lights of the studio were set up to achieve a homogeneous illumination of the face and of the background green chroma key, as shown in Figure 1. The videos were recorded with a Sony α 7S II camera at 50pfs / full HD, and a condenser microphone in front of the speaker at the height of the knees. The speech was recorded in one channel with a 48 kHz sampling rate and a 16 bit linear pulse-code modulation (LPCM) sample format. An image sample of the final video recordings is shown in Figure 1.

As shown in the schema in Figure 1, a computer was used to reproduce the original OLSA sentences, which at the same time was sending a linear time code (LTC) signal to the second audio channel of the camera. This way, the recorded speech of the session and the original sentences could be synchronized. The recording session lasted around 2 hours in total.



Figure 1: Top, positioning of the speaker and lightning in the studio. Middle, schema of the recording set-up. The red line represents the earphone cable that was reproducing the original sentences. The black cable represents the LTC signal that was sent from the computer to the camera for synchronization. Bottom, example of a frame of the video material.

2.2 Selection of the Videos

We manually discarded videos in which the speaker smiled or showed other non-neutral facial expressions. The recorded speech signals were synchronized to the reproduced original sentences using the LTC signal. When dubbing speech, there are inevitable

Development and Evaluation of Video Recordings for the OLSA Matrix Sentence Test

> asynchronies: time offsets (words spoken too early or too late) and/or words spoken slower or faster than the original words. As all these asynchronies could happen in one single sentence, we used dynamic time warping (DTW) [20] to find the best match between the recordings and the original sentences. The algorithm compares two temporal signals and provides a warping path. We computed the mel spectrograms of the signals and used them for the DTW function. The mel spectrograms were done using frame windows of 46 ms with a frame shift of 23 ms. An example of the mel spectrograms and the corresponding warping path can be seen in Figure 2 and Figure 3. Once the warping path was calculated, we used equations 1 and 2 to compute the asynchrony score:

$$wp_{ii}(n) = DTW(melSpec_i(n), melSpec_i(m))$$
 (1)

$$async_{ij}(n) = RMS(wp_{ij}(n) - n)$$
⁽²⁾

for
$$i = 1,2,3,...,150$$
 (original sentence number) and $j = 1,2,3,4$ (recording n° per original sentence)

where the $melSpec_i(n)$ is the mel spectrogram of the original sentence i, $melSpec_i(n)$ is the mel spectrogram of its corresponding recording (4 recordings j per sentence i), n is the frame number of the $melSpec_i$, m is the frame number of the $melSpec_j$, $wp_{ij}(n)$ is the warping path between the mel spectrograms in frames, $(wp_{ij}(n) - n)$ is the difference in frames, RMS is the root mean square, and $async_{ij}(n)$ is the asynchrony score between the i^{th} original sentence and the j^{th} recording of that sentence. The asynchrony score can be further expressed in seconds instead of frames, as it represents a temporal difference.



Figure 2: Mel spectrogram of original sentence and one of the four recordings of that sentence.

Preprint



Figure 3: Warping path between the original and two recorded sentences. The best match and the worst match are shown. The size of the shaded surface corresponds to the asynchrony score.

We checked the sensitivity of this measure by comparing each recording to its corresponding original sentence and to the remaining unmatching original sentences (Figure 4). For each original sentence, we chose the video recording with the smallest asynchrony score. Of the final best selections, we found three outliers, with asynchrony scores greater than 60 ms, that had to be manually corrected with time offsets. Once corrected, these outliers were shown to 5 normal-hearing participants along with the best-matched sentences; the outliers could not be distinguished from the best-matched sentences and no asynchronies were noticed. We decided that the asynchrony score was small enough (~40 ms) to avoid or at least minimize any perceptual asynchrony/dubbing effects [12][13]. Therefore, we proceeded with the evaluation of the material. The final video recordings can be found in [RefToZenodo].



Figure 4: Asynchrony scores comparing the original sentences and their best-matched recordings (left; 150 scores), the original sentences and all 4 of their recordings (middle; 600 scores) and the original sentences and the mismatched

G. Llorach et al.

recordings (right; 150x149x4 scores). The vertical axis is on a logarithmic scale.

2.3 Evaluation of the Audiovisual Material

2.3.1 Participants. 28 normal-hearing participants (14 female, 14 male) took part in the evaluation measurements. Their ages ranged from 20 to 29 years (mean age 24.9 years). They had normal or corrected-to-normal vision and their pure tone averages (PTAs) in the better ear were between -5 and 7.5 dB HL (mean - 0.31 dB HL). The PTAs were computed using the frequencies 0.5, 1, 2 and 4 kHz. Participants were recruited through the database of the Hörzentrum Oldenburg GmbH and were paid an expense allowance. Permission was granted by the ethics committee of the CvO University of Oldenburg.

2.3.2 Setup. Participants were seated in a chair inside a soundproof booth. The evaluation measurements were done using binaural headphones (Sennheiser HDA 200). A 22" touchscreen display with full HD (ViewSonic TD2220, ViewSonic Corp. Walnut, CA, USA) was placed in front of the participant within arm's reach at a height of 0.8 meters. The experiment was programmed in Matlab2016b. The videos and original sentences were reproduced with VLC 3.03. The acoustic signal was routed with RME Total Mix with an RME Fireface 400 sound card.

The acoustic levels were calibrated using a sound level meter placed at the approximate head position where participants would be seated. The sound and video reproduction was calibrated for synchronization using an external camera. For this purpose, we reproduced a video with frame numbering together with a LTC signal using the experiment setup. The external camera recorded the display screen of the experiment. The LTC signal was connected directly to the external camera instead of the headphones. Using the recording of the external camera we found a consistent asynchrony of 80 ms, which we corrected by delaying the audio signal in the experiment setup.

2.3.3 Stimuli. The acoustic stimulus was the female version of the German matrix sentence test (OLSA) [15] [19] and the visual stimuli was the best-matched video recording (see Section 2.2). For the conditions with noise, we used continuous test-specific noise (TSN) based on the female speech material. The presentation level of the noise was kept constant at 65 dB SPL. The speech level of the first sentence was 60 dB SPL for conditions with and without noise. The adaptive procedure used varied the speech presentation level depending on the responses of the participant.

2.3.4 Conditions. There were nine conditions in the experiment (see Table 1). Each condition used a list of 20 sentences. The sentences in each list were predefined by the MST. In total, we used 45 different predefined lists. The speech presentation levels were adapted after each sentence in order to reach an individual SRT of 80%, i.e. 4 out of 5 words correctly recognized per sentence [21]. During the open-set response format, participants

Development and Evaluation of Video Recordings for the OLSA Matrix Sentence Test

were asked to repeat orally what they understood after each sentence. In the closed-set response format, participants chose the words they understood from an interface displayed on the touch screen after stimulus presentation. The closed-set interface showed all 50 possible words plus one no-answer option per word category. In the visual-only condition (VONoiseClosed), there was no acoustic speech but only test-specific noise at 65 dB SPL. In this condition, the speech could only be understood through speechreading. For this condition, the percentage of correct words per sentence was averaged over 20 sentences (a list). In all conditions, no feedback was given about correctness of responses.

 Table 1. Conditions tested for the evaluation and validation of the AV-OLSA.

		Audio-only (AO)	Audiovisual (AV)	Visual-only (VO)
Noise	Closed-set response	AONoiseClosed	AVNoiseClosed	VONoiseClosed
	Open-set response	AONoiseOpen	AVNoiseOpen	
Quiet $\frac{re}{C}$	Closed-set response	AOQuietClosed	AVQuietClosed	-
	Open-set response	AOQuietOpen	AVQuietOpen	

2.3.5 Training. We added training trials prior to evaluating the nine conditions in order to assess the training effects of the AV-OLSA. We also tested the participants in two different sessions (test, retest). In the first session, 4 audiovisual lists were presented in noise (80 sentences total). Participants were randomly assigned to do the 4 training lists in open-set or closed-set formats (AVNoiseClosed or AVNoiseOpen); 13 participants completed the training in the closed-set format and 15 participants in the open-set format. In the second session, the training was a single list with the same format as the first session (20 sentences in AVNoiseClosed or AVNoiseOpen).

2.3.6 Procedure. For the test trials, participants started with the same response format (open-set or closed-set) as in the training. Next, they did the conditions with the opposite response format. The conditions of one response format were presented in pseudo-randomized order (Figure 5). On the retest session, participants performed a training list with the same response format as on training trials of the first session; then they continued with the conditions with that same format before doing the ones with the other format, as on the test session. The test and retest sessions were temporally spaced from one day to two weeks.

Preprint



Figure 5: Ordering of the trials in the test and retest sessions. Conditions stacked in columns (green and yellow) were pseudo-randomized within the column. If the participants were trained in AVNoise with the open-set format, they performed the open-set format trials before the closed-set trials; if they were trained with the closed-set format, they proceeded with the closed-set format trials before doing the open-set trials.

3 Results

3.1 Ceiling Effects

Our analysis revealed some surprising findings: participants reached SNRs below -20 dB and speech presentation levels below 0 dB SPL (no sound pressure) in the audiovisual conditions. At these levels, there is no contribution of acoustic information to speech reception: the speech detection threshold for the female OLSA is around -16.9 dB SNR in audio-only tests with TSN [22], a threshold that can be theoretically lowered by around -3 dB when adding visual speech [23]. Therefore, below these thresholds (-20 dB SNR and 0 dB SPL), participants used only visual speech in this experiment, i.e., they were speechreading. In consequence, the scores below these thresholds do not represent audiovisual speech perception, but rather visual only. Figure 6 shows that during the adaptive procedure, participants could reach levels where there was no acoustic contribution.

For the analysis of the data, we decided to limit the values that were below the acoustic speech detection thresholds, as they were not representative of audiovisual speech reception. In total, 18 out of 366 audiovisual trials (4.9%) were modified by limiting their SRTs to -20 dB SNR for speech in noise and 0 dB SPL for speech in quiet. The trials affected had varied conditions: of the 18 trials, 3 were training trials, 5 AudiovisualNoiseOpen, 5 AudiovisualNoiseClosed, 3 AudovisualQuietOpen, and 2 AudiovisualQuietClosed. Of the 28 participants, 6 were able to go below the speech detection thresholds.



Figure 6: Adaptive SNRs and speech presentation levels for AVQuiet (top) and AVNoise (bottom) conditions. The adaptive procedure changed the speech levels to reach 80% intelligibility. Below the blue horizontal line, participants understood speech using only visual cues.

3.2 Training Effects

In general, audiovisual SRTs tended to improve across trials. Figure 7 shows the SRTs during the training trials and the test and retest for the audiovisual in noise condition. On average, participants improved their SRTs by -1.6 dB SNR on their third trial. On the test trial, the total improvement was -2.9 dB SNR. On the second session, participants retained the same SRT scores as in their last trial of the first session. SRTs improved further on the trial of the retest session, by -3.8 dB SNR relative to the first trial of the first session. Figure 7 shows that there was a consistent difference of ~1.8 dB SNR between the mean SRTs of the openset and closed-set trials.

G. Llorach et al.



Figure 7: Audiovisual training effects. It should be noted that, due to the other measurement conditions, there could be up to 4 trials in between the Training 4 and Test trials and between Training Retest and Retest trials.

3.3 Speechreading and Audiovisual Benefit

Participants had a wide range of speechreading abilities. The individual VONoiseClosed scores ranged from 0 to 84% intelligibility, had an average of 50% and a standard deviation of 21.4%. Figure 8 shows the distribution of the visual-only scores. There was an average intelligibility improvement of 4.2% in the retest over the test session, although not all participants improved their scores. The intra-individual standard deviation between test and retest scores was 9.4%.



Figure 8: Speechreading scores. Word scoring percentage over 40 sentences per participant. The mean and the median are represented as a red cross and a green square, respectively.

Speechreading scores were correlated with the audiovisual benefit (i.e., the SRT difference between audiovisual and audio-

Development and Evaluation of Video Recordings for the OLSA Matrix Sentence Test

only condition). This correlation can be seen in Figure 9, where the visual-only scores are plotted against the individual SRTs in different conditions. The Pearson's r correlation scores were -0.66 (p<0.001) for AVNoiseClosed, -0.69 (p<0.001) for AVNoiseOpen, -0.65 (p<0.001) for AVQuietClosed, and -0.65 (p<0.001) for AVQuietOpen. Participants that were good speechreaders gained more from having visual information in the audiovisual trials. Whether participants were trained in open-set or closed-set formats did not make any difference for the audiovisual benefit.



Figure 9: Speechreading scores (VONoiseClosed) shown against the audiovisual benefit of each participant; each participant has two circles per plot for test and retest trials. Top left: audiovisual benefit in noise with closed-set response. This condition was the most similar to the visual-only condition, as both had noise and a closed-set format. Top right: audiovisual benefit in noise with open-set format. Bottom: audiovisual benefit in quiet with closed-set (left) and open-set formats (right).

3.4 Audio-only and Audiovisual SRTs

The average SRT differences between audio-only and audiovisual trials were 5.0 dB SNR for speech in noise and 7.0 dB SPL in quiet. The audio-only in noise conditions were the most consistent across participants (std 1.8 dB). The audiovisual trials (std 3.9 dB) and the audio-only speech in quiet trials (std 2.8 dB) were less consistent across participants. The listeners' PTAs were not significantly correlated with the audio-only in quiet scores (Pearson's r = 0.15, p = 0.11). When looking at the intra-subject differences (test minus retest SRTs), we found that audio-only SRTs were consistent from test to retest and that audiovisual SRTs improved by around 1 dB on average on the retest. The average intra-subject differences and standard deviations were 0.82 dB (std 2.02 dB) for audio-only in noise, 0.42 dB (std 2.16 dB) for audio-only in quiet, 1.39 dB (std 2.34 dB) for audiovisual

in noise, and 1.79 dB (std 4.11 dB) for audiovisual in quiet. The response format (open or closed) did not have an effect on the scores, with the exception of the audiovisual in noise trials: in those trials, the closed-set response format reached slightly lower SNRs (~0.5 dB). Furthermore, the training response format affected the AVNoiseOpen results: participants that were trained with the closed-set format, who therefore knew the complete set of words, reached an average SRT of -13.8 dB SNR (std 4.0 dB), whereas the ones who were trained using the open-set format only reached an average SRT of -12.1 dB SNR (std 2.7 dB). Mean SRTs and standard deviations of the test and retest trials are shown in Table 2.

Table 2. Mean audio-only and audiovisual SRTs and interindividual standard deviations in the test and retest sessions (56 scores per cell).

	Mean SRT / dB SNR		Mean SRT / dB SPL
AONoise Closed	-7.9 (2.5)	AOQuietClosed	17.6 (3.2)
AONoiseOpen	-8.8 (1.1)	AOQuietOpen	17.8 (2.4)
AVNoiseClosed	-13.4 (3.2)	AVQuietClosed	10.9 (4.4)
AVNoiseOpen	-12.9 (3.4)	AVQuietOpen	10.5 (4.6)

4 Discussion

4.1 Validity of the Video Material

Overall, the audiovisual speech intelligibility scores were high enough to suggest that there were no asynchronies in the video material [12][13]. Nevertheless, sentences and words should be checked individually for abnormal intelligibility scores that could be due to asynchronies. These asynchronies could be further corrected and adjusted by editing the video material. Participants were not specifically asked to detect asynchronies in the audiovisual material during the evaluation, and none reported any temporal artifacts.

The audio-only and audiovisual scores found were similar to those expected based on the literature. The reference SRT value for the female OLSA is -9.4 dB SNR at 50% intelligibility [15] and we found an SRT of -8 dB SNR at 80% intelligibility. This 1.4 dB difference represents the difference in speech intelligibility (from 50% to 80%). A difference of 3 dB between audio-only and audiovisual scores was reported previously [8], whereas we found a difference of more than 5 dB in the equivalent conditions. This difference could arise from the specific speaker, as some speakers are easier to speechread than others [24], or from language differences [3].

Speechreading scores were in concordance with the literature. In the Malay MST [6], participants reached scores from 25% to 84% word scoring with a mean of 57.1% when speechreading in quiet with a closed-set response format. Similarly, in the Dutch MST [8], the scores ranged from 15% to ~100% word scoring with the same condition. Therefore, our dubbed video recordings provide the same or similar results as video recordings with the corresponding audio. This is particularly relevant for using our method of dubbed visual speech in established audio-only speech tests.

4.2 Advantages of Dubbing

As mentioned in the introduction, one of the advantages of using existing audio material is that it maintains the validity of acoustic speech. For example, van Wanrooij et al. (2019) [8] reported a large variability in intelligibility across words, which probably arose because the word acoustic levels were not balanced and optimized, as is usually done in MSTs. Nevertheless, this does not mean that a non-optimized MST is not usable: MSTs without level adjustments [25] are used in research and can be used to evaluate speech recognition thresholds with almost the same precision.

Another advantage of using existing audio material is that it makes the recording procedure simpler. Limiting the final number of sentences (150) simplifies and speeds up the recording process. Jamaluddin (2016) did not have a final selection of sentences, and so they created all 100,000 possible sentences by re-mixing 100 recorded sentences. During the recording session, they had to ensure that the speaker's head was in the same physical position so that the videos could be cut and blended without artifacts. For this purpose, they had to fabricate a head-resting apparatus to keep the head in the same position. The material required an additional evaluation step to validate the re-mixed recordings, resulting in 600 final sentences. Had they only needed to record 150 final sentences, as in this study and in the study by van Wanrooij et al. (2019) [8], the recording procedure would have been much shorter and simpler.

Another possible solution for creating visual speech, and one that offers more flexibility and control, is virtual characters. Ideally, the virtual character's lip-syncing should achieve the same intelligibility scores as the videos of real speakers. Some of the current virtual characters used in audiological research improve speech intelligibility, but not as much as real speakers. Schreitmüller et al. (2018) [10] used the German MST with virtual characters: CI and NH participants achieved 37.7% and 12.4% average word scoring in the visual-only condition, respectively, which is below the expected scores we found in this study (50%). Similarly, Grimm et al. (2019) [39] used the German MST with virtual characters and compared it to the material presented here (AV-OLSAf), but no SRT improvements were found. Devesse et al. (2018) [9] reported an SRT improvement of 1.5 to 2 dB SNR with virtual characters, while we found a 5 dB

SNR improvement; the speech material in that study was different from ours and thus cannot be compared directly.

We would like to encourage audiovisual MSTs as a tool for evaluating the lip-syncing animations of virtual characters. Most current research in lip animation and visual speech does not consider human-computer communication and speech understanding in their evaluation procedures [26][27]. For this purpose, our videos are published in an open repository [REFtoZenodo], in the hope that other lip-syncing tools also consider speech understanding as a measure.

4.3 Speechreading

The ceiling and floor effects found in the audiovisual MST resulted from the visual speech contribution. This could be because the limited set of words in the MST is easy to learn, to differentiate visually, and to speechread. In sentences for which participants have no previous knowledge of content, one would expect lower speechreading scores, of around 30% [31][38]. Nevertheless, it can be argued that having some expectations about sentence content is probably closer to a real-life conversation.

Another possible factor is that the female speaker was easy to speechread, as there can be differences between speakers. For example, Bench et al. (1995) [24] reported that young female speakers were judged to be easier to speechread than males and older females. We did not make a selection of speakers, as we wanted to have the same person that recorded the audio-only MST. Furthermore, female speakers have been recommended as a compromise between the voice of an adult male and a child [28], so this was a reasonable starting point. Selecting speakers that are more difficult to speechread would probably reduce the ceiling and floor effects.

An interesting alternative to audiovisual MSTs would be to develop a viseme-balanced MST. The audio-only MST is designed to be phonetically balanced, but this does not mean that the visual speech is balanced, as each phoneme does not necessarily correspond to a viseme [29]. Visual cues were previously reported to affect word intelligibility and word error for the AV-OLSAf [30], demonstrating that acoustic speech and visual speech provide different information. Therefore it is possible that the visual speech found in the current MST sentences is not representative of the language tested. Language-specific viseme vocabularies [31] should be developed for this purpose.

That the audiovisual trials were correlated with the speechreading scores was expected [32] [33] [8]. The better a participant was at speechreading, the less acoustic information he or she needed to understand speech. This correlation was present in noise and in quiet conditions; the audiovisual benefit was therefore resilient to the acoustic condition.

Development and Evaluation of Video Recordings for the OLSA Matrix Sentence Test

4.4 Training Effects

The training effects found in the audiovisual MST were bigger than those found in the audio-only MSTs. An improvement of 2 dB SNR after 4-5 lists at 50% speech reception is expected in audio-only MSTs [15]. We found a \sim 3 dB SNR improvement at 80% speech reception; this additional dB probably arose from the participants learning to speechread the material and becoming familiarized with the speaker [16]. Nevertheless, this training effect was not reported in the audiovisual Dutch MST [8] after a familiarization phase with the complete set of words and a training list of 10 audiovisual sentences.

The audiovisual training served as training for the audio-only trials too. In our experiment, participants reached SRTs of -8.04 dB SNR for 80% speech reception. Had there been no training effect, their SNRs would have been higher. For example, in a previous study, participants scored 50% SRT on their first list for an SNR of -6.5 dB [15].

4.5 Individual Differences

The large variability found in audiovisual trials can be explained by individual speechreading abilities and by intrasubject differences. Speechreading scores were highly individual and the participants performed differently in the test and retest trials, with a standard deviation of 2 dB. Therefore, when testing a listener in audiovisual conditions, a standard deviation of 2 dB between trials can be expected. In the case of audio-only speech in noise SRTs, we found little variability among individuals, which was expected, as all participants were young and did not have any hearing disability [35].

Hearing thresholds and noise-induced hearing loss are usually correlated with speech in quiet scores: the worse the hearing levels, the worse the speech intelligibility in quiet [34]. Nevertheless, we did not find this correlation in our study, probably because the PTAs were all very similar and we did not include hearing-impaired participants.

Audiovisual MSTs are particularly relevant for testing severeto-profound hearing-impaired listeners in the clinic. These listeners cannot perform audio-only intelligibility tests and therefore the audiovisual MST would be useful for investigating whether hearing aid or cochlear implant provision improves their audiovisual speech comprehension. Further research should evaluate the AV-OLSAf with hearing-impaired and elderly participants, as some effects are expected: hearing-impaired listeners tend to be better speechreaders [36], and the ability to speechread decreases with age [37].

5 Conclusions

 Videos dubbed with established audio-only MSTs achieved the same performance as audiovisual recordings would have. Preprint

The method presented here keeps the validity of the original audio material while introducing concordant visual speech.

- The audiovisual MST suffers from both ceiling and floor effects, which are closely related to the speechreading abilities of the participant. These effects should be considered when designing experiments for audiovisual perception. High target SRTs such as 80% SRT are recommended instead of 50% SRT in adaptive procedures.
- Audiovisual stimuli gave an SRT benefit of 5 dB SNR in test-specific noise and 7 dB SPL in quiet in comparison to audio-only stimuli for young, normal-hearing participants. Reference values for 80% SRT found in this study were -13.2 dB SNR for audiovisual speech in noise and 10.7 dB SPL for audiovisual speech in quiet.
- At least three or four training lists should be completed in order to reduce training effects. Audiovisual SRTs may improve even further in the retest session, so training effects may continue after a certain number of training trials. More than one trial should be done to properly evaluate an audiovisual condition.
- Audiovisual SRTs correlated with speechreading abilities. The better participants could speechread, the better they performed in the audiovisual conditions.
- Variability in audiovisual SRTs was greater than in audioonly SRTs due to the individual differences in speechreading. For this reason, a larger number of participants should be used to evaluate audiovisual speech reception.

ACKNOWLEDGMENTS

This work received funding from the EU's H2020 research and innovation programme under the MSCA GA 675324 (ENRICH), from the Deutsche Forschungsgemeinschaft (DFG, German Research Foundation) – project number 352015383 (SFB 1330 B1 and C4) and from the European Regional Development Fund – Project "Innovation network for integrated, binaural hearing system technology (VIBHear)". We would like to thank the Media Technology and Production of the CvO University of Oldenburg for helping out with the recordings. Special thanks to Anja Gieseler for giving feedback on the evaluation procedures and the manuscript and to Bernd T. Meyer for counseling on the video selection metric.

REFERENCES

- Talbott, R.E. and Larson, V.D., 1983, August. Research needs in speech audiometry. In Seminars in Hearing (Vol. 4, No. 03, pp. 299-308). Copyright© 1983 by Thieme Medical Publishers, Inc..
- [2] Sanchez Lopez, R., Bianchi, F., Fereczkowski, M., Santurette, S. and Dau, T., 2018. Data-Driven Approach for Auditory Profiling and Characterization of Individual Hearing Loss. Trends in hearing, 22, p.2331216518807400.

- [3] Kollmeier, B., Warzybok, A., Hochmuth, S., Zokoll, M.A., Uslar, V., Brand, T. and Wagener, K.C., 2015. The multilingual matrix test: Principles, applications, and comparison across languages: A review. International Journal of Audiology, 54(sup2), pp.3-16.
- Sumby, W.H. and Pollack, I., 1954. Visual contribution to speech intelligibility in noise. The journal of the acoustical society of america, 26(2), pp.212-215.
 Corthals, P., Vinck, B., Vel, E.D. and Cauwenberge, P.V., 1997. Audiovisual
- [5] Corthals, P., Vinck, B., Vel, E.D. and Cauwenberge, P.V., 1997. Audiovisual speech reception in noise and self-perceived hearing disability in sensorineural hearing loss. Audiology, 36(1), pp.46-56.
- [6] Jamaluddin, S.A., 2016. Development and evaluation of the digit triplet and auditory-visual matrix sentence tests in Malay. Doctoral thesis.
 [7] Trounson, R.H., 2012. Development of the UC Auditory-visual Matrix
- [7] Trounson, R.H., 2012. Development of the UC Auditory-visual Matrix Sentence Test. Master thesis.
- [8] Van Wanrooij, M.M., Van De Rijt, L.P., Roye, A., Mylanus, E.A. and Van Opstal, A.J., 2019. The principle of inverse effectiveness in audiovisual speech perception. Frontiers in Human Neuroscience, 13, pp.335.
 [9] Devesse, A., Dudek, A., van Wieringen, A. and Wouters, J., 2018. Speech
- [9] Devesse, A., Dudek, A., van Wieringen, A. and Wouters, J., 2018. Speech intelligibility of virtual humans. International journal of audiology, 57(12), pp.914-922.
- [10] Schreitmüller, S., Frenken, M., Bentz, L., Ortmann, M., Walger, M. and Meister, H., 2018. Validating a method to assess lipreading, audiovisual gain, and integration during speech reception with cochlear-implanted and normalhearing subjects using a talking head. Ear and hearing, 39(3), pp.503-516.
- [11] Llorach, G., Grimm, G., Hendrikse, M.M. and Hohmann, V., 2018, October. Towards realistic immersive audiovisual simulations for hearing research: Capture, virtual scenes and reproduction. In Proceedings of the 2018 Workshop on Audio-Visual Scene Understanding for Immersive Multimedia (pp. 33-40). ACM.
- [12] Baskent, D. and Bazo, D., 2011. Audiovisual asynchrony detection and speech intelligibility in noise with moderate to severe sensorineural hearing impairment. Ear and hearing, 32(5), pp.582-592.
 [13] Grant, K.W. and Bernstein, J.G., 2019. Toward a Model of Auditory-Visual
- [13] Grant, K. W. and Bernstein, J.G., 2019. Toward a Model of Auditory-Visual Speech Intelligibility. In Multisensory Processes (pp. 33-57). Springer, Cham.
- [14] Wagener, K., Brand, T. and Kollmeier, B., 1999. Entwicklung und Evaluation eines Satztests für die deutsche Sprache I-III: Design, Optimierung und Evaluation des Oldenburger Satztests. ZfA, 38(1-3), pp.4-15.
 [15] Ahrlich, M., 2013. Optimierung und Evaluation des Oldenburger Satztests mit
- [15] Ahrlich, M., 2013. Optimierung und Evaluation des Oldenburger Satztests mit weiblicher Sprecherin und Untersuchung des Effekts des Sprechers auf die Sprachverständlichkeit. Optimization and evaluation of the female OLSA and investigation of the speaker's effects on speech intelligibility. Bachelor thesis.
 [16] Lander, K. and Davies, R., 2008. Does face familiarity influence
- [16] Lander, K. and Davies, R., 2008. Does face familiarity influence speechreadability?. Quarterly Journal of Experimental Psychology, 61(7), pp.961-967.
- [17] Hochmuth, S., Brand, T., Zokoll, M.A., Castro, F.Z., Wardenga, N. and Kollmeier, B., 2012. A Spanish matrix sentence test for assessing speech reception thresholds in noise. International journal of audiology, 51(7), pp.536-544.
- [18] Puglisi, G.E., Astolfi, A., Prodi, N., Visentin, C., Warzybok, A., Hochmuth, S. and Kollmeier, B., 2014. Construction and first evaluation of the Italian Matrix Sentence Test for the assessment of speech intelligibility in noise. In Forum Acusticum 2014 (Vol. 2014, pp. 1-5). European Acoustics Association, EAA.
- [19] Wagener, K., Hochmuth, S., Ahrlich, M., Zokoll, M. and Kollmeier, B., 2014. Der weibliche Oldenburger Satztest. The female version of the Oldenburg sentence test"), in Proceedings of the 17th Jahrestagung der Deutschen Gesellschaft für Audiologie, Oldenburg, Germany.
 [20] Sakoe, H. and Chiba, S., 1978. Dynamic programming algorithm optimization
- [20] Sakoe, H. and Chiba, S., 1978. Dynamic programming algorithm optimization for spoken word recognition. IEEE transactions on acoustics, speech, and signal processing, 26(1), pp.43-49.
- [21] Brand, T. and Kollmeier, B., 2002. Efficient adaptive procedures for threshold and concurrent slope estimates for psychophysics and speech intelligibility tests. The Journal of the Acoustical Society of America, 111(6), pp.2801-2810.
- [22] Schubotz, W., Brand, T., Kollmeier, B. and Ewert, S.D., 2016. Monaural speech intelligibility and detection in maskers with varying amounts of spectrotemporal speech features. The Journal of the Acoustical Society of America, 140(1), pp.524-540.
- [23] Bernstein, L.E., Auer Jr, E.T. and Takayanagi, S., 2004. Auditory speech detection in noise enhanced by lipreading. Speech Communication, 44(1-4), pp.5-18.
- [24] Bench, J., Daly, N., Doyle, J. and Lind, C., 1995. Choosing talkers for the BKB/A Speechreading Test: A procedure with observations on talker age and gender. British journal of audiology, 29(3), pp.172-187.
- [25] Nuesse, T., Wiercinski, B., Brand, T. and Holube, I., 2019. Measuring speech recognition with a matrix test using synthetic speech. Trends in hearing, 23, p.2331216519862982.
- [26] Taylor, S., Kim, T., Yue, Y., Mahler, M., Krahe, J., Rodriguez, A.G., Hodgins, J. and Matthews, I., 2017. A deep learning approach for generalized speech animation. ACM Transactions on Graphics (TOG), 36(4), p.93.

- [27] Jamaludin, A., Chung, J.S. and Zisserman, A., 2019. You said that?: Synthesising talking faces from audio. International Journal of Computer Vision, pp.1-13.
- Vision, pp.1-13.
 [28] Akeroyd, M.A., Arlinger, S., Bentler, R.A., Boothroyd, A., Dillier, N., Dreschler, W.A., Gagne, J.P., Lutman, M., Wouters, J., Wong, L. and Kollmeier, B., 2015. International Collegium of Rehabilitative Audiology (ICRA) recommendations for the construction of multilingual speech tests: ICRA Working Group on Multilingual Speech Tests. International journal of audiology, 54(sup2), pp.17-22.
- [29] Taylor, S.L., Mahler, M., Theobald, B.J. and Matthews, I., 2012, July. Dynamic units of visual speech. In Proceedings of the ACM SIGGRAPH/Eurographics Symposium on Computer Animation (pp. 275-284). Eurographics Association.
 [30] Llorach, G. and Hohmann, V., 2019. Word error and confusion patterns in an
- [30] Llorach, G. and Hohmann, V., 2019. Word error and confusion patterns in an audiovisual German matrix sentence test (OLSA). In Proceedings of the 23rd International Congress on Acoustics, Aachen.
- [31] Fernandez-Lopez, A. and Sukno, F.M., 2017. Automatic viseme vocabulary construction to enhance continuous lip-reading. In Proceedings of the 12th International Joint Conference on Computer Vision, Imaging and Computer Graphics Theory and Applications (VISIGRAPP 2017), 5, pp. 52-63.
- [32] MacLeod, A. and Summerfield, Q., 1987. Quantifying the contribution of vision to speech perception in noise. British journal of audiology, 21(2), pp.131-141.
- [33] Summerfield, Q., 1992. Lipreading and audio-visual speech perception. Philosophical Transactions of the Royal Society of London. Series B: Biological Sciences, 335(1273), pp.71-78.
- [34] Smoorenburg, G.F., 1992. Speech reception in quiet and in noisy conditions by individuals with noise-induced hearing loss in relation to their tone audiogram. The journal of the acoustical society of America, 91(1), pp.421-437.
- [35] Souza, P.E., Boike, K.T., Witherell, K. and Tremblay, K., 2007. Prediction of speech recognition from audibility in older listeners with hearing loss: Effects of age, amplification, and background noise. Journal of the American Academy of Audiology, 18(1), pp.54-65.
- [36] AuerJr, E.T. and Bernstein, L.E., 2007. Enhanced visual speech perception in individuals with early-onset hearing impairment. Journal of Speech, Language, and Hearing Research.
- [37] Tye-Murray, N., Sommers, M.S. and Spehar, B., 2007. The effects of age and gender on lipreading abilities. Journal of the American Academy of Audiology, 18(10), pp.883-892.
- [38] Duchnowski, P., Lum, D.S., Krause, J.C., Sexton, M.G., Bratakos, M.S. and Braida, L.D., 2000. Development of speechreading supplements based on automatic speech recognition. IEEE transactions on biomedical engineering, 47(4), pp.487-496.
- [39] Grimm, G., Llorach, G., Hendrikse, M.M. and Hohmann, V., 2019. Audiovisual stimuli for the evaluation of speech-enhancing algorithms. In Proceedings of the 23rd International Congress on Acoustics, Aachen.