
DIFFERENCES IN PERSPECTIVE ON INERTIAL MEASUREMENT UNIT SENSOR INTEGRATION IN MYOELECTRIC CONTROL

Evan Campbell
Institute of Biomedical Engineering
University of New Brunswick
evan.campbell1@unb.ca

Angkoon Phinyomark
Institute of Biomedical Engineering
University of New Brunswick
aphinyom@unb.ca

Erik Scheme
Institute of Biomedical Engineering
University of New Brunswick
escheme@unb.ca

November 24, 2021

ABSTRACT

Recent human computer-interaction (HCI) studies using electromyography (EMG) and inertial measurement units (IMUs) for upper-limb gesture recognition have claimed that inertial measurements alone result in higher classification accuracy than EMG. In biomedical research such as in prosthesis control, however, EMG remains the gold standard for providing gesture specific information, exceeding the performance of IMUs alone. This study, therefore, presents a preliminary investigation of these conflicting claims between these converging research fields. Previous claims from both fields were verified within this study using publicly available datasets. The conflicting claims were found to stem from differences in terminology and experimental design. Specifically, HCI studies were found to exploit positional variation to increase separation between similar hand gestures. Conversely, in clinical applications such as prosthetics, position invariant gestures are preferred. This work therefore suggests that future studies explicitly outline experimental approaches to better differentiate between gesture recognition approaches.

1 Introduction

Gesture recognition using electromyography (EMG) pattern recognition has a long history of use in biomedical and clinical applications, such as myoelectric control of prosthetic devices and other assistive or rehabilitative technologies. These devices leverage residual motor function to enhance quality of life limited by neurological (stroke [1]) or physical impairment (amputation [2]). The emerging interest in hand gesture recognition as a general human-computer interface (HCI) for consumer applications, such as virtual reality, has large commercial incentives and has therefore accelerated in recent years. The use of wrist- or forearm-worn EMG devices combined with inertial sensors (i.e., accelerometer (ACC), magnetometer (MAG), or gyroscope (GYR)) have demonstrated the potential of such gesture recognition interfaces during offline classification studies [3]. These multi-modal devices have been validated in both biomedical and general HCI studies; however, the conditions of gesture elicitation differ between the two applications.

Biomedical applications of EMG pattern recognition typically require accurate recognition of physiologically appropriate gestures that are robust to variability of daily-living; simply put, the gestures should be reliably decoded regardless of limb posture and contraction intensity, among other factors [4]. Limb posture and contraction intensity variability degrades the usability of clinical EMG pattern recognition systems meaningfully, as gesture recognition accuracies were found to decrease on the order of 13% and 20% for these factors, respectively, across several studies [5]. Interventions in the form of training strategies [6], algorithmic solutions [7], or multi-sensor approaches [8] have lessened this degradation and led to more reliable use of myoelectric control. Multi-sensor approaches using EMG and ACC measurements from many positions have altogether removed degradation caused by static limb positions in recorded positions by sequential use of a position-classifier using ACC, followed by a position-specific EMG classifier for gesture recognition [8]. No application other than position recognition, however, has been validated for non-mechanomyographic ACC measurements within clinical EMG pattern recognition studies.




Figure 1: EMG and ACC measurements (unfiltered) from the biomedical and HCI datasets. The first row contains the EMG elicited during wrist flexion (WF), wrist pronation (WP), and a turning screw (TS) gesture. The second row contains the accelerometer readings for the same contractions, where the black lines represent the x, y, and z components of a forearm mounted sensors and the blue lines represent measurements simultaneously taken at the biceps.

Alternatively, general HCI applications of EMG pattern recognition desire accurate recognition of distinct gestures; the gestures in these application are no longer required to be invariant to daily-living variability and may selectively harness position variability to become more distinct. Consequently, inertial sensors have been found to outperform EMG sensors in terms of gesture recognition accuracy [3,9–11]. For instance, gesture recognition using MAG achieved 93% accuracy across 40 motion classes, whereas EMG achieved only 65%. The different interpretation of the application and value of inertial measurements between biomedical and HCI studies is a current area of confusion in the field that warrants further clarification.

This paper aims to highlight the main differences between biomedical and HCI studies of EMG pattern recognition by examining the differences between gesture elicitation studies. Specifically, this study focused on the differences in the gestures performed and the differences in the use of inertial information. Differences in gestures are presented through visualization of signals, whereas the differing use of inertial information is presented through classification outcomes using EMG and ACC feature sets.

2 METHODS

2.1 Datasets

Two public datasets were adopted to represent biomedical and HCI gesture recognition studies; the Fougner [8] and NinaPro7 [9] datasets, respectively. The biomedical dataset was collected using 8 bipolar Ag/AgCl electrodes (EMG) and 2 tri-axis accelerometers. Twelve intact-limbed subjects performed 6 motions (wrist flexion, wrist extension, wrist pronation, wrist supination, hand close, and pinch grip) and no motion, where each motion was repeated 10 times in 5 different static limb positions. The HCI dataset contained 12 bipolar Ag/AgCl electrodes and 12 tri-axis accelerometers. Twenty intact-limbed subjects performed 40 dynamic gestures (8 finger gestures, 9 wrist gestures, and 23 grasping gestures), where each motion was repeated 6 times with limb position unspecified. The gestures of the HCI dataset were segmented into 3 gesture sets: HCI-A, a set matching the biomedical dataset gestures, HCI-B, a subset containing 8 finger gestures, and HCI-C, a subset containing 23 grasp gestures. A sample of EMG and ACC signals from both datasets is given in Fig. 1.

2.2 Data preparation

The EMG signals from both datasets were pre-processed by a 60 Hz or 50 Hz notch filter and 20-450 Hz bandpass filter to remove power-line interference and motion artefacts, respectively. The ACC signals were pre-processed using 1 Hz low-pass filters, to remove accompanying sensor noise from measurements. Both EMG and ACC signals

Table 1: Multi-gesture position recognition accuracy (mean+std of subjects) across positions of the biomedical dataset

Classifier	ACC		EMG	
	MED	RMS	TD	TDPSD
LDA	99.9+0.3	96.3+5.2	63.0+9.7	62.3+8.0
QDA	99.9+0.1	98.4+1.8	67.8+8.9	66.0+7.6
kNN	100.0+0.0	98.0+2.4	66.8+8.1	54.8+8.4
RF	99.5+0.6	96.3+3.2	66.8+8.4	63.0+8.5

of all channels were segmented into overlapping windows using window length and increment of 200 and 100 ms, respectively.

Features were extracted from each window to create 2 EMG and 2 ACC feature sets. The EMG feature sets were the Hudgins' time-domain (TD) feature set [12] (mean absolute value, zero crossings, slope sign change, and waveform length), and the time domain power spectral descriptors (TDPSD) feature set [7]. The ACC feature sets were the median feature set (MED) and root mean square (RMS) feature set.

2.3 Classification problems

The four feature sets of all four datasets (biomedical, HCI-A, HCI-B, and HCI-C) were used in three classification tasks, where applicable, to validate claims proposed by previous studies.

1. Multi-gesture position classification: Classifiers were trained with feature vectors from all gestures with the class label selected as the position of the gesture. Only the biomedical dataset was used for this analysis, as the HCI dataset did not specify any specific limb positions.
2. Within-position gesture classification: Classifiers were trained with feature vectors from an individual position with the class label being the associated gesture. This process was repeated for all positions in the case of the biomedical dataset and only a single position was assumed for the HCI dataset.
3. Sequential classification: Classifiers were first trained following the multi-gesture position classification task, where feature vectors were used to predict position. Subsequently, the position was used to select the appropriate position-specific gesture classifier, as was conducted in the within-position gesture classification task. As the HCI datasets did not provide labelled positions, they were excluded from this task.

All classification tasks were performed using within-subject leave-one-trial-out cross-validation using linear discriminant analysis (LDA), quadratic discriminant analysis (QDA), k -nearest neighbours (k NN, $k=5$), and random forest (RF, 10 trees) classifiers. Accuracies are presented as mean + standard deviation, where the mean accuracy is the mean accuracy across all subjects and cross-validations, and the standard deviation is the standard deviation across subjects.

3 RESULTS

The multi-gesture position recognition results using ACC MED, ACC RMS, EMG TD, and EMG TDPSD feature sets are shown in Table 1 for the biomedical dataset. The within-position gesture recognition results of the biomedical and HCI datasets were presented in Table 2. The LDA classifier was found to have the best performance among classifiers for all datasets in this latter classification task, again justifying its predominant use in myoelectric control [13]. The EMG TD feature set was found to be best for the biomedical dataset whereas ACC MED was found to be best for all HCI datasets. Further inspection of the performance of the EMG TD feature set with the LDA classifier is provided through the confusion matrices of the biomedical and HCI-A dataset in Table 3. Conversely, Table 3 shows a similar confusion matrix using the best feature set determined for the HCI dataset (ACC MED). Finally, the results of sequential classification of gestures from multiple-positions are presented in Table 4.

4 DISCUSSION

This study corroborates the use of ACC as an accompanying modality in biomedical/clinical applications to achieve positional robustness. Table 1 verifies that accelerometers situated on the forearm and biceps can be used with confidence to decode 5 upper-limb positions in the sagittal plane. Despite encoding similar information from the ACC modality, the MED feature set consistently encoded positional information significantly better ($p < 0.05$) than RMS.

Table 2 provides an upper-limit of accuracy that can be achieved when position recognition is performed without fault. Use of a sequential classification framework achieved no statistical difference between the within-position gesture recognition framework when using ACC MED to segment position and EMG TD to recognize gestures. Although the position recognition performance of MED was statistically better than RMS, no statistical improvement is apparent in the gesture recognition accuracy of the sequential framework using these feature sets to decode position.

This study additionally corroborates the past outcomes of biomedical and HCI studies, where EMG is best for biomedical applications and ACC is best for HCI gesture recognition. Gesture recognition for biomedical applications, such as prosthesis control, relies on class-separability provided through EMG features (96.3%). Although ACC features provide moderate class-separability for the WS (83.9%) and WP classes (87.2%), they provide only marginal class-separability for other classes (mean: 55.7%). HCI gesture recognition results found that ACC features substantially outperformed EMG features with the same set of gestures (HCI-A), a set of finger gestures (HCI-B), and a set of grasping gestures (HCI-C). In contrast to past HCI experiments where 40 gestures are used together, the use of EMG TDPSD for gesture recognition with a subset of wrist gestures provided satisfactory accuracy (91.1%).

Although findings were consistent with past studies, there remains a disconnect between the use of ACC for the recognition of gestures between the biomedical and HCI frameworks. When no positional variance was purposely included (biomedical section of Table 2), ACC provided no real gesture-specific information resulting in low accuracy. The high gesture recognition accuracy achieved using the HCI datasets is most likely an outcome of stratifying gestures across different positions to strategically reduce to improve the separability of the gestures. This use of positional variance can be seen in Fig. 1, where the HCI dataset shows distinct changes in ACC signals during contractions that are uncharacteristic of mechanomyography. This leveraging of positional variance was inferred in [3], where gestures performed “in the air” resulted in higher accuracy than gestures performed when in contact with a surface.

A limitation of this study is the use of static contractions alone in the biomedical dataset. Past studies have found that including ramp contractions can reduce the impact of contraction intensity variability by incorporating more dynamics [14]. It is possible that there may exist repeatable ACC patterns during the transient segment of such ramp contractions that could be leveraged as part of future multi-modal myoelectric control systems.

Ultimately, the consequence of different aims between biomedical and HCI applications can result in confusion when interpreting the outcomes of studies from both fields, especially the when terminology used to describe the gestures does not indicate the aim of the study. In light of this identified deficiency, it is suggested that a full review of past studies be conducted so as to develop a clear taxonomy and set of terminology that could be adopted by both of these expanding fields.

Table 2: Within-position gesture recognition rates across positions

Dataset	Classifier	ACC		EMG	
		MED	RMS	TD	TDPSD
Bio	LDA	69.8+4.4	65.8+4.5	96.2+0.7	96.0+0.4
	QDA	66.4+4.8	64.3+5.1	95.1+0.8	94.2+0.5
	kNN	63.8+5.6	60.8+5.1	94.3+0.9	85.8+1.2
	RF	61.2+4.9	59.2+3.3	92.9+0.7	91.6+0.9
HCI-A	LDA	97.1+1.5	96.6+1.9	89.1+3.5	91.1+2.7
	QDA	93.8+3.8	89.0+5.5	82.9+5.3	68.4+7.0
	kNN	94.2+2.6	94.6+2.4	82.8+4.5	70.1+4.8
	RF	92.0+3.8	92.9+2.5	85.4+3.6	82.3+3.8
HCI-B	LDA	94.4+4.0	94.2+4.1	84.7+8.1	87.5+8.6
	QDA	88.5+8.5	84.4+8.5	75.0+8.4	53.1+10.4
	kNN	87.7+8.8	87.9+8.6	68.3+9.1	50.6+9.2
	RF	84.2+6.9	84.4+7.1	78.4+7.0	73.0+8.4
HCI-C	LDA	89.1+4.4	84.5+6.6	66.5+8.5	71.9+8.5
	QDA	87.9+8.1	84.1+8.9	60.9+9.6	45.9+8.8
	kNN	80.6+9.1	81.7+9.2	52.0+9.8	34.1+7.1
	RF	77.9+8.9	78.2+8.9	62.3+8.6	54.2+8.0

Table 3: Confusion matrix of the gesture classification accuracy (%) of the biomedical and HCI-A datasets using EMG TD and ACC MED feature sets using the LDA classifier

EMG TD												
True Label	Predicted Label - Biomedical						Predicted Label - HCI-A					
	WF	WE	WP	WS	PO	PI	WF	WE	WP	WS	PO	PI
WF	97.4	0.0	1.0	0.9	0.5	0.2	94.4	0.6	0.7	1.0	1.8	1.5
WE	0.2	96.6	0.1	0.7	1.8	0.5	0.2	86.3	7.4	1.7	1.6	2.8
WP	0.2	0.3	96.2	1.9	0.8	0.5	0.0	10.2	83.5	2.3	1.0	3.0
WS	0.1	0.0	1.0	97.8	0.7	0.4	0.3	1.0	2.6	82.7	10.9	2.4
PO	0.1	0.3	0.4	0.5	95.2	3.5	0.3	1.1	1.1	10.8	85.0	1.7
PI	0.3	0.1	0.5	0.5	4.0	94.6	0.1	0.6	0.5	0.4	0.1	98.4

ACC MED												
True Label	Predicted Label - Biomedical						Predicted Label - HCI-A					
	WF	WE	WP	WS	PO	PI	WF	WE	WP	WS	PO	PI
WF	64.3	15.8	2.1	3.2	8.8	5.7	99.9	0.1	0.0	0.0	0.0	0.0
WE	15.8	59.6	1.0	3.3	10.8	9.6	1.0	97.5	1.5	0.0	0.0	0.0
WP	4.0	2.8	87.2	1.0	3.9	1.1	0.4	6.4	92.7	0.5	0.0	0.1
WS	4.4	5.3	0.6	83.9	3.6	2.3	0.0	0.4	0.6	95.1	3.9	0.0
PO	10.0	9.8	2.7	2.7	50.6	24.3	0.0	0.0	0.0	4.8	95.2	0.0
PI	8.8	11.6	2.3	3.5	25.5	48.2	0.0	0.0	0.0	0.0	0.0	100.0

Table 4: Sequential gesture recognition accuracy (mean+std of subjects) of the biomedical dataset

Position	Feature Set	Classifier			
		LDA	QDA	kNN	RF
ACC MED	ACC MED	65.5+17.2	62.4+15.8	60.2+15.5	58.2+15.1
	ACC RMS	61.9+16.4	60.5+15.7	57.5+15.5	55.8+15.4
	EMG TD	96.0+3.2	94.7+4.3	93.9+4.2	92.5+4.0
	EMG TDPSD	95.6+3.5	93.6+4.0	84.7+5.5	91.0+4.6
ACC RMS	ACC MED	63.8+15.9	61.9+15.5	59.5+14.8	57.2+14.8
	ACC RMS	60.5+15.2	60.0+15.3	56.7+14.9	54.1+14.3
	EMG TD	95.5+3.1	94.4+4.2	93.6+4.3	91.5+4.2
	EMG TDPSD	95.2+3.4	93.3+3.8	84.3+5.4	90.6+4.5
EMG TD	ACC MED	50.2+12.0	49.7+11.1	51.1+13.0	47.3+11.5
	ACC RMS	46.7+11.0	47.4+11.0	46.6+12.5	44.1+11.6
	EMG TD	93.4+4.8	92.8+5.3	94.1+4.7	91.2+4.7
	EMG TDPSD	93.2+4.7	92.2+4.5	84.3+4.3	89.8+5.0
EMG TDPSD	ACC MED	49.7+11.7	49.1+11.2	47.0+10.8	46.5+11.5
	ACC RMS	45.8+10.5	47.3+11.1	42.0+9.9	43.4+10.8
	EMG TD	93.3+5.0	92.8+5.6	91.2+5.4	90.2+5.1
	EMG TDPSD	93.5+4.7	92.1+4.5	83.1+4.3	89.2+5.3

References

- [1] N. S. K. Ho, K. Y. Tong, X. L. Hu, K. L. Fung, X. J. Wei, W. Rong, and E. A. Susanto. An emg-driven exoskeleton hand robotic training device on chronic stroke subjects: Task training system for stroke rehabilitation. In *2011 IEEE International Conference on Rehabilitation Robotics*, pages 1–5, June 2011.
- [2] E. Campbell, A. Phinyomark, A. H. Al-Timemy, R. N. Khushaba, G. Petri, and E. Scheme. Differences in emg feature space between able-bodied and amputee subjects for myoelectric control. In *2019 9th International IEEE/EMBS Conference on Neural Engineering (NER)*, pages 33–36, March 2019.
- [3] Shuo Jiang, Bo Lv, Weichao Guo, Chao Zhang, Haitao Wang, Xinjun Sheng, and Peter B Shull. Feasibility of wrist-worn, real-time hand, and surface gesture recognition via semg and imu sensing. *IEEE Transactions on*

- Industrial Informatics*, 14(8):3376–3385, 2017.
- [4] Erik Scheme and Kevin Englehart. Electromyogram pattern recognition for control of powered upper-limb prostheses: state of the art and challenges for clinical use. *Journal of Rehabilitation Research & Development*, 48(6), 2011.
 - [5] Evan Campbell, Angkoon Phinyomark, and Erik Scheme. Current trends and confounding factors in myoelectric control: Limb position and contraction intensity. *Preprints*, 1:1–47, 2020.
 - [6] D. Yang, W. Yang, Q. Huang, and H. Liu. Classification of multiple finger motions during dynamic upper limb movements. *IEEE J. Biomed. Health Inform.*, 21(1):134–141, Jan 2017.
 - [7] Rami N. Khushaba, Maen Takruri, Jaime Valls Miro, and Sarath Kodagoda. Towards limb position invariant myoelectric pattern recognition using time-dependent spectral features. *Neural Networks*, 55:42–58, 2014.
 - [8] ADCCKEA Fougner, Erik Scheme, Adrian DC Chan, Kevin Englehart, and Øyvind Stavdahl. A multi-modal approach for hand motion classification using surface emg and accelerometers. In *2011 Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, pages 4247–4250. IEEE, 2011.
 - [9] Agamemnon Krasoulis, Iris Kyranou, Mustapha Suphi Erden, Kianoush Nazarpour, and Sethu Vijayakumar. Improved prosthetic hand control with concurrent use of myoelectric and inertial measurements. *Journal of neuroengineering and rehabilitation*, 14(1):71, 2017.
 - [10] Rami N Khushaba, Agamemnon Krasoulis, Adel Al-Jumaily, and Kianoush Nazarpour. Spatio-temporal inertial measurements feature extraction improves hand movement pattern recognition without electromyography. In *2018 40th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*, pages 2108–2111. IEEE, 2018.
 - [11] Philipp Koch, Mark Dreier, Martina Böhme, Marco Maass, Huy Phan, and Alfred Mertins. Inhomogeneously stacked rnn for recognizing hand gestures from magnetometer data. In *2019 27th European Signal Processing Conference (EUSIPCO)*, pages 1–5. IEEE, 2019.
 - [12] B. Hudgins, P. Parker, and R. N. Scott. A new strategy for multifunction myoelectric control. *IEEE Transactions on Biomedical Engineering*, 40(1):82–94, Jan 1993.
 - [13] Evan Campbell, Angkoon Phinyomark, and Erik Scheme. Linear discriminant analysis with bayesian risk parameters for myoelectric control. In *2019 IEEE Global Conference on Signal and Information Processing (GlobalSIP)*, pages 1–5. IEEE, 2019.
 - [14] Erik Scheme and Kevin Englehart. Training strategies for mitigating the effect of proportional control on classification in pattern recognition based myoelectric control. *Journal of prosthetics and orthotics: JPO*, 25(2):76, 2013.

This figure "test.png" is available in "png" format from:

<http://arxiv.org/ps/2003.03424v1>