

Low frame rate laser Doppler holography

Léo Puyo,^{1,2,3} Michel Paques,^{2,3} and Michael Atlan⁴

¹Corresponding author: gl.puyo@gmail.com

²Centre Hospitalier National d'Ophthalmologie des Quinze-Vingts,
INSERM-DHOS CIC 1423. 28 rue de Charenton, 75012 Paris France

³Institut de la Vision-Sorbonne Universités. 17 rue Moreau, 75012 Paris France

⁴Institut Langevin. Centre National de la Recherche Scientifique (CNRS). Paris Sciences & Lettres (PSL University).

École Supérieure de Physique et de Chimie Industrielles (EPCIP Paris) - 1 rue Jussieu. 75005 Paris France

(Dated: April 2, 2020)

Laser Doppler holography (LDH) is an interferometric blood flow imaging technique based on full-field measurements of the Doppler broadening. So far, LDH has been demonstrated in the retina with ultrafast cameras, typically at 75 kHz. We show here that it can be implemented with camera frame rates 10 times slower than before. Low and high frequency local power Doppler signals have opposite variations due to energy conservation. A simple contrast inversion of the low frequency power Doppler reveals fast blood flow largely beyond the camera detection bandwidth for conventional laser Doppler measurements. Relevant blood flow variations and color composite power Doppler images can be obtained with camera frame rates down to a few kHz.

Laser Doppler holography (LDH) is a digital holographic method where blood flow is measured from the interference of coherent light backscattered by the eye with a reference beam [1–3]. The Doppler broadening is measured over the full-field array of an ultrafast camera thanks to the coherent gain brought by the reference beam. The power Doppler is calculated pixel-wise as the integral of the high-pass filtered Doppler power spectrum density (DPSD) to reveal blood flow from the larger Doppler broadening of light scattered in blood vessels. By using a sliding short-time window, variations of retinal blood flow over cardiac cycles can be measured in power Doppler units with a few milliseconds of temporal resolution [4]. The method was previously introduced in ophthalmology to image retinal and choroidal blood flow in the human eye [4–6]. LDH measurements demonstrated so far have been performed with ultrahigh camera frame rates, as required to sample Doppler shifts up to a few tens of kHz. Undersampling the Doppler broadening otherwise leads to artifactual blood flow measurements, for example in large vessels lumen or during systole [4]. This frame rate requirement is a serious limitation for the technique as ultrafast camera are expensive, and generate large amounts of data that are challenging to deal with. In this Letter, we demonstrate that blood flow measurements of similar quality can be performed with frame rates 10 times slower than before by taking advantage of the energy conservation of the power spectrum density. We have found that low and high frequency local power Doppler signals have opposite variations, and that simply reversing the contrast of the very low frequency power Doppler allows to obtain relevant blood flow measurements with camera frame rates down to a few kHz. This scheme allows to reveal fast blood flow that is largely beyond the camera detection bandwidth for conventional laser Doppler measurements [7, 8]. The method is first validated by using the low frequency con-

tent of ultrahigh speed measurements and comparing it to the usual blood flow measurements obtained from high frequency power Doppler, which serves as ground truth. The technique is then demonstrated with real low frame rate measurements.

We use the LDH setup presented in [4] to image retinal and choroidal blood flow in human eyes. Informed consent was obtained from the subjects, experimental procedures adhered to the tenets of the Declaration of Helsinki, the study authorization was obtained from the appropriate local ethics review boards CPP and ANSM, and the clinical trial was registered under the references IDRCB 2019-A00942-5, and NCT04129021. Interferograms are recorded using a CMOS camera (Ametek - Phantom V2511, quantum efficiency 40%, 12-bit pixel depth, pixel size 28 μm), which is used in a 512×512 format with ultrafast (i.e. between 60 and 75 kHz), and slow (i.e. between 4 and 8 kHz) frame rates. The ocular exposure to the 785 nm single frequency laser diode is constant, and the camera exposure time is set to the maximum possible. The data processing is essentially the same for both the fast and slow frame rates. The digital holograms are numerically propagated by angular spectrum propagation, and analyzed by short-time Fourier transform. In order to remove the Doppler contribution of eye motion, each short-time window is first filtered by singular value decomposition (SVD), as detailed elsewhere [9]. The eigenvectors of the holograms space-time matrix associated to the eigenvalues of highest energy are rejected, which allows the access to low frequency blood flow signals, otherwise dominated by spurious contributions. Then, for each time point t_n the power Doppler M_0 is calculated as the integral of the power spectrum density S over a given frequency range $[f_1, f_2]$:

$$M_0(x, y, t_n) = \int_{f_1}^{f_2} S(x, y, t_n, f) df \quad (1)$$




FIG. 1: $f_S = 67$ kHz LDH measurement. (a) Power Doppler on the range 0-1 kHz. (b) RC power Doppler on the range 0-1 kHz. (c) The power Doppler on the range 6-33 kHz exhibits very similar contrast. (d) Energy conservation in LDH: the power spectrum is more broadened in blood vessels so the energy is depleted in the low frequency range, so low and high frequency power Doppler images are oppositely contrasted.

Finally, power Doppler images are corrected for non-uniform illumination, and the baseline signal is subtracted to the power Doppler movie [6]. To study blood flow variations with low frame rate acquisitions, the negative power Doppler is considered, which amounts to reversing the contrast of images (inversion of the image grayscale). This process is illustrated in Fig. 1 with a LDH measurement performed at $f_S = 67$ kHz. As noted in previous work [4], blood vessels are visible with a dark contrast on low frequency power Doppler images. This is shown in Fig. 1(a) with a 0-1 kHz power Doppler image: the power Doppler is lower in vascularized structures than in the surrounding tissue. In Fig. 1(b), the contrast of the power Doppler image has been inverted, and this reverse contrast (RC) power Doppler image is very similar to the usual high frequency (6-33 kHz) power Doppler image shown in Fig. 1(c). Retinal and choroidal blood vessels with fast blood flow can be observed on the two images with a similar contrast. This phenomenon can be explained by energy conservation. All pixels correspond to retinal areas that have received the same amount of energy from the laser beam illumination, and the Parseval-Plancherel identity states that the total power Doppler (area under the curve of the power spectrum density) is a quantity that is conserved in the temporal and spectral domains. Thus, aside from the difference of reflectivity and absorption discussed further, the total area under the curve of the DPSD should be the same in vascular and tissue structures because it corresponds to the laser beam energy. In blood vessels, the power spectrum energy is more broadened towards higher frequencies due to the larger Doppler shifts induced by faster scattering parti-

cles. Consequently the spectral energy is more depleted at low frequency in the blood flow areas. Conversely, there is less Doppler broadening occurring in tissue structures, so the power spectrum density remains high at low frequency. Therefore, because the power spectrum density is greater in blood flow areas than in tissue at high frequency, it is also greater in tissue than in blood flow areas at low frequency. A larger Doppler broadening, which is visible with a positive contrast on high frequency power Doppler images, is also visible on low frequency power Doppler images with a negative contrast. In other words, low and high frequency power Doppler images are oppositely contrasted. Other parameters such as the difference of intrinsic structural reflectivity, absorption, forward scattering of the various features also come into play and invalidate the exact energy conservation. For example the RC low frequency power Doppler image does not reflect the content of the high frequency power Doppler image in the area pointed by the arrow in Fig. 1(b) and (c): the absorption is locally greater because of the exposed retinal pigment epithelium (RPE). In this area the strong absorption invalidates the energy conservation, which is why it is not observed with the same contrast. However, the experimental results shown in this Fig. demonstrate that as far as blood flow is concerned, the low frequency power Doppler can be exploited to satisfactorily reveal the high frequency power Doppler content.

In Fig. 1, we demonstrated that reversing the contrast of low frequency power Doppler images produces a similar contrast to high frequency power Doppler images. Now in Fig. 2, we show that it is also true for dynamic changes of power Doppler induced by blood flow variations. With a LDH measurement acquired at $f_S = 67$ kHz, we have measured blood flow in a retinal artery and vein from both the variations of RC power Doppler at low frequency, and normal power Doppler at high frequency. The variations have been normalized for each range, but the relative artery/vein amplitudes have been preserved. Empirically we have determined that for large blood vessels, the 0-4 kHz frequency range is adequate to measure RC power Doppler variations similar to those measured on the range 6-33 kHz. This was the case for all considered LDH measurements. The corresponding power Doppler images are shown in Fig. 2(a) and (b). As shown in Fig. 2(c), the blood flow variations measured with both frequency bands are strikingly similar in both vessels. The variations are slightly noisier on the low frequency range, possibly because it is averaged on a smaller ensemble length in Fourier space. Some minor perturbation can also be observed at low frequency, probably resulting from residual bulk motion that is not totally suppressed by the SVD filtering. These results show that negative variations of power Doppler at low frequency reflect positive variations of power Doppler at very high frequency. Over the course of cardiac cycles,




FIG. 2: $f_S = 67$ kHz, validation of the use of reversed contrast low frequency power Doppler to measure pulsatile blood flow. (a) RC power Doppler on the 0-4 kHz range. (b) Power Doppler on the 6-33 kHz range. (c) Variations measured in an artery and a vein. Corresponding movies in [Visualization 1](#).

when blood flow is increased, the energy is shifted towards higher frequencies, which means the low frequency power Doppler is depleted. Thus, probing the negative power Doppler variations at low frequency allows to retrieve information about the positive variations of high frequency power Doppler.

The results shown in Fig. 1 and 2 suggest that using a low frame rate and reversing the power Doppler contrast could be used to perform blood flow measurements similar to those obtained from ultrafast frame rates LDH acquisitions. This technique is verified in Fig. 3 with a LDH measurement recorded at low frame rate. In order to access the frequency band 0-4 kHz, the sampling frequency was set to $f_S = 8$ kHz. It is possible to see on the RC power Doppler image shown in Fig. 3(a) that fast blood flow can indeed be revealed as expected. The corresponding variations of negative power Doppler in an artery (red) and a vein (blue) are plotted in Fig. 3(c). It is possible to observe the systolic increase, diastolic decrease in the artery, and a cycloidal flow profile in the vein, which are the waveforms typically measured with high frequency LDH. The coefficient of variation map that was previously introduced to differentiate retinal arteries and veins is shown in Fig. 3(b), and shows a quality similar to what was obtained at high frequency in our previous work [6]. The results presented in this Fig. demonstrate that perfusion maps and blood flow




FIG. 3: $f_S = 8$ kHz LDH measurement. (a) RC power Doppler image on the frequency range 0-4 kHz. (b) The local power Doppler coefficient of variation differentiates arteries and veins. (c) The RC power Doppler variations in arteries and veins exhibit the waveforms typically observed with high frequency LDH measurements. Movie in [Visualization 2](#).

variations of fast blood flow largely beyond the detection bandwidth of the camera can be revealed by inverting the contrast of low frame rate LDH measurements. The results shown in Fig. 1 and 2 are obtained from high frequency LDH measurements, where it can be assumed that the Doppler broadening is properly sampled. This means the measured power spectrum density reflects the real Doppler broadening of light. In the case where the measurement is done at low frequency, the undersampled part of the spectrum is aliased, and it could be expected the ability of the RC low frequency power Doppler to reflect positive high frequency power Doppler would be jeopardized. However the results obtained with the low frame rate LDH measurement shown in Fig. 3 show that it is not the case.

In our previous work [4], we showed that composite power Doppler images from low and high frequency ranges that reveal slow and fast blood on a single image can be very useful. The typical frequency ranges used for these images are 1-6 kHz and $6 \text{ kHz} - f_S/2$, where f_S is usually between 60 and 75 kHz. Arterial and venous choroidal vessels are differently contrasted on these composite images, thanks to the natural discrepancy of flow between these vessels. We demonstrate in Fig. 4 that composite images with very similar contrast can be generated from RC low frequency LDH measurement. LDH measurements were performed in the optic nerve head




FIG. 4: Making low/high flow composite images with high and low frame rates. (a), (b), and (c) show the usual composite images previously demonstrated [5]. (d), (e), and (f) show the result with a 8 kHz frame rate: the high flow image is obtained from the RC low frequency power Doppler. The composite movies are shown in [Visualization 3](#) and [Visualization 4](#).

(ONH) of a same eye with sampling frequencies of 67 kHz and 8 kHz. In Fig. 4(a) and (b) are shown the usual power Doppler images for the low and high flow for $f_s = 67$ kHz, and in Fig. 4(c) the resulting composite. Then for the measurement at $f_s = 8$ kHz, we show in Fig. 4(d) that a similar low flow power Doppler image is obtained from the 1-4 kHz frequency range. Then, as shown in Fig. 4(e), the RC power Doppler image on the range 0-1 kHz produces an image equivalent to the very high frequency power Doppler, which reveals the fast flow. Consequently, combining these two images into the color composite Doppler image displayed in Fig. 4(f) leads to an image contrasted very similarly to the composite image obtained with the 67 kHz measurement. Retinal vessels can be observed on both images with the same contrast. More importantly, as shown by the arrows 'CA' and 'CV', the choroidal artery and vein




FIG. 5: $f_s = 4$ kHz LDH measurement. (a) RC power Doppler image on the 0-2 kHz range, the corresponding movie is shown in [Visualization 5](#). (b) The high dynamic range composite color image reveals both slow and fast blood flow in the ONH microvasculature and in large vessels.

in the vicinity of the ONH are similarly differentiated on both composite images. This shows that colored composite images of equivalent quality to those obtained from ultrafast LDH measurements can be obtained from low frame rate LDH measurements.

In Fig. 5, we demonstrate blood flow measurement with RC-LDH by using a camera frame rate set to $f_s = 4$ kHz. The average RC power Doppler image for the 0-2 kHz range shown in Fig. 5(a) is able to reveal the perfusion map of fast blood flow. To obtain a color composite Doppler image from this measurement, the 0-1 kHz RC power Doppler and the 1-2 kHz power Doppler have been used to reveal fast (red) and low (cyan) flow. The resulting image shown in Fig. 5(b) reveals the ONH microvasculature as well as fast blood flow in large vessels. The 0-1 kHz range in RC-LDH produces perfusion maps of satisfying quality. However, the pulsatile variations of blood flow in large vessels are revealed by higher frequencies than those revealing the steady component of blood flow [6]. This is why we used the 0-4 kHz frequency range to measure pulsatile variations in Fig. 3 and 4. When the local velocity is greater, the Doppler shifts are greater, so the spectrum energy is broader and the spectral depletion revealing blood flow by RC power Doppler is increased. This means that higher velocities are also revealed by greater frequencies in RC-LDH. Roughly speaking, we found that blood flow revealed with normal power Doppler in the range of a few kHz and tens of kHz, seems to be revealed with RC power Doppler in the range of a few hundreds of Hz, and a few kHz, respectively. Slow blood is very challenging to reveal with the RC power Doppler because ultralow frequencies are very corrupted by spurious contributions. Thus, it is advantageous to use frequencies of a few kHz for the dual purpose of measuring the pulsatile variations of blood flow with RC power Doppler, and revealing slow blood flow with normal power Doppler. Interestingly, the low frequency range, around 1 to 4 kHz, is able to reveal

both slow and fast blood flow, by using RC or normal power Doppler. This frequency range is very affected by spurious interferometric contributions, so the use of spatio-temporal filtering is critical to reveal the signal. We have found the SVD to be efficient to reject the specular reflections from instrument diopters and the anterior segment. It is particularly important to filter the pulsatile axial motion of the eye because they are synchronous with blood flow and occur in the direction that maximizes the Doppler effect.

The comparisons we have conducted between RC power Doppler and normal power, from low and high frequency LDH measurement, empirically validate the use of the contrast inversion. Considering there is a 10 fold frame rate reduction, the similarity between the obtained blood flow traces and color composite images is very satisfying. For measurements of a same duration, it seems that the signal quality of blood flow measurements is slightly lower with RC power Doppler, possibly because there are less pixels involved in the signal detection. However this is counterbalanced by the possibility to perform longer acquisitions. The quality of the perfusion maps, blood flow variations, color composite Doppler images obtained by RC-LDH seems of sufficient quality for clinical use. As current streaming cameras have an available throughput allowing measurement at frame rates of a few kHz, this work leaves the way open towards a real-time LDH system.

In conclusion, the detection of low frequency light fluctuations by digital holography enables efficient characterization of high frequency Doppler broadening beyond the Shannon-Nyquist sampling limit. Perfusion maps and pulsatile blood flow variations of flow velocities largely out of the reach of conventional wideband laser Doppler measurements can be obtained. Color composite Doppler images of low/high flow can also be produced with surprisingly low camera frame rates.

Funding Information

This work was supported by the European Research Council (ERC Synergy HELMHOLTZ #610110), the In-

stitut Hospitalo-Universitaire ForeSight (ANR 18-001), and the Sesame program of the Region Ile-de-France (4DEye project). The Titan Xp used for this research was donated by the NVIDIA Corporation.

Disclosures

The authors declare no conflicts of interest.

-
- [1] M. Simonutti, M. Paques, J. A. Sahel, M. Gross, B. Samson, C. Magnain, and M. Atlan. Holographic laser doppler ophthalmoscopy. *Opt. Lett.*, 35(12):1941–1943, 2010.
 - [2] Caroline Magnain, Amandine Castel, Tanguy Boucneau, Manuel Simonutti, Isabelle Ferezou, Armelle Rancillac, Tania Vitalis, José-Alain Sahel, Michel Paques, and Michael Atlan. Holographic laser doppler imaging of microvascular blood flow. *JOSA A*, 31(12):2723–2735, 2014.
 - [3] Mathilde Pellizzari, Manuel Simonutti, Julie Degardin, J-A Sahel, Mathias Fink, Michel Paques, and Michael Atlan. High speed optical holography of retinal blood flow. *Optics Letters*, 41(15):3503–3506, 2016.
 - [4] L. Puyo, M. Paques, M. Fink, J.-A. Sahel, and M. Atlan. In vivo laser doppler holography of the human retina. *Biomedical Optics Express*, 9(9):4113–4129, Sep 2018.
 - [5] Léo Puyo, Michel Paques, Mathias Fink, José-Alain Sahel, and Michael Atlan. Choroidal vasculature imaging with laser doppler holography. *Biomedical Optics Express*, 10(2):995–1012, 2019.
 - [6] Léo Puyo, Michel Paques, Mathias Fink, José-Alain Sahel, and Michael Atlan. Waveform analysis of human retinal and choroidal blood flow with laser doppler holography. *Biomedical Optics Express*, 10(10):4942–4963, 2019.
 - [7] T. Tanaka, C. Riva, and I. Ben-Sira. Blood velocity measurements in human retinal vessels. *Science*, 186:830–831, November 1974.
 - [8] M. D. Stern. In vivo evaluation of microcirculation by coherent light scattering. *Nature*, 254(5495):56–58, March 1975.
 - [9] Leo Puyo, Michel Paques, and Michael Atlan. Spatio-temporal filtering in laser doppler holography. arXiv preprint, 2020.