Do CNNs solve the CT inverse problem?

Emil Y. Sidky, *Member, IEEE*, Iris Lorente, Jovan G. Brankov, *Senior Member, IEEE*, and Xiaochuan Pan, *Fellow, IEEE*

Abstract—Objective: This work examines the claim made in the literature that the inverse problem associated with image reconstruction in sparse-view computed tomography (CT) can be solved with a convolutional neural network (CNN).

Methods: Training and testing image/data pairs are generated in a dedicated breast CT simulation for sparse-view sampling, using two different object models. The trained CNN is tested to see if images can be accurately recovered from their corresponding sparse-view data. For reference, the same sparse-view CT data is reconstructed by the use of constrained total-variation (TV) minimization (TVmin), which exploits sparsity in the gradient magnitude image (GMI).

Results: Using sparse-view data from images either in the training or testing set, there is a significant discrepancy between the image obtained with the CNN and the image that generated the data. For the same simulated scanning conditions, TVmin is able to accurately reconstruct the test image.

Conclusion: We find that the sparse-view CT inverse problem cannot be solved for the particular published CNN-based methodology that we chose and the particular object model that we tested. Furthermore, this negative result is obtained for conditions where TVmin is able to recover the test images.

Significance: The inability of the CNN to solve the inverse problem associated with sparse-view CT, for the specific conditions of the presented simulation, draws into question similar unsupported claims being made for the use of CNNs to solve inverse problems in medical imaging.

Index Terms—CT image reconstruction, sparse view sampling, total variation, convolutional neural networks, deep-learning, inverse problems

I. INTRODUCTION

UCH recent literature on computed tomography (CT) image reconstruction has focused on data-driven approaches to "solve" the associated inverse problem. In particular, deep-learning with convolutional neural networks (CNN) is being developed for image reconstruction from sparse-view projection data [1]–[3]. As a procedure for finding an image from sparsely-sampled projection data, there is no fundamental logical problem in the use of CNNs. The claim, however, that CNNs *solve* the associated inverse problem remains unsubstantiated in the literature.

In the sparse-view CT literature using CNNs, there is not clear evidence that an associated inverse problem is being solved. There does not seem to be a framework that makes it

This work is supported in part by NIH Grant Nos. R01-EB026282, R01-EB023968, and R01-EB023969. The contents of this article are solely the responsibility of the authors and do not necessarily represent the official views of the National Institutes of Health.

E. Y. Sidky and X. Pan are with the Department of Radiology at The University of Chicago, Chicago, IL, 60637 (e-mails: sidky@uchicago.edu, xpan@uchicago.edu).

I. Lorente and J. G. Brankov are with the Department of Electrical and Computer Engineering at the Illinois Institute of Technology, Chicago, IL, 60616 (e-mails: irislorente@gmail.com, brankov@iit.edu).

clear how to determine CNN parameters, network design, and training sets to obtain an accurate inverse problem solution. In fact, in the majority of sparse-view CT CNN literature there is not even a clear statement of what is the inverse problem being solved. The lack of rigorous study of CNNs as an inverse problem solver for CT has motivated us to perform our own investigation.

We focus specifically on the problem of sparse-view CT image reconstruction, because many of the published works on CNN-based image reconstruction address this problem. The only known framework that has been shown to solve sparse-view CT inverse problems is sparsity-exploiting image reconstruction. This type of image reconstruction has typically been formulated as a large-scale non-smooth convex optimization, where a substantial theoretical framework has been built-up over the past few years.

We design a simulation based on dedicated breast CT [4], where minimizing X-ray dose is of paramount importance and the question of reducing the necessary number of projections provides an avenue for achieving this dose reduction. Exploiting gradient-sparsity has proven particularly effective, and we employ constrained total-variation (TV) minimization to show inversion of the sparse-view CT inverse problem and to provide a reference for the CNN-based investigation. Using a stochastic 2D breast model phantom, we generate ideal data for training the CNN to solve the sparse-view CT inverse problem.

In this work, we define the sparse-view CT inverse problem, explain the evidence that TVmin solves this inverse problem, and demonstrate that we are unable to solve this inverse problem using a published CNN technique that claims to do exactly that. A brief synopsis of inverse problems is given in Sec. II. The measurement model and model inverse for sparsity-exploiting image reconstruction is presented in Sec. III. Sparse-view image reconstruction with CNN-based deeplearning is outlined in Sec. IV from an inverse problem perspective. The object model for the breast CT simulation is specified in Sec. V. Results for the sparse-view image reconstruction using TVmin and the CNN are shown in Sec. VI. Finally, we conclude in Sec. VII.

II. SOLVING AN INVERSE PROBLEM

What is meant by "solving an inverse problem" requires explanation, and we follow the notation of [5]. Inverse problems start with specifying a measurement/data model

$$y = \mathcal{M}(x),$$

where \mathcal{M} is an operator that yields data y given model parameters x. An important part of the measurement model also involves specification of any constraints on x.

The investigation of the measurement model begins with addressing measurement model injectivity; namely, that the measurement model is one-to-one,

$$\mathcal{M}(x_1) = \mathcal{M}(x_2) \Rightarrow x_1 = x_2. \tag{1}$$

In words, if two measurements in the range of \mathcal{M} are the same then they correspond to the same parameter set. As an example, if \mathcal{M} is a linear transform then injectivity is obtained if there are no non-trivial null vectors on the right of \mathcal{M} .

Once injectivity is established, an inverse to the measurement model \mathcal{M}^{-1} can be constructed. It is then logical to address stability of the inverse of \mathcal{M} ; namely, if the difference between two data vectors y_1 and y_2 is small then the difference between their corresponding parameter estimates $\mathcal{M}^{-1}(y_1)$ and $\mathcal{M}^{-1}(y_2)$ is also small

$$\|\mathcal{M}^{-1}(y_1) - \mathcal{M}^{-1}(y_2)\| \le \omega(\|y_1 - y_2\|),$$
 (2)

where $\omega(\cdot)$ is a monotonic function, mapping positive reals to positive reals and $\omega(0)=0$. The measurement model inverse \mathcal{M}^{-1} is stable if $\omega(t)=\epsilon t$, where ϵ is a small positive number [5].

Stability of the inverse plays a large role in applying the model inverse to real data or simulated data, which originates from a measurement model with greater realism. These possibilities can be conceptually summarized as a "physical" measurement model

$$\mathcal{M}_{\text{phys}}(x) = \mathcal{M}(x) + \epsilon_{\text{det}}(x) + \epsilon_{\text{noise}}(x),$$

where $\epsilon_{\text{det}}(x)$ and $\epsilon_{\text{noise}}(x)$ represent, respectively, deterministic and stochastic error, which are both possibly dependent on x. Applying \mathcal{M}^{-1} to $\mathcal{M}_{\text{phys}}$, yields

$$\mathcal{M}^{-1}\mathcal{M}_{\text{phys}}(x) = x + \delta, \tag{3}$$

where the magnitude of δ is bounded with the stability inequality in Eq. (2) with $y_1 = \mathcal{M}(x)$ and $y_2 = \mathcal{M}_{phys}(x)$.

With this brief background, there is context for discussing what is meant by solving an inverse problem. The only step in the analysis of an inverse problem that involves solving an equation is deriving an inverse for the measurement model

$$y = \mathcal{M}(x)$$
.

The model inverse is said to solve the measurement model equation if it can be shown that

$$\mathcal{M}^{-1}\left(\mathcal{M}(x)\right) = x. \tag{4}$$

Solving an inverse problem should not be confused with applying this model inverse to a different model that perhaps describes the same physical system, with a greater degree of realism, i.e.

$$\mathcal{M}^{-1}\left(\mathcal{M}_{phys}(x)\right) \neq x.$$

The way that \mathcal{M}_{phys} enters the inverse problem analysis is through the stability inequality in Eq. (3). This inequality in turn relies on deriving the inverse to the idealized measurement model \mathcal{M} .

Numerical investigation of inverse problems

Many inverse problems of interest are not amenable to analytic methods and are investigated numerically through the development of numerical algorithms. In the numerical setting, the parameter and data spaces are necessarily finite-dimensional and accordingly x and y are finite length vectors. A numerical algorithm for inverting \mathcal{M} codes the action of a proposed model inverse. In numerical investigation of an inverse problem, we have a measurement model and a proposed inverse to that model. We may not know: (1) that the measurement model is injective, (2) that the algorithm implements the proposed model inverse, or (3) that a proposed inverse actually inverts the measurement model.

The stages of numerical inverse problem investigation starts with verifying that the algorithm does indeed implement the proposed model inverse. Having verified the algorithm, it is possible that the proposed model inverse does not actually invert the measurement model; thus the next step is to generate simulated data

$$y_{\text{test}} = \mathcal{M}(x_{\text{test}})$$

from a realization $x_{\rm test}$ and observe that $x_{\rm test}$ can be recovered by applying the algorithm to $y_{\rm test}$. If, however, the verified algorithm fails to recover $x_{\rm test}$ there are two possible reasons for the failure: the proposed model inverse might not invert the measurement model, or the measurement model itself might not be injective and it is not possible to construct a model inverse at all. Numerical investigation of the model inverse stability is also important, and to some extent, accurate recovery of the test image also provides evidence for stability in that numerical error does not get magnified substantially, but a thorough investigation on stability entails a characterization of the inverse in response to data inconsistency and noise.

There are limitations to numerical inverse problem studies. They are empirical by nature. Knowledge of successful measurement model inversion is limited to the set of trial model parameter sets x_{test} that are successfully recovered. Furthermore, in most cases of interest, x_{test} is only recovered to within some level of computational error. One can only amass evidence that a proposed model inverse actually inverts the measurement model of interest. As part of this evidence it is crucial to report metrics that reflect the verification of the algorithm and metrics that quantitate the recovery of x_{test} .

Convex optimization-based inverses

For CT image reconstruction, there has been much recent work on the use of convex optimization to provide an inverse to a relevant measurement model. In this setting, the model inverse is formally expressed as

$$x^* = \mathcal{M}^{-1}(y_{\text{test}})$$

$$= \underset{x}{\arg \min} \Phi[y_{\text{test}}](x) \text{ such that } x \in S_{\text{convex}}[y_{\text{test}}]$$

where Φ is a convex objective function and x is restricted to convex set $S_{\rm convex}$, encoding prior knowledge on x. Either Φ or $S_{\rm convex}$ may depend on data vector $y_{\rm test}$. In the vast majority of cases of interest to CT image reconstruction, this optimization problem is solved by an iterative algorithm.

Numerical algorithm verification consists of (1) showing that the iterates approach the optimality conditions of this optimization problem and (2) showing that the iterates generate data that approaches y_{test} . Approaching the optimality conditions ensures that

$$x_k \to x^* \text{ as } k \to \infty,$$
 (5)

where k is the iteration index of the algorithm. Approaching the test data

$$\mathcal{M}(x_k) \to y_{\text{test}}$$

can be measured by the data root-mean-square-error (RMSE)

$$\sqrt{\|y_{\text{test}} - \mathcal{M}(x_k)\|_2^2/\text{size}(y_{\text{test}})} \to 0.$$
 (6)

Showing evidence for Eqs. (5) and (6) verifies that the algorithm implements the proposed inverse.

Evidence that the proposed inverse actually inverts the measurement model is established by showing that the iterates of the numerical algorithm approaches test image

$$x_k \to x_{\text{test}}$$
.

This can be measured by the image root-mean-square-error (RMSE)

$$\sqrt{\|x_{\text{test}} - x_k\|_2^2/\text{size}(x_{\text{test}})} \to 0.$$
 (7)

We reiterate that a failure to show Eq. (7) can either be a result of the non-injectivity of the measurement model – i.e. there is more than one image that yields y_{test} – or the proposed inverse does not invert \mathcal{M} .

III. SPARSE-VIEW CT AND COMPRESSED SENSING

The most common measurement model for the CT system involves line integration over an object model. The line-integration model takes the form of the X-ray or Radon transform for a divergent ray or parallel ray geometry, respectively. The present study employs a parallel beam geometry in a 2D CT scanning setting. For iterative image reconstruction, the continuous line-integration model is converted to a discrete-to-discrete transform by representing the object function with an image array of pixels. The measurement model parameters are the values of the image pixels and the measurement model ${\cal M}$ is

$$g = Rf, (8)$$

where f denotes the pixel values; R is a discrete-to-discrete linear transform representing the action of the Radon transform on the image pixels; and g indicates the sinogram values. Analyzing Eq. (8) as an inverse problem is straight-forward. If the matrix R is left-invertible then the measurement model is one-to-one and two different sets of pixel values, f, yield different sinograms g and a left-inverse R^{-1} can be constructed. Stability of performing this inversion depends on the singular value spectrum of the particular discrete-to-discrete version of the Radon transform [6].

For sparse-view sampling, the number of projections in the sinogram is small enough that R is a short, fat matrix; i.e. there are more columns than rows. In this case, injectivity is lost and it can occur that two different images f can yield

the same sinogram under the action of R. The conjecture of Compressed Sensing [7], [8] is that injectivity of Eq. (8) can be restored by restricting the image f to be sparse or sparse in some transform Tf.

The modified measurement model ${\mathcal M}$ studied in Compressed Sensing is thus

$$g = Rf$$
 where $f \in F_{\text{sparse}}$ and $F_{\text{sparse}} = \{f \mid ||Tf||_0 \le s\},$ (9

where T is the sparsifying transform; $\|\cdot\|_0$ is the counting norm, which yields the number of non-zero elements in the argument vector; and s is an integer bounding the number of non-zeros in Tf. Direct restriction in the sparsity of f is the special case where T=I, the identity matrix. Note that the the measurement model $\mathcal M$ includes both the transform R and the sparsity restriction on f.

The canonical inverse to this measurement model is specified implicitly by the formal optimization problem

$$f^* = \underset{f}{\arg\min} \|Tf\|_0 \text{ such that } g = Rf.$$
 (10)

From a computational point of view this inverse is not practical to evaluate, especially for full-scale CT systems, due to the ℓ_0 -norm in the objective function.

As suggested in Ref. [7], a more practical inverse to the sparsity-restricted measurement model may be obtained by the convex relaxation of the ℓ_0 -norm in Eq. (10) to the ℓ_1 -norm.

$$f^* = \underset{f}{\arg\min} \|Tf\|_1 \text{ such that } g = Rf.$$
 (11)

where $\|\cdot\|_1$ is the ℓ_1 -norm, which sums the absolute value of the vector components. While analytic results for Compressed Sensing have been established for certain classes of measurement models [9], these results do not cover the line-integration based measurement models used in CT and we must turn to numerical investigation and amass results from CT simulations.

With numerical simulation, the algorithm is verified by showing that the solution of Eq. (11) is attained. Successful recovery of test image f_{test} from data $g_{\text{test}} = Rf_{\text{test}}$ is evidence that Eq. (11) provides the inverse of Eq. (9) and that measurement model Eq. (9) is one-to-one. Failure to recover f_{test} could mean either that the measurement model is not injective or that the proposed inverse does not actually invert the measurement model. The convex relaxation of the ℓ_0 -norm to the ℓ_1 -norm, in particular, does raise the distinct possibility that the Compressed Sensing model could be injective while Eq. (11) might not provide its inverse.

For Compressed Sensing style studies the goal is to establish a relationship between the number of samples needed for accurate image recovery and sparsity in the image or transformed image. A practical approach to achieving this relationship numerically is to perform a type of Monte Carlo investigation where test images are generated randomly at various sparsity levels. The information on the sampling level required to recover the test images as a function of their sparsity can be summarized in a Donoho-Tanner phase diagram [10], [11]. For full-scale CT systems it is of interest to perform numerical inverse problem related investigations on configurations and




Fig. 1. Three realizations of a computerized breast phantom with stochastic distribution of fibroglandular tissue. The top row shows realizations of the binary phantom, and the bottom row shows the corresponding smooth-edge phantom, generated by blurring with a gaussian kernel of one pixel width. The breast phantom is composed of a 16 cm disk containing background fat tissue, attenuation 0.194 cm $^{-1}$, skin-line and fibroglandular tissue at attenuation 0.233 cm $^{-1}$. The generated phantom images are 512 \times 512 pixels and are shown in a gray scale window of [0.18, 0.24] cm $^{-1}$.

test objects that reflect properties of a CT system of interest. We present such studies here so as to have a reference for numerical inverse problem studies with CNNs.

Gradient sparsity exploiting image reconstruction

A particularly successful sparsity model for CT imaging is seeking sparsity in the image gradient magnitude, because most of the variation in medical CT images occurs at the organ or tissue boundaries [12], [13]. For this type of sparsity the most common sparsifying transform is

$$Tf = |Df|_{\text{mag}},$$

where D is the finite differencing approximation to the image gradient, and $|\cdot|_{\rm mag}$ represents the spatial vector magnitude; i.e. Df is the gradient of f and $|Df|_{\rm mag}$ is the gradient magnitude image (GMI).

The measurement model \mathcal{M} is

$$g=Rf \text{ where } f\in F_{\text{GMI-sparse}}$$
 and
$$F_{\text{GMI-sparse}}=\{f\mid \|(|Df|_{\text{mag}})\|_0\leq s\},\quad (12)$$

where again the measurement model includes the restriction that $f \in F_{\text{GMI-sparse}}$. The posited inverse of this measurement model is defined implicitly with the optimization problem

$$f^* = \underset{f}{\operatorname{arg\,min}} \|(|Df|_{\operatorname{mag}})\|_1 \text{ such that } g = Rf. \tag{13}$$

where $\|(|Df|_{mag})\|_1$ is also known as the image total variation (TV). We refer to this optimization problem as equality constrained TV minimization (TVmin). Numerical investigations are carried out by generating simulation data according to Eq. (12) and numerically solving Eq. (13) to see if the test phantom is recovered accurately. The TVmin problem can be efficiently solved by the Chambolle-Pock primal-dual (CPPD) algorithm [14]–[16] explained in Appendix A with the pseudocode listed in Algorithm 1.




Fig. 2. Non-zero pixels, indicated in white, of the gradient magnitude images (GMI) corresponding to the phantoms shown in Fig. 1. The GMI sparsity is listed above each panel. For reference, the total number of pixels in the image array is $512^2 = 262,144$. Thus the corresponding percentage of GMI non-zero pixels is 3.7%, 3.7%, and 2.37% from left to right.

Accurate recovery of the test phantom provides evidence for injectivity of the measurement model, because it is unlikely that the original GMI-sparse test image would be recovered if there were other GMI-sparse images that yield the same data. Of course, recovery of the test image also provides evidence that TVmin inverts Eq. (12). Also, stability is tested to the extent that numerical error does not get amplified and interfere with accurate recovery of the test phantom.

IV. CNN-BASED DEEP-LEARNING FOR SPARSE-VIEW CT

We implement CNN-based deep-learning for image reconstruction for a 2D sparse-view problem, where a 512×512 image is reconstructed from a sinogram of dimension 128 views by 512 detector bins. In order to implement CNN image reconstruction, 4,000 simulation phantoms are generated from the stochastic models described in Sec. V. From each phantom, a 128-view sinogram is generated by use of Eq. (8). The 128-view sinograms are further processed by applying filtered back-projection (FBP). The resulting FBPreconstructed images, which have prominent streak artifacts due to the sparse-view sampling, are paired with the ground truth images and used to train the CNN. The data available to training, validating, and testing the CNN thus consists of 4,000 paired images, where each pair consists of the true phantom and its 128-view FBP reconstructed counterpart. The goal of the CNN training is to obtain a model that will yield the true phantom from the FBP reconstruction.

The CNN uses a standard U-net architecture as described in Figure 4a of Ref. [3]. The output of the U-net estimates the difference between the 128-view FBP images and its phantom counterpart. Given a 128-view FBP image, the CNN-predicted residual is subtracted from the FBP input to obtain the predicted reconstruction from the 128-view data.

For each dataset of 4000 pairs of images, 3990 are used for training and validation, and 10 images are left out for independent testing. Of the 3990 pairs, 80% or 3192 are used for training the residual U-net and 20% or 798 are used for validation. The models are trained by minimizing the mean-square-error (MSE) between the residual labels and predictions using a stochastic gradient descent algorithm (SGD) with momentum [17], learning rate decay and Xavier weight initialization [18]. Model training is done for 450 epochs using the MATLAB MatConvNet toolbox [19] and a NVIDIA GeForce GTX 1080 Ti graphics processing unit (GPU) for computation.




Fig. 3. Results for the CNN on reconstruction of a binary test phantom. Shown are FBP applied to the 128-view sinogram that serves as input to the CNN, the CNN reconstruction, and the corresponding difference images. The whole images appear in the top row and a corresponding ROI image is displayed in the bottom row. The gray scale windows are specified below the ROI panels in units of cm $^{-1}$. The RMSE for the CNN result is 1.15×10^{-3} cm $^{-1}$ and there is obvious discrepancy in the difference image where the shown grayscale window is 10% of the contrast between fat and fibroglandular tissue. The CNN image error is highly non-uniform and the maximum pixel error is $0.04~{\rm cm}^{-1}$, which is the same as the phantom tissue contrast.

For the present computations, the training took one week to complete.

One of the challenging aspects of analyzing deep-learning in the framework of inverse problems is that the measurement model itself is not completely specified. For the sparse-view CT setting, we do know that the data are generated by the discrete-to-discrete Radon transform, but the restriction on the images in the domain of the transform is not described mathematically. Instead there is the vague notion that the image domain is a set S that is somehow defined by the set of training data $F_{\rm train}$

$$g = Rf$$
 where $f \in S(F_{\text{train}})$. (14)

While it is not clear from the literature how to define S in the CT setting, S must include at least the training data if CNNs are truly capable of inverting Eq. (14). Thus, in the absence of a completely specified measurement model we can still check that a training image is recovered if the CNN is given the FBP reconstruction of data corresponding to that training image.

Because we do not know what the precise measurement model is for deep-learning, we employ a measurement model that is based on gradient-sparsity restriction. Knowledge of the measurement model is made available to the CNN by generating training and testing image/data pairs that adhere perfectly to the idealized gradient-sparsity restricted model. We test the deep-learning methodology on its ability to learn and invert this measurement model.

V. Breast Phantom Realizations

The breast CT slice phantom is designed to provide an illustrative test for both the gradient sparsity exploiting TVmin




Fig. 4. Results for TVmin on reconstruction of a binary test phantom. Shown are the TVmin reconstruction and the corresponding difference image in two different gray scale windows. The narrow gray scale range in the middle column is selected so that the difference from the truth can be seen, and the gray scale window in the right column is the same as that of the corresponding CNN result. The whole images appear in the top row and a corresponding ROI image is displayed in the bottom row. The gray scale windows are specified below the ROI panels in units of cm $^{-1}$. The RMSE for the TVmin result is $6.43\times10^{-8}~{\rm cm}^{-1}$ and the maximum deviation from the truth over all pixels is $7.11\times10^{-6}~{\rm cm}^{-1}$.

algorithm and CNN-based deep-learning. The phantom consists of modeled fat, fibroglandular, and skin tissues. The distribution of the fibroglandular tissue is generated by thresholding a power-law noise realization described in Ref. [20]. The phantom images are generated on a 512×512 pixel grid.

From the basic phantom realizations, two image classes are defined. The first class, called "binary", consist of the image realizations themselves, and the transition between fat and fibroglandular tissues is sharp. The second class, called "smooth-edge", is derived from the binary class by convolving with a gaussian kernel where the full width at half maximum (FWHM) is set to one pixel

$$f_{\text{smooth-edge}} = G(w_0) f_{\text{binary}}$$
 and $w_0 = 1$,

where G(w) symbolizes the operation of gaussian convolution with width parameter w. Three realizations of the "binary" and "smooth-edge" phantoms are shown in Fig. 1. Having the two classes of images allow us to test TVmin and the CNN under each of the two classes of images and under a generalized object model where the images can be selected from either class.

The binary phantom realizations have a sparse gradient magnitude image (GMI) as shown in Fig. 2, where the three shown realizations all indicate a high degree of GMI sparsity. The maximum number of GMI non-zeros over all 4000 phantom realizations is 12053, which corresponds to 4.58% of the whole image. Sparse GMI should allow for accurate image reconstruction by TVmin with significantly reduced sampling even though the tissue borders have a complex geometry.

The stochastic nature of the phantom makes it ideal for testing data-driven image reconstruction because the number of phantom realizations is only limited by storage space and




Fig. 5. Results for the CNN on reconstruction of a smooth-edge test phantom. The format of the figure is the same as that of Fig. 3. The RMSE for the CNN result is $6.76\times10^{-4}~\rm cm^{-1}$ and the maximum pixel error is $1.52\times10^{-2}~\rm cm^{-1}$.

computation time. Furthermore, the image truth is known exactly, eliminating error due to imperfect labeling of the training data.

VI. RESULTS

For the sparse-view CT problem of study, we generate parallel-beam projections of the binary or smooth-edge breast phantom, where $R=R_{128}$ is specified as a circular scan over 360 degrees with 128 projections. The linear detector array has length 18 cm and consists of 512 bins. The image array is also $18\times18~\rm cm^2$ consisting of 512x512 pixels, the same as the image array used for the breast phantom. With this sampling scheme, it is clear that the measurement model

$$g = R_{128}f$$

without any restriction on f is not injective because the size of g is 128×512 measurements and the number of unknowns pixel values is 512×512 . There are four times as many unknowns as knowns, and accordingly R has a non-trivial nullspace meaning that there can be two images that yield the same data on applying R_{128} .

For gradient-sparsity exploiting image reconstruction, the additional sparsity restriction on the image possibly allows the measurement model to be inverted. For deep-learning with CNNs, the training data are exploited to "learn" the measurement model which should include learning the subset S to which the object model images are restricted.

A. Image reconstruction with the binary class

The images, which are drawn from the binary breast phantom class, have GMI sparsity and consequently the data are generated from an instance of the measurement model in Eq. (12)

$$g = R_{128} f_{\text{obj}}$$
 where $f_{\text{obj}} \in F_{\text{binary}}$
and $F_{\text{binary}} = \{ f \mid \|(|Df|_{\text{mag}})\|_0 \le 12053 \},$ (15)




Fig. 6. Results for TVmin on reconstruction of a smooth-edge test phantom. The format of the figure is the same as that of Fig. 4. The RMSE for the TVmin result is $1.15\times 10^{-6}~{\rm cm}^{-1}$ and the maximum deviation from the truth over all pixels is $7.64\times 10^{-5}~{\rm cm}^{-1}$.

where 12053 is the largest number of GMI non-zeros in the 4000 binary phantom image realizations. To demonstrate image reconstruction from the binary phantom, TVmin and the deep-learning CNN are applied to the 128-view sinogram generated from the second phantom shown in Fig. 1.

The CNN results are shown in Fig. 3, where the input FBP, reconstructed, and difference images are shown. The top row shows the full reconstructed image array, and the bottom row shows a corresponding region of interest (ROI). While the FBP image looks as if it is generated from noisy data, it is actually not. The complexity of the edge structure and the severe angular undersampling causes the wavy and noisy appearance.

On the global and ROI views the CNN prediction images appear to be accurate in the shown gray scale window. Some discrepancies are visible in the ROI image. The difference images, however, reveal substantial discrepancy between the CNN reconstruction and the true phantom image in a gray scale window that is 10% of the fat/fibroglandular contrast. Furthermore, the maximum pixel error is at the level of this soft-tissue contrast. As a result, the claim that the CNN provides a numerically accurate reconstruction and numerically solves Eq. (15) can not be made.

The corresponding results for TVmin are shown in Fig. 4, where it is clear that the numerical accuracy of TVmin is much greater than that of the CNN. The image RMSE is five orders of magnitude smaller than the soft-tissue contrast, and the maximum pixel error is three orders of magnitude smaller than this contrast level. As a result, there is no perceptible difference between the reconstructed image and true phantom for any ROI in the image. To emphasize the difference in the accuracy of the recovery, the last column of Fig. 4 displays the TVmin difference images in the gray scale used for the CNN difference images.

The conclusion from this particular experiment is that TVmin is able to numerically recover the test phantom. This result is not too surprising because the number of non-zero




Fig. 7. The left and right sets of results follow the same format as Figs. 3 and 5, respectively. The respective CNNs are applied to images used in the training. For the binary training image on the left, the RMSE is 7.90×10^{-4} cm⁻¹ and the maximum pixel error is 0.04 cm⁻¹. For the smooth-edge training image on the right, the RMSE is 4.66×10^{-4} cm⁻¹ and the maximum pixel error is 0.011 cm⁻¹.

pixels in the GMI of the phantom is 9720 pixels and the total number of samples is $128 \times 512 = 65536$ or approximately 6.5 times the GMI sparsity. The CNN, on the other hand, does not provide a numerically accurate inverse. Furthermore, the CNN is not able to recover the test phantom in a situation where we *know* there exists an accurate numerical inverse as provided by TVmin.

B. Image reconstruction with the smooth-edge class

The smooth-edge breast phantom class images have GMI sparsity in the object model and the corresponding restricted measurement model is

$$g = R_{128} f_{\text{obj}} \text{ where } f_{\text{obj}} \in F_{\text{smooth-edge}}$$

$$\text{and } F_{\text{smooth-edge}} = \{ G(w_0) f \mid \|(|Df|_{\text{mag}})\|_0 \le 12053 \},$$

$$\tag{16}$$

and $w_0=1$ in pixel units. In terms of methodology, there is no change in implementing the training of the CNN for this class of images other than switching to the set of smoothedge training pairs. To perform sparsity-exploiting image reconstruction for this measurement model, the optimization problem in Eq. (13) is altered slightly in the data equality constraint, becoming

$$f^* = \underset{f}{\arg\min} \|(|Df|_{\max})\|_1 \text{ such that } g = R_{128}G(w_0)f.$$
 (17)

The reconstructed image is obtained from f^* by applying the Gaussian blurring

$$f_{\text{recon}} = G(w_0) f^*$$
.

Generalizing the implementation of TVmin to account for the blur is straight-forward [21], [22], and the necessary changes are discussed in Appendix A.

The results for image reconstruction by the CNN and TVmin are shown in Figs. 5 and 6, respectively. Again, both reconstructed images are visually accurate on the global view and in the chosen ROI. For the CNN results, the difference

images reveal a wider band of discrepancy at the transition between fat and fibroglandular tissue than was the case for the binary class study. There is a corresponding distortion in the shapes of the fibroglandular regions that is visible in the ROI. The numerical results are similar to that of the binary class study. There is some numerical differences in that the smooth-edge RMSE values are lower than that of the binary phantom case, but they are still far from being able to claim numerically accurate image reconstruction.

The RMSE values for TVmin are larger than the corresponding results with the binary phantom, but the numerical error is still orders of magnitude below the soft-tissue contrast level of the smooth-edge phantom. The conclusions from this study are similar to that of the binary phantom; namely, the CNN images do not show accurate numerical recovery of the test phantom while the TVmin images do.

C. Applying the CNN to training images

As noted in Sec. IV, the training images must belong to the restricted set of images that are part of the learned measurement model. Testing the CNN with data from a training image reveals the ability of the CNN to approximate the transform from the FBP image to the reconstructed image. We perform this test with a training image from the binary phantom class. The results are shown on the left set of six panels of Fig. 7. Likewise, we show the result for the smooth-edge phantom class on the right set of six panels. In both cases, we note accurate visual recovery, but the image reconstruction is not numerically accurate. From an inverse problem perspective, the CNN must be able to return the true training image from the corresponding training data in order to be able to claim that it inverts the sparse-view CT measurement model.

D. Expanding the object model

In the next set of results, we investigate the impact of generalizing the object model by combining the binary and smooth-




Fig. 8. The left and right sets of results follow the same format as Figs. 3 and 5, respectively. The difference here is that both sets of results are generated with the same CNN, trained with a mixture of binary and smooth-edge images and tested on an independent binary image. The RMSE for this CNN applied to a binary test image is 1.24×10^{-3} cm⁻¹ and the maximum pixel error is 0.041 cm⁻¹. The RMSE for this CNN applied to a smooth-edge test image is 7.60×10^{-4} cm⁻¹ and the maximum pixel error is 0.0184 cm⁻¹.

edge phantom classes. The corresponding measurement model becomes

$$g = R_{128} f_{\text{obj}}$$
 where $f_{\text{obj}} \in F_{\text{binary}} \cup F_{\text{smooth-edge}}$. (18)

For the CNN, the training and testing set is chosen again to be 4000 image/data pairs, where 2000 pairs are selected from the binary and smooth-edge phantoms. In this way, the training effort is commensurate with the previous CNN results. The sparsity exploiting reconstruction in this case requires a double optimization, TVmin over f

$$f^{\star}(w) = \underset{f}{\arg\min} \|(|Df|_{\text{mag}})\|_{1} \text{ such that } g = R_{128}G(w)f,$$
(19)

followed by GMI sparsity optimization over w

$$w^* = \underset{w}{\arg\min} \|(|Df^*(w)|_{\max})\|_0,$$
 (20)

where the ℓ_0 -norm can be efficiently optimized over w. Because it is a scalar argument, the minimum can be found by bracketing and bisection. The reconstructed image is obtained from f^* and w^* by computing

$$f_{\text{recon}} = G(w^*) f^*(w^*).$$

Application of TVmin in this setting does require more effort, depending on how many function evaluations are needed in the second optimization. The minimizer in the second optimization can be accurately determined in ten to twenty iterations.

We apply the trained CNN to a binary and smooth-edge phantom in Fig. 8. Note that in this case it is the same network that yields both sets of results, while previously, the networks are trained with data from the corresponding object classes. The RMSE and maximum pixel error metrics are at a level that is similar to the previous results where all 4000 training/testing pairs came from the same class. Thus, at least visually, the accuracy of the reconstruction seems to not be compromised, supporting the potential for object model generalization. As the error numbers and difference

images show, however, numerically accurate inversion of the measurement model is not obtained.

For the generalized TVmin results, we select a pair of phantoms from the binary and smooth-edge class; hence the true value of w are 0 and 1, respectively. The GMI sparsity optimization over w is shown in Fig. 9. To perform the double optimization, the TVmin algorithm iteration is halted when the data RMSE reaches 1×10^{-6} so that the estimate of $f^*(w)$ achieves similar levels of convergence as a function of w. The number of "non-zeros" in the GMI is computed by using a threshold of 0.1% of the maximum value in the GMI. The resulting number of GMI pixels above this threshold is plotted in the top panel of Fig. 9. The minimum number of nonzeros occurs at the value w=0 and w=1 which coincides with the truth values for the binary and smooth-edge images, respectively. Having recovered the true w, the actual image is obtained from TVmin using w = 0 and w = 1 and the result is the same as what is shown in Figs. 4 and 6, respectively.

It is also instructive to plot the image RMSE between the phantoms and the image estimates, $G(w)f^*(w)$, for the whole range of w that is searched. The image RMSE is minimized at the corresponding w-values that minimize the number of GMI non-zeros. We do note, however, that the image RMSE is still quite low for other values of w; in fact all the plotted values are less than the image RMSE obtained with the CNN. These results raise the interesting question on how to determine whether or not other values of w can still lead to recovery of the test phantom. This issue is taken up in Appendix B.

E. Numerical evidence for solving the inverse problem

Numerically, it is not possible to show mathematical inversion in this setting because there is always some level of numerical error. We have implicitly been appealing to the image RMSE or maximum pixel error being small compared to the inherent contrast of the test phantom. That the maximum pixel error for the TVmin result is several orders of magnitude lower than the phantom contrast implies that there is no ROI where




Fig. 9. GMI sparsity optimization of generalized TVmin over w. The top panel shows the GMI sparsity $\|(|Df^*(w)|_{\text{mag}})\|_0$ as a function of w, where a threshold of 0.1% of the GMI maximum is used to distinguish non-zeros from numerically small values. The bottom row plots the image RMSE between the smooth-edge phantom and the TVmin image estimate $G(w)f^*(w)$ as a function of w.

the difference between the true and reconstructed phantom is apparent in a gray scale matched to the phantom contrast. Further discussion of TVmin convergence and sparsity model inversion is presented in Appendix B.

For the CNN, we have not managed to achieve an accuracy level where the image error is imperceptible for a gray scale matched to the phantom contrast level. To illustrate this point clearly, we have searched all 24×24 pixel ROIs in the ten testing images and selected the one with the greatest image RMSE for both the binary and smooth-edge results presented in Secs. VI-A and VI-B, respectively. The resulting ROIs are displayed in Fig. 10. On the one hand, it is a remarkable achievement of the methodology that this is the worst ROI over ten 512×512 pixel testing images. On the other hand, there is obvious discrepancy and it is difficult to take this result as evidence for solving a sparse-view CT inverse problem.

It may be possible that there are methods to improve the CNN accuracy. One could imagine increasing the training set size, using transfer learning, or an entirely different deep-learning network architecture. As far as we know, however, there is no specific quantitative methodology in the literature for estimating how much additional training, or what modifications to the deep-learning, are needed to achieve a desired accuracy goal.

VII. CONCLUSION

We have presented a focused study on solving the sparse-view CT inverse problem using GMI sparsity-exploiting image reconstruction and deep-learning with CNNs. The numerical evidence for solving sparse-view CT between TVmin and the CNN is contrasted. The TVmin results have an error many orders of magnitude smaller than that of the CNN results. In an absolute sense the scale of the TVmin error is demonstrated




Fig. 10. ROI images for the 24×24 ROI with the greatest image RMSE over all ten testing images. The top and bottom rows show the ROI images for the CNN reconstructed binary and smooth-edge phantoms. There is visible discrepancy between the phantom and CNN reconstructed ROIs in a gray scale appropriate for the fat/fibroglandular contrast. Most notably there are structures, indicated by the red arrow, that are not present in the phantom ROIs. Note that the gray scale window used for the difference image is ten times wider than the gray scale windows used for the previous CNN difference images.

to be much smaller than the inherent contrast levels of the relevant CT application, while this is not the case for the CNN.

For TVmin there is a general inverse problem framework for sparsity exploiting image reconstruction that underpins the algorithm so that inverse problem results are testable and repeatable. The CNN results show some level of accuracy, but it is not clear what it takes, nor that it is even possible, to drive the image RMSE arbitrarily close to zero. We have done our best to implement the residual U-net methodology described in Refs. [1], [3] but could not repeat the claim that the CT inverse problem is solved. There remain questions: What exactly is the measurement model that is being inverted? How much training data is needed for a given application and system configuration? How can the loss function be optimized so that there is arbitrarily small error on recovering the training images? What does it take to achieve arbitrarily small error on the testing image? How exactly should the network architecture be designed to achieve the inverse problem solution? More generally, what are all the parameters involved in developing the CNN for inverse problems and how are they determined?

We point out that this negative result for the CNN methodology as an inverse problem solver employs a much larger set of training images with perfect knowledge of the truth than would be available from actual CT scans. Furthermore, the training set is generated respecting gradient sparsity where we know there exists an inverse for the 128-view CT problem as demonstrated by the TVmin algorithm. In fact, the CNN is tested on the class of images generated by our breast model, which is a subset of GMI-sparse images.

No other study in the literature has presented convincing evidence that CNNs solve a CT related inverse problem. Because this has not been demonstrated, the generalizability, conferred by inverse problem results on injectivity and stability, cannot be applied to CNNs in the CT image reconstruction setting.

We also point out that similar claims on using CNNs to invert full-data CT measurement models or other inverse problems in imaging are also drawn into question by their lack of supporting evidence.

This result, however, does not mean that CNNs cannot be used for applications in CT imaging. It only means that it is paramount to recognize and embrace the fact that there is a high degree of variability inherent to deep-learning and CNN methodology and, commensurately, a high degree of variability in the possible qualities of the CNN processed images. We must recognize and take seriously that the output of the CNN is a function of the training set, the training methodology, network structure, and properties of the input data. Each of these broad areas needs to be precisely specified if CNN research in the CT setting is to be repeatable.

More importantly, the images obtained from the trained CNNs must be evaluated and demonstrated in a task-based fashion [23]. Given the high degree of CNN variability and lack of mathematical characterization through inverse problem results on injectivity and stability, it is not clear what is the scope of applicability of a particular trained CNN. A small change in scan configuration or object properties may or may not compromise the utility of the CNN for CT imaging tasks. Thus, in evaluating a particular CNN, the purpose of the scan must be precisely defined and evaluated with quantitative, objective metrics relevant to that purpose.

APPENDIX

A. Primal-dual algorithm for solving TVmin

Algorithm 1 Pseudocode for the CPPD-TVmin updates. The convergence criteria are explained in the text of Appendix A, where the additional computation of splitting variables y_s and y_g is needed, as specified by Eq. (27).

1:
$$f^{(k+1)} \leftarrow f^{(k)} - \tau \left(\nu_s R^\top \lambda_s^{(k)} + \nu_g D^\top \lambda_g^{(k)} \right)$$

2: $\bar{f} \leftarrow 2f^{(k+1)} - f^{(k)}$
3: $\lambda_s^{(k+1)} \leftarrow \lambda_s^{(k)} + \sigma(\nu_s R\bar{f} - \nu_s g)$
4: $\lambda_g^+ \leftarrow \lambda_g^{(k)} + \sigma \nu_g D\bar{f}$
5: $\lambda_g^{(k+1)} \leftarrow \lambda_g^+ / \max\left(1, \left|\lambda_g^+\right|_{\text{mag}}\right)$

The Chambolle-Pock primal dual (CPPD) algorithm for solving the TVmin problem

$$f^{\star} = \underset{f}{\arg\min} \, \|(|Df|_{\text{mag}})\|_1 \text{ such that } g = Rf, \qquad (21)$$

is presented in Algorithm 1. The same algorithm can be straight-forwardly generalized to solve Eq. (19) where the equality constraint is replaced by g = RG(w)f. To make the generalization, R can be replaced by R(w)

$$R(w) = RG(w)$$
 and $R^{\top}(w) = G(w)R^{\top}$,

where we have used the fact that G(w) is symmetric, i.e. $G^{\top}(w) = G(w)$.

The parameters and notation used in Algorithm 1 are explained here. The parameter ρ is the ratio of the primal

and dual update step-size parameters, τ and σ , respectively. The integer k denotes the iteration index. The parameters ν_s and ν_q normalize the linear transforms R and D

$$\nu_s = 1/\|R\|_2, \ \nu_q = 1/\|D\|_2,$$

where the ℓ_2 -norm of a matrix is its largest singular value. The scaling is performed so that algorithm efficiency is optimized and so that results are independent of the physical units used in implementing R and D. Note that the data g must also be multiplied by ν_s . The step-size parameters σ and τ are

$$\sigma = \rho/L, \ \tau = 1/(\rho L),$$

where

$$A = \begin{pmatrix} \nu_s R \\ \nu_g D \end{pmatrix} \ L = \|A\|_2,$$

and the matrix A is constructed by stacking $\nu_s R$ on $\nu_g D$. Due to the normalization of R and D, and the fact that R and D approximately commute, L should be close to 1. The symbol $|\cdot|_{\rm mag}$ acts on the spatial vector at each pixel, yielding the spatially dependent vector magnitude. For example, if f is an image, Df is the gradient of f and $|Df|_{\rm mag}$ is the gradient-magnitude image (GMI). The function "max" acts component-wise on the vector argument.

The only free algorithm parameters are the step-size ratio ρ and the total number of iterations K. The step-size ratio must be tuned for algorithm efficiency, and clearly larger K leads to greater solution accuracy. For the results presented in the main text, we set $\rho = 2 \times 10^4$ and K = 5000 iterations.

CPPD-TVmin theory and convergence metrics: We present a summary of the primal, saddle, and dual optimization inherent to the CPPD framework in order to explain the origin of the convergence criteria. The CPPD framework is based on the use of splitting, where Eq. (21) is modified to

$$f^\star=\mathop{\arg\min}_{f,y_s,y_g}\|(|y_g|_{\rm mag})\|_1$$
 such that
$$g=y_s,\, Df=y_g,\, Rf=y_s. \quad \mbox{(22)}$$

The introduction of the splitting variables y_g and y_s simplifies the potentials and constraints. We drop the use of the normalization factors ν_g and ν_s to simplify the presentation. This equality constrained optimization is converted to a saddle point problem by introducing Langrange multipliers for the splitting equality-constraints

$$\max_{\lambda_s, \lambda_g} \min_{f, y_s, y_g} \|(|y_g|_{\text{mag}})\|_1 + \lambda_s^\top (Rf - y_s) + \lambda_g^\top (Df - y_g)$$
such that $g = y_s$, (23)

where λ_s and λ_g are the dual variables, a.k.a. Lagrange multipliers, for X-ray projection and the image gradient, respectively. This saddle point problem can be simplified by performing the minimization over the splitting variables y_s and y_a analytically

$$\max_{\lambda_s, \lambda_g} \min_{f} \lambda_s^{\top} (Rf - g) + \lambda_g^{\top} Df \text{ such that } \|(|\lambda_g|_{\text{mag}})\|_{\infty} \le 1,$$
(24)

where $||v||_{\infty} = \max_{i} |v_{i}|$ yields the magnitude of the largest component of the argument vector. It is this saddle point

problem that is solved with the CPPD-TVmin algorithm. The minimization over f in Eq. (24) can also be performed analytically, yielding the dual maximization problem to TVmin

$$\max_{\lambda_s, \lambda_g} -\lambda_s^{\top} g \text{ such that } \|(|\lambda_g|_{\text{mag}})\|_{\infty} \le 1, \ D^{\top} \lambda_g + R^{\top} \lambda_s = 0.$$
(25)

It turns out that the transversality constraint

$$D^{\top} \lambda_q + R^{\top} \lambda_s = 0$$

of the dual maximization in Eq. (25) and the splitting equalities

$$y_s = Rf, \ y_q = Df, \tag{26}$$

in Eq. (22) provide a complete set of first order optimality checks.

In the CPPD framework, the splitting variable iterates can be computed from the iterates of the primal and dual variables by adding the following two lines to the pseudocode in Algorithm 1

$$y_g^{(k+1)} = \frac{1}{\sigma} \left(\lambda_g^{(k)} - \lambda_g^{(k+1)} \right) + D\bar{f},$$

$$y_s^{(k+1)} = \frac{1}{\sigma} \left(\lambda_s^{(k)} - \lambda_s^{(k+1)} \right) + R\bar{f}.$$
(27)

A complete set of first-order convergence conditions are given by the splitting gap

$$\sqrt{\left\|y_s^{(k)} - Rf^{(k)}\right\|_2^2 + \left\|y_g^{(k)} - Df^{(k)}\right\|_2^2} \to 0, \quad (28)$$

which clearly derives from Eq. (26), and the transversality condition [24]

$$\left\| R^{\mathsf{T}} \lambda_s^{(k)} + D^{\mathsf{T}} \lambda_g^{(k)} \right\|_2 \to 0. \tag{29}$$

The splitting gap is equivalent to what is called the dual residual in Ref. [25]. These convergence criteria are particularly useful for determining the step-size ratio parameter ρ , which must be tuned for algorithm efficiency. Increasing ρ , increases σ and decreases τ , which has the effect of slowing progress on decreasing the splitting gap while improving progress toward transversality. Decreasing ρ has the opposite effect.

B. TVmin verification and measurement model inversion

The TVmin algorithm results of Sec. VI-D provides an opportunity to explain further the numerical evidence for algorithm verification and measurement model inversion. We consider the case where the test object image comes from the smooth-edge class, and we attempt to recover the test image using Eq. (19) and

$$f_{\text{recon}}(w) = G(w)f^{\star}(w),$$

for w=0 and w=1. The results for the image RMSE for these two w values can be read from the green curve in the bottom panel of Fig. 9. The image RMSE for w=0 is higher than that of the true value at w=1. Yet, both image RMSEs are still small. We explore image reconstruction for these two w values further to see if the corresponding measurement models can be inverted. The measurement model for both cases is

$$g = R_{128} f_{\text{obj}}$$
 where $f_{\text{obj}} \in F_{\text{smooth-edge}}$.




Fig. 11. Plot of convergence metrics for $f_{\text{recon}}(w)$ when the data are generated from projection of a smooth-edge phantom. Shown are the image and data RMSE along with the transversality and splitting gap convergence criteria. The latter two quantities are normalized to 1 for their value at the first iteration.

102

10

iterations

10⁴

The proposed model inverse for the w=0 case reduces to

$$f_{\mathrm{recon}}(0) = \mathop{\arg\min}_{f} \|(|Df|_{\mathrm{mag}})\|_1 \text{ such that } g = R_{128}f,$$

and the model inverse for the w=1 case is the same as Eq. (17)

$$f^* = \underset{f}{\arg\min} \|(|Df|_{\max})\|_1 \text{ such that } g = R_{128}G(w_0)f$$
$$f_{\text{recon}}(1) = G(w_0)f^*,$$

where $w = w_0 = 1$.

From a Compressed Sensing perspective we do not expect $f_{\rm recon}(0)$ to invert the measurement model because it does not exploit the GMI sparsity optimally. The number of nonzeros in the GMI of the smooth-edge phantom is 57050 and the number of samples in the 128×512 sinogram is 65536. From empirical studies performed in Ref. [11] the number of samples should be at least twice the number GMI non-zeros for this sparsity regime and the ratio of samples to GMI non-zeros in this case is 1.15. On the other hand, we do expect $f_{\rm recon}(1)$ to invert the measurement model because it does exploit GMI sparsity optimally as demonstrated in Sec. VI-B.

For this study we extend the iteration number in computing $f_{\rm recon}(w)$ to 100000 and change the step-size ratio to $\rho=2\times10^5$. The resulting image and data RMSE plots are shown in Fig. 11 along with the transversality and splitting gap convergence criteria.

The splitting gap and transversality provide the algorithm verification metrics, and both of these curves show a downward trend verifying that the CPPD-TVmin algorithm is solving the optimization problems associated with $f_{\rm recon}(0)$ and $f_{\rm recon}(1)$. The data RMSE is a redundant algorithm verification metric in this case where the data are ideal, but we show it also to provide reference for the image RMSE.

It is the image RMSE that indicates whether or not the measurement model is inverted. The image RMSE curves in Fig. 11 show a plateau for $f_{\rm recon}(0)$ and a decreasing trend for $f_{\rm recon}(1)$, indicating that measurement model inversion is not attained for the former case but is attained for the latter. Interpreting the image RMSE trend does need context, however, from the data RMSE. Inspecting the numerical values of the image RMSE for $f_{\rm recon}(0)$, they are actually decreasing over the 100000 iterations, and considering image RMSE alone it can be difficult to distinguish slow algorithm convergence to zero from convergence to a small positive value, i.e. plateauing.

When plotting the image and data RMSE together, the picture is clearer. For the $f_{\rm recon}(0)$ case there is a clear divergence between the image and data RMSE curves, ending at a nearly four order of magnitude gap for 100000 iterations. Accordingly, the most reasonable extrapolation for this case is that the image RMSE limits to a small positive number and does not tend to zero. For $f_{\rm recon}(1)$, on the other hand, the image and data RMSE track together, and the most reasonable extrapolation is that they will continue to tend to zero as the iteration continues. We acknowledge that these conclusions involve extrapolation of the presented results because it is only possible to execute a finite number of iterations and there is limitations imposed by the floating point precision.

REFERENCES

- K. H. Jin, M. T. McCann, E. Froustey, and M. Unser, "Deep convolutional neural network for inverse problems in imaging," *IEEE Trans. Image Proc.*, vol. 26, pp. 4509–4522, 2017.
- [2] B. Zhu, J. Z. Liu, S. F. Cauley, B. R. Rosen, and M. S. Rosen, "Image reconstruction by domain-transform manifold learning," *Nature*, vol. 555, pp. 487–492, 2018.
- [3] Y. Han and J. C. Ye, "Framing U-Net via deep convolutional framelets: Application to sparse-view CT," *IEEE Trans. Med. Imag.*, vol. 37, pp. 1418–1429, 2018.
- [4] J. M. Boone, T. R. Nelson, K. K. Lindfors, and J. A. Seibert, "Dedicated breast CT: radiation dose and image quality evaluation," *Radiology*, vol. 221, pp. 657–667, 2001.
- [5] G. Bal, "Introduction to inverse problems," 2012, URL: https://statistics.uchicago.edu/~guillaumebal/PAPERS/IntroductionInverseProblems.pdf.
- [6] J. S. Jørgensen, E. Y. Sidky, and X. Pan, "Quantifying admissible undersampling for sparsity-exploiting iterative image reconstruction in X-ray CT," *IEEE Trans. Med. Imag.*, vol. 32, pp. 460–473, 2012.
- [7] E. J. Candès, J. Romberg, and T. Tao, "Robust uncertainty principles: Exact signal reconstruction from highly incomplete frequency information," *IEEE Trans. Inf. Theory*, vol. 52, pp. 489–509, 2006.
- [8] D. L. Donoho, "Compressed sensing," *IEEE Trans. Inf. Theory*, vol. 52, pp. 1289–1306, 2006.
- [9] E. J. Candès and M. B. Wakin, "An introduction to compressive sampling," *IEEE Sig. Proc. Mag.*, vol. 25, pp. 21–30, 2008.
- [10] D. L. Donoho and J. Tanner, "Sparse nonnegative solution of underdetermined linear equations by linear programming," *Proc. Nat. Acad.* Sci., vol. 102, pp. 9446–9451, 2005.
- [11] J. S. Jørgensen and E. Y. Sidky, "How little data is enough? Phase-diagram analysis of sparsity-regularized X-ray computed tomography," Philos. Trans. Royal Soc. A, vol. 373, pp. 20140387, 2015.
- [12] E. Y. Sidky, C.-M. Kao, and X. Pan, "Accurate image reconstruction from few-views and limited angle data in divergent-beam CT," J. X-ray Sci. Technol., vol. 14, pp. 119–39, 2006.

- [13] E. Y. Sidky and X. Pan, "Image reconstruction in circular cone-beam computed tomography by constrained, total-variation minimization," *Phys. Med. Biol.*, vol. 53, pp. 4777–4807, 2008.
- Phys. Med. Biol., vol. 53, pp. 4777–4807, 2008.
 [14] A. Chambolle and T. Pock, "A first-order primal-dual algorithm for convex problems with applications to imaging," J. Math. Imag. Vis., vol. 40, pp. 120–145, 2011.
- [15] T. Pock and A. Chambolle, "Diagonal preconditioning for first order primal-dual algorithms in convex optimization," in *International Con*ference on Computer Vision (ICCV 2011), 2011, pp. 1762–1769.
- [16] E. Y. Sidky, J. H. Jørgensen, and X. Pan, "Convex optimization problem prototyping for image reconstruction in computed tomography with the Chambolle-Pock algorithm," *Phys. Med. Biol.*, vol. 57, pp. 3065–3091, 2012.
- [17] D. E. Rumelhart, G. E. Hinton, and R. J. Williams, "Learning representations by back-propagating errors," *Nature*, vol. 323, pp. 533– 536, 1986.
- [18] X. Glorot and Y. Bengio, "Understanding the difficulty of training deep feedforward neural networks," in *Proceedings of the thirteenth* international conference on artificial intelligence and statistics, 2010, pp. 249–256.
- [19] A. Vedaldi and K. Lenc, "Matconvnet: Convolutional neural networks for matlab," in *Proceedings of the 23rd ACM international conference* on Multimedia, 2015, pp. 689–692.
- [20] I. Reiser and R. M. Nishikawa, "Task-based assessment of breast tomosynthesis: effect of acquisition parameters and quantum noise," *Med. Phys.*, vol. 37, pp. 1591–1600, 2010.
- [21] P. A. Wolf, T. G. Schmidt, J. S. Jørgensen, and E. Y. Sidky, "Few-view single photon emission computed tomography (SPECT) reconstruction based on a blurred piecewise constant object model," *Phys. Med. Biol.*, vol. 58, pp. 5629–5652, 2013.
- [22] Z. Zhang, S. Rose, J. Ye, A. E. Perkins, B. Chen, C.-M. Kao, E. Y. Sidky, C.-H. Tung, and X. Pan, "Optimization-based image reconstruction from low-count, list-mode TOF-PET data," *IEEE Trans. Biomed. Eng.*, vol. 65, pp. 936–946, 2018.
- [23] H. H. Barrett and K. J. Myers, Foundations of image science, Wiley, Hoboken, New Jersey, 2004.
- [24] J.-B. Hiriart-Urruty and C. Lemaréchal, Convex analysis and minimization algorithms I, Springer, Berlin, Germany, 1993.
- [25] T. Goldstein, M. Li, X. Yuan, E. Esser, and R. Baraniuk, "Adaptive primal-dual hybrid gradient methods for saddle-point problems," arXiv preprint arXiv:1305.0546, 2013.