

# Noise and dose reduction in CT brain perfusion acquisition by projecting time attenuation curves onto lower dimensional spaces

Vojtěch Kulvait<sup>a</sup>, Philip Hoelter<sup>b</sup>, Arnd Doerfler<sup>b</sup>, and Georg Rose<sup>a</sup>

<sup>a</sup>Institute for Medical Engineering and Research Campus STIMULATE, University of Magdeburg, Magdeburg, Germany

<sup>b</sup>Department of Neuroradiology, University Hospital Erlangen, Friedrich-Alexander-Universität (FAU) Erlangen-Nürnberg, Erlangen, Germany

## ABSTRACT

CT perfusion imaging (CTP) plays an important role in decision making for the treatment of acute ischemic stroke with large vessel occlusion. Since the CT perfusion scan time is approximately one minute, the patient is exposed to a non-negligible dose of ionizing radiation. However, further dose reduction increases the level of noise in the data and the resulting perfusion maps. We present a method for reducing noise in perfusion data based on dimension reduction of time attenuation curves. For dimension reduction, we use either the fit of the first five terms of the trigonometric polynomial or the first five terms of the SVD decomposition of the time attenuation profiles. CTP data from four patients with large vessel occlusion and three control subjects were studied. To compare the noise level in the perfusion maps, we use the wavelet estimation of the noise standard deviation implemented in the scikit-image package. We show that both methods significantly reduce noise in the data while preserving important information about the perfusion deficits. These methods can be used to further reduce the dose in CT perfusion protocols or in perfusion studies using C-arm CT, which are burdened by high noise levels.

**Keywords:** CT perfusion imaging, dose reduction, noise reduction, dimension reduction, SVD decomposition

## 1. INTRODUCTION

CT perfusion imaging (CTP) helps select population of patients with acute ischemic stroke and large vessel occlusion within 6 to 24 hours from the symptoms onset who are suitable for mechanical thrombectomy.<sup>1,2</sup> A typical perfusion scan takes approximately one minute and involves a significant dose of radiation for the patient. Low dose protocols cause additional noise in the data and resulting perfusion maps. Dimension reduction of the temporal attenuation profiles is a tool that can reduce noise in the data, as has been shown in the case of C-arm CT perfusion studies.<sup>3,4</sup> In this paper, we apply dimension reduction techniques to CT perfusion data to decrease noise with applications to radiation dose reduction.

## 2. METHODS

CT perfusion data from 4 acute ischemic stroke patients and 3 control subjects were used. Time attenuation curves (TACs) were obtained from individual CTP acquisitions. For noise reduction we used two different dimension reduction techniques. In the first approach, we fit the first 5 terms of the trigonometric polynomial including the constant to the TACs. In the second approach, we first perform an SVD decomposition of the TACs, see Fig. 1. We then fit the 5 profiles corresponding to the highest singular values to the data. A Gaussian blurr with  $\sigma = 2$  px was added into the original volume data and the reduced-dimension data sets and these data were processed using a standard deconvolution-based perfusion processing method.<sup>5</sup> Perfusion maps of CBF, CBV, MTT and TTP were generated. To compare the noise levels in the perfusion maps, we use a wavelet-based estimator `skimage.restoration.estimate_sigma` of the noise standard deviation implemented in the scikit-image package.

---

Send correspondence to: kulvait@gmail.com




Figure 1. Example SVD basis derived from the patient data. In the legend, there is a singular value corresponding to each temporal vector.

### 3. RESULTS

Tables 1 and 2 summarize Pearson correlation coefficients when comparing data with and without dimension reduction. Table 1 contains the correlation coefficients for dimension reduction using trigonometric functions and table 2 contains the correlation coefficients for dimension reduction using SVD decomposition. Figure 2 shows the differences between data processed with and without dimension reduction on the example of a patient with large vessel occlusion. Table 3 summarizes the estimates of the standard deviation of the noise using the `skimage.restoration.estimate_sigma` function of the scikit-image package. Paired T-test showed a significant reduction in noise level with  $p < 0.01$  values for all perfusion parameters and for both dimension reduction methods.

	Patient				Control		
	1A	2A	3A	4A	C1	C2	C3
CBF	0.96	0.92	0.96	0.97	0.96	0.97	0.99
CBV	0.94	0.90	0.92	0.92	0.93	0.95	0.97
MTT	0.81	0.78	0.79	0.78	0.76	0.73	0.74
TTP	0.58	0.59	0.59	0.58	0.56	0.39	0.49

Table 1. Pearson correlation coefficients comparing data without dimension reduction with data of reduced dimension using first five terms of trigonometric polynomial.

	Patient				Control		
	1A	2A	3A	4A	C1	C2	C3
CBF	0.97	0.97	0.95	0.95	0.96	0.98	0.98
CBV	0.96	0.96	0.92	0.92	0.96	0.97	0.97
MTT	0.82	0.79	0.73	0.75	0.75	0.82	0.74
TTP	0.59	0.70	0.60	0.55	0.59	0.49	0.51

Table 2. Pearson correlation coefficients comparing data without dimension reduction with data of reduced dimension using first five terms of SVD decomposition.

	No dimension reduction	Trigonometric functions	SVD decomposition
CBF	$0.21 \pm 0.01$	$0.12 \pm 0.01$	$0.14 \pm 0.01$
CBV	$0.033 \pm 0.001$	$0.016 \pm 0.001$	$0.021 \pm 0.001$
MTT	$0.19 \pm 0.01$	$0.13 \pm 0.01$	$0.10 \pm 0.01$
TTP	$0.99 \pm 0.07$	$0.31 \pm 0.01$	$0.30 \pm 0.01$

Table 3. Noise standard deviation estimates using the `skimage.restoration.estimate_sigma` function of the scikit-image package for data without dimension reduction and data with dimension reduction using trigonometric functions and SVD decomposition.

#### 4. CONCLUSIONS

Table 3 shows that both dimension reduction methods significantly reduce the noise in the perfusion maps. Dimension reduction using trigonometric polynomial and SVD decomposition had very comparable effects on the data used, see tables 1 to 3. The image data, see Figure 2, show that the noise reduction using both dimension reduction methods is most significant for the CBV map. The data suggest that SVD reduction can stabilize the TTP and MTT maps better.

Both dimension reduction methods preserve well the essential patterns of perfusion deficits in perfusion maps as seen, for example, in Figure 2. Our proposed method of noise reduction in CT perfusion data can be used to




Figure 2. Comparison of the original data (A) with the data of reduced dimension using trigonometric functions (B) or SVD decomposition of temporal profiles (C).

further reduce the dose in CT perfusion acquisitions. We will also investigate these dimension reduction methods to process interventional perfusion data from C-arm CT systems.

## ACKNOWLEDGMENTS

The work of this paper is partly funded by the Federal Ministry of Education and Research within the Research Campus STIMULATE under the number 13GW0473A.

## REFERENCES

- [1] Albers, G. W., Marks, M. P., Kemp, S., Christensen, S., Tsai, J. P., Ortega-Gutierrez, S., McTaggart, R. A., Torbey, M. T., Kim-Tenser, M., Leslie-Mazwi, T., Sarraj, A., Kasner, S. E., Ansari, S. A., Yeatts, S. D., Hamilton, S., Mlynash, M., Heit, J. J., Zaharchuk, G., Kim, S., Carrozzella, J., Palesch, Y. Y., Demchuk, A. M., Bammer, R., Lavori, P. W., Broderick, J. P., and Lansberg, M. G., “Thrombectomy for stroke at 6 to 16 hours with selection by perfusion imaging,” *New England Journal of Medicine* **378**, 708–718 (feb 2018).
- [2] Powers, W. J., Rabinstein, A. A., Ackerson, T., Adeoye, O. M., Bambakidis, N. C., Becker, K., Biller, J., Brown, M., Demaerschalk, B. M., Hoh, B., et al., “Guidelines for the early management of patients with acute ischemic stroke: 2019 update to the 2018 guidelines for the early management of acute ischemic stroke: a guideline for healthcare professionals from the American Heart Association/American Stroke Association,” *Stroke* **50**(12), e344–e418 (2019).
- [3] Bannasch, S., Frysck, R., Pfeiffer, T., Warnecke, G., and Rose, G., “Time separation technique: Accurate solution for 4d c-arm-CT perfusion imaging using a temporal decomposition model,” *Medical Physics* **45**, 1080–1092 (feb 2018).
- [4] Haseljić, H., Kulvait, V., Frysck, R., Hensen, B., Wacker, F., Rose, G., and Werncke, T., “The Application of Time Separation Technique to Enhance C-arm CT Dynamic Liver Perfusion Imaging,” in [*16th International Meeting on Fully Three-Dimensional Image Reconstruction in Radiology and Nuclear Medicine*], (2021).
- [5] Fieselmann, A., Kowarschik, M., Ganguly, A., Hornegger, J., and Fahrig, R., “Deconvolution-based CT and MR brain perfusion measurement: Theoretical model revisited and practical implementation details,” *International Journal of Biomedical Imaging* **2011**, 1–20 (2011).

*This work has not been submitted, published or presented elsewhere.*