

# VORTEX: Physics-Driven Data Augmentations Using Consistency Training for Robust Accelerated MRI Reconstruction

Arjun D. Desai\*, Beliz Gunel\*, Batu M. Ozturkler, Harris Beg†, Shreyas Vasanawala, Brian A. Hargreaves, Christopher Ré, John M. Pauly, Akshay S. Chaudhari

Stanford University

{arjundd,bgunel,ozt,beg,vasanawala,bah,chrismre,pauly,akshaysc}@stanford.edu

## Abstract

Deep neural networks have enabled improved image quality and fast inference times for various inverse problems, including accelerated magnetic resonance imaging (MRI) reconstruction. However, such models require a large number of fully-sampled ground truth datasets, which are difficult to curate, and are sensitive to distribution drifts. In this work, we propose applying *physics-driven* data augmentations for consistency training that leverage our domain knowledge of the forward MRI data acquisition process and MRI physics to achieve improved label efficiency and robustness to clinically-relevant distribution drifts. Our approach, termed VORTEX, (1) demonstrates strong improvements over supervised baselines with and without data augmentation in robustness to signal-to-noise ratio change and motion corruption in data-limited regimes; (2) considerably outperforms state-of-the-art purely image-based data augmentation techniques and self-supervised reconstruction methods on both in-distribution and out-of-distribution data; and (3) enables composing heterogeneous image-based and physics-driven data augmentations. Our code is available at <https://github.com/ad12/meddlr>.

## 1 Introduction


Magnetic resonance imaging (MRI) is a powerful diagnostic imaging technique; however, acquiring clinical MRI data typically requires long scan durations (30+ minutes). To reduce these durations, MRI data acquisition can be accelerated by undersampling the requisite spatial frequency measurements, referred to as *k-space* data. Reconstructing images without aliasing artifacts from such undersampled *k-space* measurements is ill-posed in the Hadamard sense [27]. To address this challenge, previous methods utilized underlying image priors to constrain the optimization, most notably by enforcing sparsity in a transformation domain, in a process called compressed sensing (CS) [40]. However, CS methods provide limited acceleration and require long reconstruction times and parameter-specific tuning [39, 3].

Deep learning (DL) based accelerated MRI reconstruction methods have recently enabled higher acceleration factors than traditional methods, with fast reconstruction times and improved image quality [28]. However, these approaches rely on large amounts of paired undersampled and fully-sampled reference data for training, which is often costly or simply impossible to acquire. State-of-the-art reconstruction methods use large fully-sampled (labeled) datasets, with only a handful of methods leveraging prospectively undersampled (unlabeled) data [11] or using image-based data augmentation schemes [21] to mitigate data paucity. These DL-based MR reconstruction methods are also highly sensitive to clinically-relevant distribution drifts, such as scanner-induced drifts, patient-induced artifacts, and forward model changes [17]. Despite being a critical need, sensitivity to distribution drifts remains largely unexplored, with only a few studies considering simple forward model alterations such as undersampling mask changes at inference time [23].

In this work, we leverage domain knowledge of the forward MRI data acquisition model and MRI physics through *physics-driven, acquisition-based* data augmentations for consistency training to build label-efficient networks that are robust to clinically-relevant distribution drifts such as signal-to-noise ratio (SNR) changes and motion artifacts. Our proposal builds on the Noise2Recon framework that conducts joint reconstruction for supervised scans and denoising for unsupervised scans [18] by extending the original consistency denoising objective to a generalized

\*Equal contribution.

†Work done as Summer Undergraduate Research Fellow.



**Figure 1:** VORTEX, a semi-supervised consistency training framework for robust accelerated MRI reconstruction, enforces invariance to physics-driven and equivariance to image-based data augmentations. **Supervised.** Fully-sampled (i.e. labeled) examples can be retrospectively undersampled, and model reconstructions can be optimized with respect to ground-truth references (orange arrow). **Consistency.** Undersampled-only (i.e. unlabeled, unsupervised) examples are reconstructed by the model to produce pseudo-references (red arrow). These examples are augmented (yellow box) and reconstructed by the model. Model reconstructions of the augmented version are optimized with respect to the pseudo-references. VORTEX supports both curriculum learning to schedule augmentation difficulty (blue box) and augmentation composition to pair image-based and physics-based transforms.

data augmentation pipeline. Specifically, we propose a semi-supervised consistency training framework, termed VORTEX, that uses a data augmentation pipeline to enforce invariance to physics-driven data augmentations of noise and motion, and equivariance to image-based data augmentations of flipping, scaling, rotation, translation, and shearing (Fig. 1). VORTEX supports curriculum learning based on the difficulty of physics-driven augmentations and composing heterogeneous augmentations. We demonstrate this generalized consistency paradigm increases robustness to varying test-time perturbations without decreasing the reconstruction performance on non-perturbed, in-distribution data. We also show that VORTEX outperforms both the state-of-the-art self-supervised training strategy SSDU [62] and the data augmentation scheme MRAugment [21], which solely relies on image-based data augmentations. Unlike MRAugment, which requires preservation of training data noise statistics, VORTEX can operate on a broader family of MRI physics-based augmentations without noise statistics constraints. Our contributions are the following:

1. We propose VORTEX, a semi-supervised consistency training framework for accelerated MRI reconstruction that enables composing image-based data augmentations with *physics-driven* data augmentations. VORTEX leverages domain knowledge of MRI physics and the MRI data acquisition forward model to improve label efficiency and robustness.
2. We demonstrate strong improvements over supervised and self-supervised baselines in robustness to two clinically-relevant distribution drifts: scanner-induced SNR change and patient-induced motion artifacts. Notably, we obtain +0.106 structural similarity (SSIM) and +5.3dB complex PSNR (cPSNR) improvement over supervised baselines on heavily motion-corrupted scans in label-scarce regimes.
3. We outperform MRAugment, a state-of-the-art, purely image-based data augmentation technique for MRI reconstruction. We achieve improvements of +0.061 SSIM and +0.2dB cPSNR on in-distribution data, +0.089 SSIM and +2.5dB cPSNR on noise-corrupted data, and +0.125 SSIM and +7.8dB cPSNR on motion-corrupted data.

4. We conduct ablations comparing image space and latent space consistency and designing curricula for data augmentation difficulty.

Our code, experiments, and pretrained models are publicly available<sup>1</sup>.

## 2 Related Work

Supervised accelerated MRI reconstruction methods map zero-filled images obtained from undersampled measurements to fully-sampled ground truth images using fully-convolutional networks (e.g. U-Net [52]) or unrolled networks modeling iterative proximal-update optimization methods [1, 2, 54]. However, such approaches rely on a large corpus of fully-sampled (i.e. labeled) scans. Although lagging in performance to supervised techniques, several studies have leveraged unsupervised data including using generative adversarial networks [35, 14], self-supervised learning [62], and dictionary learning [33]. Other works have proposed image-based data augmentations to reduce model overfitting when labeled training data is limited [21].

Several methods have also explored building neural networks robust to distribution drifts for image classification [57, 48, 24] and natural language processing tasks [42, 25]. For accelerated MRI reconstruction, [17] recently demonstrated trained deep neural networks, un-trained networks [16], and traditional iterative approaches are sensitive to adversarial perturbations and distribution drifts. However, unlike what we propose with consistency training, no mitigation approaches were discussed.

Consistency training was initially proposed to train models to be invariant to noisy input data [43, 53, 13] or hidden representations [5, 34]. Desai et al. [18] extended these methods to a consistency training framework for joint MRI image reconstruction and denoising, where complex-valued additive Gaussian noise was applied to undersampled k-space. Beyond noise-based consistency, Xie et al. [61] showed that semantics-preserving data augmentation consistency (RandAugment [15] for vision tasks and back-translation for language tasks [20]) led to significant performance boosts.

Motion correction for MRI is an active research area, as scans corrupted by patient motion affect the diagnostic image quality and clinical outcomes [12, 6]. Pawar et al. [46] proposed a supervised DL method to map simulated motion-corrupted scans to clean scans as a post-processing method after reconstruction. Liu et al. [36] extended iterative application of image denoisers as imaging priors [51] for general artifact removal such as that of motion. Gan et al. [22] extended this method by training the model in the measurement domain without supervised data. However, these methods require multiple measurements of the same object undergoing nonrigid deformation, which is unrealistic in most clinical settings. An extended discussion on related work in comparison to our work is available in Appendix B.

## 3 Background on Accelerated Multi-coil MRI Reconstruction


In MRI, measurements are acquired in the spatial frequency domain, referred to as *k-space*. In this work, we consider the clinically-relevant case of accelerated multi-coil MRI acquisition where multiple receiver coils are used to acquire spatially-localized k-space measurements modulated by corresponding *sensitivity maps*. Sensitivity maps are generally unknown and vary per patient, and thus, need to be estimated to perform reconstruction [47]. In accelerated MRI reconstruction, the goal is to reduce scan times by decreasing the number of samples acquired in k-space. The undersampling operation can be represented by a binary mask  $\Omega$  that indexes acquired samples in k-space. The forward problem for multi-coil accelerated MRI can be written as:

$$y = \Omega F S x^* + \epsilon = A x^* + \epsilon$$

where  $y$  is the measured signal in k-space,  $F$  is the discrete Fourier transform matrix,  $S$  are the receiver coil sensitivity maps,  $x^*$  is the ground-truth signal in image-space, and  $\epsilon$  is additive complex Gaussian noise.  $A = \Omega F S$  is the known

---

<sup>1</sup><https://github.com/ad12/meddlr>



**Figure 2:** Examples of image-based, physics-driven (motion where  $\alpha = 0.2$ , noise where  $\sigma = 0.2$ ), and composed (image + motion + noise) augmentations applied to a fully-sampled image.

forward operator during acquisition (see Appendix A for notation). This problem is ill-posed in the Hadamard sense [27] as we have fewer measurements than variables to recover. It does not satisfy the three conditions of 1) existence of a solution, 2) uniqueness, and 3) continuous dependence on measurements to be defined as well-posed. As such, prior assumptions of the characteristics of the data (e.g. sparsity in compressed sensing [40]) are required to uniquely recover the underlying image  $x^*$ .

## 4 Methods

We propose VORTEX, a semi-supervised consistency training framework that integrates a generalized data augmentation pipeline for accelerated MRI reconstruction (Fig. 1). We consider the setup with dataset  $\mathcal{D}$  that consists of (1) fully-sampled examples in k-space  $y^{(s)}$  with corresponding supervised reference ground truth images  $x^{(s)}$ , and (2) undersampled-only k-space examples  $y^{(u)}$ , which do not have supervised references.  $f_\theta$  is the learned reconstruction model with the forward operator  $A$ . A pixel-wise  $\ell_1$  supervised loss  $\mathcal{L}_{sup}$  is computed for supervised examples  $y^{(s)}$ . Undersampled examples  $y^{(u)}$  are passed through the *Augmentation Pipeline* (see §4.1). We consider the case where there are more unsupervised examples than supervised examples, a common observation in clinical practice.

Let  $T_I$  denote the set of invariant transformations consisting of MR physics-driven data augmentations such as additive complex Gaussian noise and motion corruption (see §4.1.1). Similarly, let  $T_E$  denote the equivariant transformations that include image-based data augmentations such as flipping, rotation, translation, scaling, and shearing (see §4.1.2). We define our use of the terms *invariance* and *equivariance* and how these definitions motivate the structure of augmentations in Appendix C. A pixel-wise  $\ell_1$  consistency loss  $\mathcal{L}_{cons}$  is computed between the model reconstruction outputs of input undersampled examples with and without augmentation (Eq. 1). The overall training objective is shown in Eq. 2.

$$\mathcal{L}_{cons} = \begin{cases} \|f_\theta(y_i^u, A) - f_\theta(g(y_i^u), A)\|_1, & \text{if } g \in T_I \\ \|g(f_\theta(y_i^u, A)) - f_\theta(g(y_i^u), A)\|_1, & \text{if } g \in T_E \end{cases} \quad (1)$$

$$\mathcal{L}_{VORTEX} = \sum_i \|f_\theta(y_i^s, A) - x_i^s\|_1 + \lambda \mathcal{L}_{cons} \quad (2)$$

### 4.1 Generalized Data Augmentation Pipeline for Consistency Training

Our augmentation pipeline enables composing state-of-the-art image-based data augmentations with physics-driven data augmentations motivated by the MRI data acquisition forward model.

#### 4.1.1 Physics-Driven Data Augmentations

**Noise.** One of the most common artifacts in MRI is noise; thus, practical MRI reconstruction methods should be robust to variations in SNR. Because noise in MRI is dominated by thermal fluctuations in the subject and due to



receiver coil electronics, magnetic field strength, and specific imaging parameters, it can be modeled as an additive complex-valued Gaussian distribution [41]. Given it can be modeled well in the MRI acquisition process, we leverage noise for consistency training [41]. Specifically, we sample  $\sigma$  from a range  $\mathcal{R}(\sigma) = [\sigma^{\text{LN}}, \sigma^{\text{HN}})$  where **LN** (light noise) and **HN** (heavy noise) are chosen based on visual inspections of clinical scans by a board-certified clinical radiologist. We normalize each sampled  $\sigma$  to the magnitude of the image to induce the same relative SNR changes across scans. We denote the operation of adding noise to the k-space as  $g_N$ , where the noise-augmented unsupervised example is given by  $g_N(y_i^{(u)}) = y_i^{(u)} + \epsilon_i$ . We provide an example of a noise-augmented scan in Fig. 2.

**Motion.** Rigid patient motion, which is prevalent among pediatric and claustrophobic patients, during MRI scans can considerably degrade image quality and result in ghosting artifacts. While navigator-based sequences that sample low-resolution motion can be used for motion correction, they require custom sequences that often lead to increased acquisition time, reduced SNR, and complicated reconstruction [64]. Many multi-shot MRI acquisitions sample data over multiple *shots* where consecutive k-space lines are acquired in separate excitations [4]. Here, motion across every shot manifests as additional phase in k-space and as translation in image space. Thus, one-dimensional translational motion artifacts across the phase dimension can be modeled using random phase errors that alter odd and even lines of k-space separately. We leverage motion for consistency training since we can precisely model rigid motion in k-space. We denote the phase error due to motion for  $i^{\text{th}}$  example by  $e^{-j\phi_i}$  that corresponds to a translational motion. We sample two random numbers from the uniform distribution  $m_o, m_e \sim U(-1, 1)$  which is chosen from a specified range  $\mathcal{R}(\alpha) = [\alpha^{\text{LM}}, \alpha^{\text{HM}})$  where  $\alpha$  denotes the amplitude of the phase errors and **LM** (light motion) and **HM** (heavy motion) are chosen based on visual inspections of clinical scans by a board-certified clinical radiologist. For a given k-space readout  $k^{\text{th}}$ , the phase error is:

$$\phi_i^k = \begin{cases} \pi\alpha m_o, & \text{if } k \text{ is odd} \\ \pi\alpha m_e, & \text{if } k \text{ is even} \end{cases}$$

We denote the operation of adding motion to the k-space as  $g_M$ , in which case the motion-augmented unsupervised example is given by  $g_M(y_i^{(u)}) = y_i^{(u)} e^{-j\phi_i}$  (example scan shown in Fig. 2).

#### 4.1.2 Image-based Data Augmentations

In contrast to classification problems where labels are invariant with respect to the augmentations, data augmentations in the MR reconstruction task need to transform the target images and their corresponding k-space and coil sensitivity measurements. Unlike physics-driven augmentations that occur in k-space, image-based augmentations occur in the image domain. Since the training data initially exists as k-space measurements, we transform it into the image domain using coil sensitivity maps, and subsequently apply a cascade of the image-based data augmentations to both the image and the sensitivity maps. Image-based data augmentations include pixel-preserving augmentations such as flipping, translation, arbitrary and 90 degree multiple rotations, translation, as well as isotropic and anisotropic scaling. Using the augmented image and transformed sensitivity maps, we run the forward operator  $A$  to generate the corresponding undersampled k-space measurements.

#### 4.1.3 Composing Augmentations

Our *Augmentation Pipeline* allows for composing different combinations of physics-driven and image-based data augmentations, with example composed augmentations shown in Fig. 2. It is important to note that composing multiple physics-driven augmentations such as noise and motion corruption represents a real-world scenario as multiple artifacts can occur simultaneously during MRI acquisition. Appendix C discusses augmentation composition and the generalized multi-augmentation consistency loss in detail.

## 4.2 Augmentation Scheduling

We adopt curriculum learning [26] for physics-driven data augmentations, where we seek to schedule the *task difficulty*. Difficulty is denoted by  $\sigma$ , the standard deviation of the additive zero-mean complex-valued Gaussian noise, and  $\alpha$ ,

**Table 1:** Average test results for in-distribution data and out-of-distribution data with heavy motion ( $\alpha=0.4$ ) and heavy noise ( $\sigma=0.4$ ) perturbations. Physics augmentations are compositions of noise and motion in their *heavy* training difficulty configurations. A detailed set of ablations comparing latent space vs pixel-level consistency, variations in augmentation scheduling, effect of training time, and robustness to changes in acceleration factors are provided in Appendix E.

Perturbation	None		Motion (heavy)		Noise (heavy)	
Model	SSIM	cPSNR (dB)	SSIM	cPSNR (dB)	SSIM	cPSNR (dB)
Compressed Sensing	0.847 (0.011)	35.4 (0.4)	0.724 (0.090)	24.5 (1.4)	0.708 (0.014)	30.5 (0.2)
Supervised	0.798 (0.038)	35.8 (0.4)	0.706 (0.048)	27.0 (0.8)	0.807 (0.015)	32.2 (0.3)
MRAugment	0.811 (0.043)	36.2 (0.5)	0.660 (0.040)	24.0 (1.0)	0.742 (0.005)	30.8 (0.3)
SSDU	0.787 (0.026)	34.9 (0.4)	0.734 (0.009)	31.9 (1.7)	0.716 (0.023)	32.5 (0.3)
Aug (Physics)	0.789 (0.045)	35.7 (0.4)	0.739 (0.010)	31.9 (2.4)	0.739 (0.051)	33.4 (0.3)
Aug (Image+Physics)	0.785 (0.050)	36.1 (0.5)	0.742 (0.022)	<b>32.8 (2.4)</b>	0.727 (0.051)	33.7 (0.4)
VORTEX (Image)	0.862 (0.030)	36.4 (0.3)	0.648 (0.080)	26.1 (0.7)	0.767 (0.016)	31.5 (0.2)
VORTEX (Physics)	<b>0.872 (0.033)</b>	<b>36.4 (0.3)</b>	<b>0.785 (0.019)</b>	31.8 (2.8)	0.817 (0.034)	<b>33.9 (0.2)</b>
VORTEX (Image+Physics)	0.861 (0.036)	36.4 (0.4)	0.777 (0.034)	31.1 (2.7)	<b>0.831 (0.023)</b>	33.3 (0.1)

the amplitude of the phase errors for motion. Note that this is in contrast to the MRAugment scheduling strategy, which only schedules the probability  $p$  of an augmentation. Concretely, for noise, we consider a time-varying range  $\mathcal{R}(\sigma(t)) = [\sigma^L, \sigma^H(t))$ , where  $t$  indexes the iteration number during training. The upper-bound  $\sigma^H(t)$  increases monotonically to ensure task difficulty increases during training. We consider two scheduling techniques  $\beta(t)$  such that  $\sigma^H(t) = \sigma^L + \beta(t)(\sigma^H - \sigma^L)$ : (1) **Linear**:  $\beta(t) = t/M$ , and (2) **Exponential**:  $\beta(t) = \frac{1-e^{-t/\tau}}{1-e^{-M/\tau}}$ , where  $M$  is the number of epochs until which task difficulty increases and  $\tau$  is the time-constant for exponential scheduling. After  $M$  epochs, training proceeds with constant upper bound  $\sigma^H$ . Scheduling for motion is the same where  $\sigma$  is replaced with  $\alpha$ . Image-based data augmentations follow the scheduling strategy proposed in MRAugment as there is no explicit sense of difficulty for that family of data augmentations. Fig. 5 in Appendix E.2 shows simulated  $\beta(t)$  for different curricula configurations.


## 5 Experiments

We evaluate our method using the publicly available 3D mridata fast-spin-echo (FSE) multi-coil knee dataset [45]. 3D MRI scans were decoded into a hybrid k-space ( $x \times k_y \times k_z$ ) using the 1D orthogonal inverse Fourier transform along the readout direction  $x$ . All methods reconstructed 2D  $k_y \times k_z$  slices. Sensitivity maps were estimated for each slice using JSENSE [63]. 2D Poisson Disc undersampling masks were used for training and evaluation.  $N_s$  training scans were randomly selected to be fully-sampled (labeled) examples while  $N_u$  scans were used to simulate undersampled-only (unlabeled) scans. All methods used a 2D U-Net network with a complex- $\ell_1$  training objective for both supervised and consistency losses. Reconstructions were evaluated with two image quality metrics: magnitude structural similarity (SSIM) [60] and complex peak signal-to-noise ratio (cPSNR) in decibels (dB). Appendix D discusses the experimental setup in further detail.

We also evaluate our method on the 2D fastMRI multi-coil brain dataset [65]. Details for experimental setup and results on this dataset are provided in Appendix D.1 and Appendix F, respectively. Appendix E discusses ablation experiments comparing latent space vs pixel-level consistency, variations in augmentation scheduling, effect of training time, and robustness to changes in acceleration factors.

### 5.1 Robustness to Clinically Relevant Distribution Drifts

Unlike many other ML domains, the source of possible distribution drifts in accelerated MRI reconstruction can be well characterized based on the known, physics-driven forward data acquisition process. This enables accurate simulations of many of these distribution drifts. In this work, we simulate SNR and motion corruptions, two common MRI artifacts, at inference time using models described in Section 4.1.1. Specifically, we use  $\sigma = 0.2$  for light noise,  $\sigma = 0.4$  for heavy noise,  $\alpha = 0.2$  for light motion, and  $\alpha = 0.4$  for heavy motion.



**Figure 3:** Example reconstructions for simulated scans with heavy motion (top) and heavy noise (bottom). Compressed sensing (CS), *Supervised*, MRAugment, SSDU, and Noise2Recon amplify motion ghosting artifacts (blue arrow). While supervised training with motion augmentations (*Aug (Motion)*) reduces these artifacts, it still suffers from artifacts (red arrow) and extensive blurring. *VORTEX (Motion)* and *VORTEX (Physics)* (i.e. Motion+Noise) suppress these artifacts. Methods without noise augmentations (CS, *Supervised*, MRAugment, SSDU, *Aug (Motion)*, *VORTEX (Motion)*) amplify image noise. *VORTEX (Physics)* suppresses noise without over-blurring the image.

In Table 1, we compare VORTEX’s performance for in-distribution and OOD data at 16x acceleration to supervised methods using both physics-driven and the state-of-the-art image-based MRAugment augmentations, and to the state-of-the-art self-supervised via data undersampling (SSDU) reconstruction method [62]. We describe all baseline implementations in detail in Appendix D.2. We isolate the benefits of consistency training with VORTEX from the utility of the augmentations (*Aug*) themselves by separately comparing MRAugment (i.e. *Aug (Image)*), *Aug (Physics)*, and *Aug (Image + Physics)* (details in Appendix D.3.1).

In Table 2, we compare supervised baselines without and with the physics-based augmentations to VORTEX at different OOD motion and noise levels at inference time. We note that Noise2Recon is a specialized case of the VORTEX framework (i.e. *VORTEX (Noise)*) where the Augmentation Pipeline only consists of noise augmentations. For the supervised training with augmentation methods, augmentation is applied with probability  $p = 0.2$  during training for noise, motion, and composition corresponding to  $g_N(g_M(\cdot))$ . For enforcing consistency, we used  $\lambda = 0.1$  for  $\mathcal{L}_{cons}$  weighting for noise, motion, and composition that we refer to as *Physics*. Both *Aug (Motion)* and *VORTEX (Motion)* models were trained with  $\mathcal{R}(\alpha) = [0.2, 0.5]$ , and both *Aug (Noise)* and *VORTEX (Noise)* models were trained with  $\mathcal{R}(\sigma) = [0.2, 0.5]$ . *Aug (Physics)* and *VORTEX (Physics)* setting also follow these ranges. We used a balanced data sampling approach where unsupervised and supervised examples are sampled at a fixed ratio of 1:1 during training [18]. All consistency training approaches used augmentation curricula with highest validation cPSNR as described in Section 4.2.

All experiments are conducted in the semi-supervised setting, where both fully-sampled (labeled) and undersampled-only (unlabeled) data are available. Results are shown with 5x more unsupervised slices than supervised (1600 vs 320), which is a realistic clinical scenario. We show results for different accelerations, training times and augmentation curricula in Appendices D and E.

## 5.2 VORTEX vs. Baseline Results

**Label efficiency among in-distribution data.** As shown in Table 1, *VORTEX (Physics)* demonstrated substantial improvements of +0.074 SSIM and +0.6dB cPSNR over the *Supervised* baseline, +0.061 SSIM and +0.2dB cPSNR over *MRAugment*, and +0.085 SSIM and +1.5dB cPSNR over *SSDU* for in-distribution data. As *VORTEX (Image)* also considerably improves over *Supervised* and *MRAugment*, a dominant mechanism of the benefits may be attributed to the consistency training even for the in-distribution setting.

**Generalizing to motion and noise corrupted data.** For both heavy motion and noise settings, including physics augmentations is vital for robust performance, as *MRAugment*, *SSDU*, and *VORTEX (Image)* perform worse, even compared to the *Supervised* baseline. For heavy motion, we show an improvement of +0.079 SSIM and +4.8dB cPSNR over the *Supervised* and +0.125 SSIM and +7.8dB cPSNR over *MRAugment* with *VORTEX (Physics)*. Similarly, for

**Table 2:** Results at 16x acceleration for in-distribution and OOD with SNR and motion perturbations. VORTEX uses curriculum learning (see Appendix E.2).

Perturbation	Metric	Supervised	Aug (Motion)	Aug (Noise)	Aug (Physics)	VORTEX (Motion)	Noise2Recon (i.e. VORTEX (Noise))	VORTEX (Physics)
None	SSIM	0.798 (0.038)	0.793 (0.041)	0.805 (0.045)	0.789 (0.045)	0.877 (0.029)	<b>0.882 (0.031)</b>	0.869 (0.030)
	cPSNR (dB)	35.8 (0.4)	35.9 (0.5)	35.8 (0.4)	35.7 (0.4)	36.4 (0.3)	<b>36.4 (0.3)</b>	36.4 (0.3)
Motion (light)	SSIM	0.809 (0.028)	0.793 (0.028)	0.799 (0.039)	0.785 (0.036)	<b>0.867 (0.021)</b>	0.854 (0.015)	0.854 (0.029)
	cPSNR (dB)	33.6 (0.2)	35.1 (0.5)	34.1 (0.4)	35.0 (0.4)	<b>35.8 (0.5)</b>	32.8 (1.3)	35.4 (0.1)
Motion (heavy)	SSIM	0.706 (0.048)	0.751 (0.025)	0.722 (0.042)	0.739 (0.010)	<b>0.812 (0.021)</b>	0.731 (0.014)	0.803 (0.017)
	cPSNR (dB)	27.0 (0.8)	31.5 (2.7)	29.6 (1.4)	31.9 (2.4)	<b>32.3 (2.5)</b>	27.1 (1.7)	32.3 (2.6)
Noise (light)	SSIM	0.830 (0.024)	0.786 (0.032)	0.778 (0.049)	0.761 (0.049)	<b>0.857 (0.015)</b>	0.854 (0.033)	0.840 (0.034)
	cPSNR (dB)	33.8 (0.3)	33.7 (0.3)	34.2 (0.3)	34.2 (0.3)	34.0 (0.1)	<b>34.8 (0.2)</b>	34.8 (0.2)
Noise (heavy)	SSIM	0.807 (0.015)	0.758 (0.024)	0.745 (0.054)	0.739 (0.051)	0.823 (0.008)	<b>0.830 (0.033)</b>	0.812 (0.033)
	cPSNR (dB)	32.2 (0.3)	32.0 (0.3)	33.5 (0.3)	33.4 (0.3)	32.4 (0.1)	<b>34.0 (0.3)</b>	33.9 (0.3)

heavy noise, we show an improvement of +0.024 SSIM and +1.1dB cPSNR over the Supervised baseline and +0.089 SSIM and +2.5dB cPSNR over *MRAugment* with *VORTEX (Image + Physics)*. In Table 2, we observe consistent improvements over both *Supervised* and *Aug* baselines for light and heavy motion cases, with a large improvement of +0.106 SSIM and +5.3dB cPSNR with *VORTEX (Motion)* over *Supervised* for heavy motion-corruptions, demonstrating the strength of our method at varying levels of OOD corruptions at inference time. Also, we depict a large improvement of +0.079 SSIM/+0.6dB pSNR with *VORTEX (Motion)* and +0.084 SSIM/+0.6dB cPSNR with *VORTEX (Noise)* compared to the supervised baseline for in-distribution data, while none of the *Aug* baselines show any meaningful improvement. Similar trends were observed with the 2D fastMRI brain dataset (Appendix F.2). Example reconstructions are shown in Fig. 3.

The substantial performance gain with VORTEX in both in-distribution and OOD settings suggests that the consistency training framework is amenable to both image-based and physics-driven augmentations. While supervised training requires noise statistics-preserving augmentations, consistency training can relax this constraint and allow for more diverse augmentations (see Appendix B.1). We highlight that our proposed consistency-based improvements are considerably larger than prior reported values for DL methods that use different architectures, loss functions, or data consistency schemes [65, 29]. We also note that SSIM is clinically preferred to cPSNR for quantifying perceptual quality [32].

## 6 Conclusion

We propose VORTEX, a semi-supervised consistency training framework for accelerated MRI reconstruction that uses a generalized data augmentation pipeline for improved label-efficiency and robustness to clinically relevant distribution drifts. VORTEX enforces invariance to *physics-driven*, *acquisition-based* augmentations and enforces equivariance to image-based augmentations, enables composing data augmentations of different types, and allows for curriculum learning based on the difficulty of physics-driven augmentations. We demonstrate strong improvements over fully-supervised baselines and state-of-the-art data augmentation (MRAugment) and self-supervised (SSDU) methods on both in-distribution and OOD data. Our framework is model-agnostic and could be used with any other MRI reconstruction models or even for other image-to-image tasks with appropriate data augmentations. Besides the strengths of our method, we also note several limitations. While our framework is flexible to work with any motion model, we utilized a simpler 1D motion model. We also did not evaluate VORTEX on prospective data or consider complex artifacts that are challenging to model such as  $B_0$  variations. In future work, we plan to extend physics-driven, acquisition-based augmentations to account for more complex motion models, additional OOD MRI artifacts, and non-Cartesian undersampling patterns in prospectively acquired clinical data to work towards building robust DL-based MR reconstruction models that can be safely deployed in clinics.

## 7 Acknowledgements

This work was supported by R01 AR063643, R01 AR AR077604, R01 EB002524, R01 EB009690, R01 EB026136, K24 AR062068, and P41 EB015891 from the NIH; the Precision Health and Integrated Diagnostics Seed Grant from Stanford University; DOD – National Science and Engineering Graduate Fellowship (ARO); National Science Foundation (GRFP-DGE 1656518, CCF1763315, CCF1563078); Stanford Artificial Intelligence in Medicine and Imaging GCP grant; Stanford Human-Centered Artificial Intelligence GCP grant; GE Healthcare and Philips.

## References

- [1] Jonas Adler and Ozan Öktem. Learned primal-dual reconstruction. *IEEE transactions on medical imaging*, 37(6): 1322–1332, 2018.
- [2] Hemant K Aggarwal, Merry P Mani, and Mathews Jacob. Modl: Model-based deep learning architecture for inverse problems. *IEEE transactions on medical imaging*, 38(2):394–405, 2018.
- [3] Thai Akasaka, Koji Fujimoto, Takayuki Yamamoto, Tomohisa Okada, Yasutaka Fushumi, Akira Yamamoto, Toshiyuki Tanaka, and Kaori Togashi. Optimization of regularization parameters in compressed sensing of magnetic resonance angiography: can statistical image metrics mimic radiologists’ perception? *PloS one*, 11(1): e0146548, 2016.
- [4] Adam W Anderson and John C Gore. Analysis and correction of motion artifacts in diffusion weighted imaging. *Magnetic resonance in medicine*, 32(3):379–387, 1994.
- [5] Philip Bachman, Ouais Alsharif, and Doina Precup. Learning with pseudo-ensembles. *Advances in neural information processing systems*, 27:3365–3373, 2014.
- [6] GJ Barker. Technical issues for the study of the optic nerve with mri. *Journal of the neurological sciences*, 172: S13–S16, 2000.
- [7] Joshua Batson and Loic Royer. Noise2self: Blind denoising by self-supervision. In *International Conference on Machine Learning*, pages 524–533. PMLR, 2019.
- [8] Yoshua Bengio, Jérôme Louradour, Ronan Collobert, and Jason Weston. Curriculum learning. In *Proceedings of the 26th annual international conference on machine learning*, pages 41–48, 2009.
- [9] Robert Bridson. Fast poisson disk sampling in arbitrary dimensions. *SIGGRAPH sketches*, 10(1), 2007.
- [10] Elena Celledoni, Matthias Joachim Ehrhardt, Christian Etmann, Brynjulf Owren, Carola-Bibiane Schonlieb, and Ferdia Sherry. Equivariant neural networks for inverse problems. *Inverse Problems*, 37, 2021.
- [11] Akshay S Chaudhari, Christopher M Sandino, Elizabeth K Cole, David B Larson, Garry E Gold, Shreyas S Vasanawala, Matthew P Lungren, Brian A Hargreaves, and Curtis P Langlotz. Prospective deployment of deep learning in mri: A framework for important considerations, challenges, and recommendations for best practices. *Journal of Magnetic Resonance Imaging*, 2020.
- [12] Govind B Chavhan, Paul S Babyn, and Shreyas S Vasanawala. Abdominal mr imaging in children: motion compensation, sequence optimization, and protocol organization. *Radiographics*, 33(3):703–719, 2013.
- [13] Kevin Clark, Minh-Thang Luong, Christopher D. Manning, and Quoc V. Le. Semi-supervised sequence modeling with cross-view training. In *EMNLP*, 2018.
- [14] Elizabeth K Cole, John M Pauly, Shreyas S Vasanawala, and Frank Ong. Unsupervised mri reconstruction with generative adversarial networks. *arXiv preprint arXiv:2008.13065*, 2020.

- [15] Ekin D Cubuk, Barret Zoph, Jonathon Shlens, and Quoc V Le. Randaugment: Practical automated data augmentation with a reduced search space. In *Proceedings of the IEEE/CVF Conference on Computer Vision and Pattern Recognition Workshops*, pages 702–703, 2020.
- [16] Mohammad Zalbagi Darestani and Reinhard Heckel. Accelerated mri with un-trained neural networks. *IEEE Transactions on Computational Imaging*, 7:724–733, 2021.
- [17] Mohammad Zalbagi Darestani, Akshay S. Chaudhari, and Reinhard Heckel. Measuring robustness in deep learning based compressive sensing. In *International Conference on Machine Learning (ICML)*, 2021.
- [18] Arjun D Desai, Batu M Ozturkler, Christopher M Sandino, Shreyas Vasanawala, Brian A Hargreaves, Christopher M Re, John M Pauly, and Akshay S Chaudhari. Noise2recon: A semi-supervised framework for joint mri reconstruction and denoising. *arXiv preprint arXiv:2110.00075*, 2021.
- [19] Arjun D Desai, Andrew M Schmidt, Elka B Rubin, Christopher Michael Sandino, Marianne Susan Black, Valentina Mazzoli, Kathryn J Stevens, Robert Boutin, Christopher Re, Garry E Gold, et al. Skm-tea: A dataset for accelerated mri reconstruction with dense image labels for quantitative clinical evaluation. In *Thirty-fifth Conference on Neural Information Processing Systems Datasets and Benchmarks Track (Round 2)*, 2021.
- [20] Sergey Edunov, Myle Ott, Michael Auli, and David Grangier. Understanding back-translation at scale. In *EMNLP*, 2018.
- [21] Zalan Fabian, Reinhard Heckel, and Mahdi Soltanolkotabi. Data augmentation for deep learning based accelerated mri reconstruction with limited data. In *International Conference on Machine Learning*, pages 3057–3067. PMLR, 2021.
- [22] Weijie Gan, Yu Sun, Cihat Eldeniz, Jiaming Liu, Hongyu An, and Ulugbek S Kamilov. Modir: Motion-compensated training for deep image reconstruction without ground truth. *arXiv preprint arXiv:2107.05533*, 2021.
- [23] Davis Gilton, Greg Ongie, and R. Willett. Model adaptation for inverse problems in imaging. *IEEE Transactions on Computational Imaging*, 7:661–674, 2021.
- [24] Karan Goel, Albert Gu, Yixuan Li, and Christopher Ré. Model patching: Closing the subgroup performance gap with data augmentation. *arXiv preprint arXiv:2008.06775*, 2020.
- [25] Beliz Gunel, Jingfei Du, Alexis Conneau, and Veselin Stoyanov. Supervised contrastive learning for pre-trained language model fine-tuning. In *International Conference on Learning Representations*, 2021. URL <https://openreview.net/forum?id=cu7IUihUjH>.
- [26] Guy Hach Cohen and Daphna Weinshall. On the power of curriculum learning in training deep networks. In *International Conference on Machine Learning*, pages 2535–2544. PMLR, 2019.
- [27] Jacques Hadamard. Sur les problèmes aux dérivées partielles et leur signification physique. *Princeton university bulletin*, pages 49–52, 1902.
- [28] Kerstin Hammernik, Teresa Klatzer, Erich Kobler, Michael P Recht, Daniel K Sodickson, Thomas Pock, and Florian Knoll. Learning a variational network for reconstruction of accelerated mri data. *Magnetic resonance in medicine*, 79(6):3055–3071, 2018.
- [29] Kerstin Hammernik, Jo Schlemper, Chen Qin, Jinming Duan, Ronald M Summers, and Daniel Rueckert. Systematic evaluation of iterative deep neural networks for fast parallel mri reconstruction with sensitivity-weighted coil combination. *Magnetic Resonance in Medicine*, 2021.
- [30] Kyong Hwan Jin, Ji-Yong Um, Dongwook Lee, Juyoung Lee, Sung-Hong Park, and Jong Chul Ye. Mri artifact correction using sparse+ low-rank decomposition of annihilating filter-based hankel matrix. *Magnetic resonance in medicine*, 78(1):327–340, 2017.
- [31] Diederik P Kingma and Jimmy Ba. Adam: A method for stochastic optimization. *arXiv preprint arXiv:1412.6980*, 2014.

- [32] Florian Knoll, Tullie Murrell, Anuroop Sriram, Nafissa Yakubova, Jure Zbontar, Michael Rabbat, Aaron Defazio, Matthew J Muckley, Daniel K Sodickson, C Lawrence Zitnick, et al. Advancing machine learning for mr image reconstruction with an open competition: Overview of the 2019 fastmri challenge. *Magnetic resonance in medicine*, 84(6):3054–3070, 2020.
- [33] Anish Lahiri, Guanhua Wang, Saiprasad Ravishankar, and Jeffrey A Fessler. Blind primed supervised (blips) learning for mr image reconstruction. *arXiv preprint arXiv:2104.05028*, 2021.
- [34] Samuli Laine and Timo Aila. Temporal ensembling for semi-supervised learning. *arXiv preprint arXiv:1610.02242*, 2016.
- [35] Ke Lei, Morteza Mardani, John M Pauly, and Shreyas S Vasanawala. Wasserstein gans for mr imaging: from paired to unpaired training. *IEEE transactions on medical imaging*, 40(1):105–115, 2020.
- [36] Jiaming Liu, Yu Sun, Cihat Eldeniz, Weijie Gan, Hongyu An, and Ulugbek S Kamilov. Rare: Image reconstruction using deep priors learned without groundtruth. *IEEE Journal of Selected Topics in Signal Processing*, 14(6): 1088–1099, 2020.
- [37] Ilya Loshchilov and Frank Hutter. Decoupled weight decay regularization. *arXiv preprint arXiv:1711.05101*, 2017.
- [38] Wenmiao Lu, Kim Butts Pauly, Garry E Gold, John M Pauly, and Brian A Hargreaves. Semac: slice encoding for metal artifact correction in mri. *Magnetic Resonance in Medicine: An Official Journal of the International Society for Magnetic Resonance in Medicine*, 62(1):66–76, 2009.
- [39] Michael Lustig, David Donoho, and John M Pauly. Sparse mri: The application of compressed sensing for rapid mr imaging. *Magnetic Resonance in Medicine: An Official Journal of the International Society for Magnetic Resonance in Medicine*, 58(6):1182–1195, 2007.
- [40] Michael Lustig, David L Donoho, Juan M Santos, and John M Pauly. Compressed sensing mri. *IEEE signal processing magazine*, 25(2):72–82, 2008. doi: 10.1109/MSP.2007.914728.
- [41] Albert Macovski. Noise in mri. *Magnetic resonance in medicine*, 36(3):494–497, 1996.
- [42] John Miller, Karl Krauth, Benjamin Recht, and Ludwig Schmidt. The effect of natural distribution shift on question answering models. In *International Conference on Machine Learning*, pages 6905–6916. PMLR, 2020.
- [43] Takeru Miyato, Shin-ichi Maeda, Masanori Koyama, and Shin Ishii. Virtual adversarial training: a regularization method for supervised and semi-supervised learning. *IEEE transactions on pattern analysis and machine intelligence*, 41(8):1979–1993, 2018.
- [44] F Ong and M Lustig. Sigpy: a python package for high performance iterative reconstruction. In *Proceedings of the ISMRM 27th Annual Meeting, Montreal, Quebec, Canada*, volume 4819, 2019.
- [45] F Ong, S Amin, S Vasanawala, and M Lustig. Mridata.org: An open archive for sharing mri raw data. In *Proc. Intl. Soc. Mag. Reson. Med*, volume 26, page 1, 2018.
- [46] Kamlesh Pawar, Zhaolin Chen, N. Jon Shah, and Gary F. Egan. Suppressing motion artefacts in mri using an inception-resnet network with motion simulation augmentation. *NMR in Biomedicine*, Dec 2019. ISSN 1099-1492. doi: 10.1002/nbm.4225. URL <http://dx.doi.org/10.1002/nbm.4225>.
- [47] Klaas P Pruessmann, Markus Weiger, Markus B Scheidegger, and Peter Boesiger. Sense: sensitivity encoding for fast mri. *Magnetic Resonance in Medicine: An Official Journal of the International Society for Magnetic Resonance in Medicine*, 42(5):952–962, 1999.
- [48] Benjamin Recht, Rebecca Roelofs, Ludwig Schmidt, and Vaishaal Shankar. Do imagenet classifiers generalize to imagenet? In *International Conference on Machine Learning*, pages 5389–5400. PMLR, 2019.

- [49] Philip M Robson, Aaron K Grant, Ananth J Madhuranthakam, Riccardo Lattanzi, Daniel K Sodickson, and Charles A McKenzie. Comprehensive quantification of signal-to-noise ratio and g-factor for image-based and k-space-based parallel imaging reconstructions. *Magnetic Resonance in Medicine: An Official Journal of the International Society for Magnetic Resonance in Medicine*, 60(4):895–907, 2008.
- [50] Peter B Roemer, William A Edelstein, Cecil E Hayes, Steven P Souza, and Otward M Mueller. The nmr phased array. *Magnetic resonance in medicine*, 16(2):192–225, 1990.
- [51] Yaniv Romano, Michael Elad, and Peyman Milanfar. The little engine that could: Regularization by denoising (red). *SIAM Journal on Imaging Sciences*, 10(4):1804–1844, 2017.
- [52] Olaf Ronneberger, Philipp Fischer, and Thomas Brox. U-net: Convolutional networks for biomedical image segmentation. In *International Conference on Medical image computing and computer-assisted intervention*, pages 234–241. Springer, 2015.
- [53] Mehdi Sajjadi, Mehran Javanmardi, and Tolga Tasdizen. Regularization with stochastic transformations and perturbations for deep semi-supervised learning. *Advances in neural information processing systems*, 29:1163–1171, 2016.
- [54] Christopher M. Sandino, Joseph Y. Cheng, Feiyu Chen, Morteza Mardani, John M. Pauly, and Shreyas S. Vasanawala. Compressed sensing: From research to clinical practice with deep neural networks: Shortening scan times for magnetic resonance imaging. *IEEE Signal Process. Mag.*, 37(1):117–127, 2020. doi: 10.1109/MSP.2019.2950433. URL <https://doi.org/10.1109/MSP.2019.2950433>.
- [55] Christopher M Sandino, Peng Lai, Shreyas S Vasanawala, and Joseph Y Cheng. Accelerating cardiac cine mri using a deep learning-based espirit reconstruction. *Magnetic Resonance in Medicine*, 85(1):152–167, 2021.
- [56] Anuroop Sriram, Jure Zbontar, Tullie Murrell, Aaron Defazio, C Lawrence Zitnick, Nafissa Yakubova, Florian Knoll, and Patricia Johnson. End-to-end variational networks for accelerated mri reconstruction. In *International Conference on Medical Image Computing and Computer-Assisted Intervention*, pages 64–73. Springer, 2020.
- [57] Rohan Taori, Achal Dave, Vaishaal Shankar, Nicholas Carlini, Benjamin Recht, and Ludwig Schmidt. Measuring robustness to natural distribution shifts in image classification. *arXiv preprint arXiv:2007.00644*, 2020.
- [58] Martin Uecker, Peng Lai, Mark J Murphy, Patrick Virtue, Michael Elad, John M Pauly, Shreyas S Vasanawala, and Michael Lustig. Espirit—an eigenvalue approach to autocalibrating parallel mri: where sense meets grappa. *Magnetic resonance in medicine*, 71(3):990–1001, 2014.
- [59] Muhammad Usman, Siddique Latif, Muhammad Asim, Byoung-Dai Lee, and Junaid Qadir. Retrospective motion correction in multishot mri using generative adversarial network. *Scientific reports*, 10(1):1–11, 2020.
- [60] Zhou Wang, Alan C Bovik, Hamid R Sheikh, and Eero P Simoncelli. Image quality assessment: from error visibility to structural similarity. *IEEE transactions on image processing*, 13(4):600–612, 2004.
- [61] Qizhe Xie, Zihang Dai, Eduard Hovy, Thang Luong, and Quoc Le. Unsupervised data augmentation for consistency training. *Advances in Neural Information Processing Systems*, 33, 2020.
- [62] Burhaneddin Yaman, Seyed Amir Hossein Hosseini, Steen Moeller, Jutta Ellermann, Kâmil Uğurbil, and Mehmet Akçakaya. Self-supervised physics-based deep learning mri reconstruction without fully-sampled data. In *2020 IEEE 17th International Symposium on Biomedical Imaging (ISBI)*, pages 921–925. IEEE, 2020.
- [63] Leslie Ying and Jinhua Sheng. Joint image reconstruction and sensitivity estimation in sense (jsense). *Magnetic Resonance in Medicine: An Official Journal of the International Society for Magnetic Resonance in Medicine*, 57(6):1196–1202, 2007.
- [64] Maxim Zaitsev, Karl Zilles, and Nadim Joni Shah. Shared k-space echo planar imaging with keyhole. *Magnetic Resonance in Medicine: An Official Journal of the International Society for Magnetic Resonance in Medicine*, 45(1):109–117, 2001.
- [65] Jure Zbontar, Florian Knoll, Anuroop Sriram, Tullie Murrell, Zhengnan Huang, Matthew J Muckley, Aaron Defazio, Ruben Stern, Patricia Johnson, Mary Bruno, et al. fastmri: An open dataset and benchmarks for accelerated mri. *arXiv preprint arXiv:1811.08839*, 2018.



## A Glossary

Table 3 provides the notation used in the paper.

**Table 3:** Summary of notation used in this work.

	Notation	Description
<b>MRI forward model</b>	$x, y$	Image, k-space measurements
	$y_i^{(s)}, y_i^{(u)}$	Fully-sampled (supervised) k-space, prospectively undersampled (unsupervised) k-space
	$\Omega, F, S$	Undersampling mask, fourier transform matrix, coil sensitivity maps
	$A$	The forward MRI acquisition operator
	$\epsilon$	Additive complex-valued Gaussian noise
<b>Augmentation transforms</b>	$T$	Set of data transforms
	$T_I, T_E$	Set of invariant and equivariant data transforms
	$g, g_E, g_I$	Transform, equivariant transform, invariant transform
	$G_E, G_I$	Sequence of sampled invariant and equivariant data transforms
	$\mathcal{N}(0, \sigma)$	Complex gaussian distribution with zero-mean, variance $\sigma^2$
	$\alpha$	Motion-induced phase error amplitude
	$\phi_i^k$	Phase error for $k^{\text{th}}$ phase encode line in example $i$
	$\mathcal{R}(\cdot)$	Range
	$\beta(t)$	Difficulty scale
	LM, HM	Light motion ( $\alpha=0.2$ ), heavy motion ( $\alpha=0.4$ )
	LN, HN	Light noise ( $\sigma=0.2$ ), heavy noise ( $\sigma=0.4$ )
<b>Model components and losses</b>	$\mathcal{L}_{sup}, \mathcal{L}_{cons}$	Supervised, consistency loss
	$\lambda$	Consistency loss weight
	$R_i$	U-Net resolution level $i$

## B Extended Related Work

In this section, we summarize the key differences between VORTEX and prior work in data augmentations (i.e. MRAugment) and in consistency training (i.e. Noise2Recon). Specifically, we highlight two advantages of VORTEX:

1. **Image-based and Acquisition-based Augmentations.** VORTEX can relax the assumption that augmentations must preserve the noise statistics of the data [21]. This allows VORTEX to leverage both image-based and acquisition-based augmentations, which do not preserve the noise statistics of the data.
2. **Regularization Beyond Noise.** VORTEX can leverage physics-driven augmentations beyond the standard denoising regularization used in prior work in both consistency [18] and pre-training [51]. Thus, it is feasible to extend VORTEX to other relevant clinical artifacts while maintaining the regularization properties of the well-studied denoising task.

### B.1 VORTEX vs MRAugment

MRAugment proposes a framework for applying image-based augmentations on fully-supervised training data. This approach showed improved performance in data-limited settings, which may suggest the family of image-based augmentations are helpful in reducing model overfitting. It also suggests scheduling the likelihood of applying an augmentation can be helpful for reducing the number of augmented examples in early stages of training.

**Image vs Acquisition Augmentations.** MRAugment focuses on the use of image-based augmentations for supervised training. In VORTEX, both image-based and MRI acquisition-based augmentations are used for semi-supervised consistency training to 1) reduce dependence on supervised training data and 2) increase robustness to physics-driven perturbations that are frequently observed during MRI acquisition.

**Relaxing Assumption of Preserved Noise Statistics.** MRAugment notes that the family of image-based augmentations were selected to ensure that noise statistics of the training data were preserved. However, this constraint excludes acquisition-based augmentations, particularly noise and motion, which are needed to build robustness to noise and motion artifacts in MRI. However, these acquisition-based augmentations inherently change the effective noise floor (and thus SNR) of the scan, and thus violate this constraint. We empirically validate this claim in supervised settings, where acquisition-based augmentations perform worse than standard supervised training in in-distribution settings (Table 2). This tradeoff between in-distribution performance and OOD robustness would preclude the application of acquisition-based augmentations in practice.

However, with VORTEX, not only is this tradeoff mitigated but the performance in both in-distribution and OOD settings is significantly improved (Table 10). This improved performance empirically demonstrates that the assumption that augmentations must preserve noise statistics can be relaxed in the VORTEX framework. Thus, both image-based and acquisition-based augmentations can be leveraged simultaneously, which leads to improvements in performance over either family of augmentations alone (Table 10).

**Precomputing Coil Sensitivity Maps.** Integrating coil sensitivity maps is standard clinical practice to help constrain the optimization problem for MRI image reconstruction [55, 49, 50]. MRAugment utilizes the end-to-end VarNet, which *learns* to jointly estimate coil sensitivities and reconstruct images [56]. Thus, the augmentation pipeline in MRAugment does not need to explicitly account for the effect of image-based transformations on sensitivity maps. It also has the added benefit of optimizing sensitivity map estimation with respect to augmented data. In practice, precomputing coil sensitivities is feasible and routine with sensitivity map estimation methods such as ESPIRiT [58] and JSENSE [63]. Additionally, precomputed maps are important in multi-coil datasets where the number of coils are not constant across different scans, which is critical when patients with heterogenous anatomies are being imaged [19].

VORTEX utilizes precomputed sensitivity maps estimated from auto-calibration regions in each scan. Because image-based augmentations are designed to emulate shifts in the imaging target, they also impact the coil geometry and sensitivity maps that are estimated. In contrast to the MRAugment sensitivity map formulation, which assumes sensitivity maps are fixed, VORTEX integrates physics-based modeling to appropriately warp sensitivity maps based on image-based augmentations. Given some equivariant image-based transform  $g_E$ , the augmented image for coil  $i$  ( $\tilde{x}_i$ ) can be defined as

$$\tilde{x}_i = g_E(\mathbf{S}_i)g_E(x)$$

**Scheduling Augmentation Difficulty.** MRAugment and VORTEX also differ in the mechanism of how augmentations are scheduled. MRAugment proposes an augmentation scheduling method that schedules the probability of applying an augmentation. Thus, training can occur predominantly on collected data in earlier stages of training and augmentations can help reduce overfitting at later training stages.

VORTEX is designed to build robustness to OOD perturbations, where the *extent* (and, more generally, *difficulty*) of these perturbations will be unknown at test time. In this framework, augmentations must not only function as a regularization method for improved performance on in-distribution data, but also appropriately model a separate distribution of data with respect to which the model can be trained. Thus, the model must learn to jointly optimize for both in-distribution (default training data) and OOD (perturbation-corrupt data) examples simultaneously. Intuitively, we need to design an augmentation scheduling scheme that will allow the model to gradually learn to generalize to higher extents (more difficult) perturbations over time while still ensuring examples from both distributions are sampled for joint optimization. To ensure that augmentations are always applied but at different extents, we propose a curriculum learning strategy for scheduling the *difficulty* of the augmentation.

## B.2 VORTEX vs Noise2Recon

Noise2Recon proposes a semi-supervised consistency based framework for joint denoising and reconstruction. This approach showed improved performance in label-limited settings, where the training dataset consists of both supervised and unsupervised data. VORTEX 1) extends this consistency training paradigm to a broader family of acquisition-based perturbations, 2) exhaustively studies how this framework can be leveraged for *both* image and acquisition-based

augmentations, and 3) proposes a curriculum learning strategy to gradually increase reconstruction difficulty.

**Robustness to Motion.** Noise2Recon proposes a novel consistency framework for semi-supervised MRI reconstruction but solely focuses on applications to noise artifacts. While denoising is a well-known regularizer for inverse problems [51, 7], many other acquisition-related artifacts in MRI are commonplace. In VORTEX, we explore the utility of motion augmentations as 1) a regularizer to improve robustness in label-limited settings and 2) a method to increase robustness to OOD motion artifacts. We demonstrate that motion artifact removal is as effective of a regularizer as denoising (Tables 1 and 2).

**Composing Augmentations for Multi-Artifact Correction.** Existing MRI artifact correction or removal methods, including Noise2Recon, separately handle reconstruction and artifact removal tasks, are limited to correcting for a single artifact, or require multiple unique workflows to correct for different artifacts [59, 38, 30]. However, in practice, effects of multiple acquisition-related artifacts can be compounded even in accelerated MRI. Thus a unified framework for removing these artifacts is desirable. VORTEX establishes a framework for both image-based and acquisition-based augmentations that can be utilized to jointly reconstruct and remove multiple artifacts with a single approach.

**Curriculum Learning for Augmentations.** VORTEX extends basic consistency training to include a scheduling protocol for increasing the difficulty of augmentations over the training cycle. Results demonstrate that designing curricula for augmentations in the consistency framework can lead to considerable performance improvements in OOD settings without losing performance among in-distribution scans (Table 5). Such curricula can be helpful for joint optimization of both artifactual and artifact-free images, particularly when example difficulty is extensive [8].

### B.3 Summary of Technical Contributions

In this work, we characterize the interface between physics-based MRI acquisition-motivated and image-based augmentations to 1) reduce label dependency and 2) increase robustness to clinically-relevant distribution shifts that are pervasive during MRI acquisition. We extend the semi-supervised consistency framework in Noise2Recon to handle both acquisition and image based perturbations in a way that is motivated by the physics-driven forward model of MRI acquisition. To ensure that we are inclusive of a broader family of acquisition-based perturbations than was available in Noise2Recon, we propose extending the semi-supervised consistency framework to handle motion, a common artifact in MRI. We exhaustively study the interaction between physics/acquisition based and image based augmentations in both fully supervised training with augmentations and semi-supervised training with the proposed consistency.

## C Equivariant and Invariant Transforms

In this section, we provide an extended discussion of the choice of and interaction between equivariant and invariant transforms.

**Equivariance.** We simplify the precise definition of equivariance that requires group theory [10] to denote  $f_\theta(g(x)) = g(f_\theta(x))$  for all  $g \in T$  where  $T$  is the set of data augmentation transformations. Intuitively, if a trained model  $f_\theta$  is equivariant to a transformation  $g$ , then the transformation of the input directly corresponds to the transformation of the model output.

**Invariance.** Similarly, we simplify the definition of invariance to  $f_\theta(g(x)) = f_\theta(x)$  for all  $g \in T$  where  $T$  is the set of transformations we use for data augmentation. Intuitively,  $f_\theta$  is invariant to transformation  $g$  if the output of the model does not change upon applying  $g$  to the input. Details on how these definitions motivate the structure of augmentations in VORTEX are provided below.

**Choosing Equivariance or Invariance.** It is important to note that, practically, specifying to which transforms the model should be equivariant or invariant is a design choice and often task-dependent. In the case of MRI, image-based augmentations proposed in MRAugment are meant to simulate differences in patient positioning and spatial scan parameters (e.g. field-of-view, nominal resolution). The differences are typically prescribed at scan time (i.e. scan parameters) or are correctable prior to the scan. In contrast, motion and noise are perturbations that occur *during acquisition*, and therefore cannot be corrected a priori. Thus, building networks that are invariant to these perturbations are critical. Based on this paradigm of transforms in MRI, spatial image transforms are classified as equivariant transforms while the physics-based transforms we propose are classified as invariant transforms.

**Composing Transforms (Extended).** Section 4.1 introduces the intuition for equivariant and invariant transformations. In this section, we formalize how transforms from these families are composed.

Let  $g_1, \dots, g_K$  be an ordered sequence of unique transforms sampled from a set of transforms  $T$ . Let  $\bar{G}_E, \bar{G}_I$  be the sequence of sampled equivariant and invariant transforms, respectively. Thus,  $\bar{G}_E = (g_i \text{ if } g_i \in T_E \forall i = 1, \dots, K)$  (similarly for  $\bar{G}_I$ ). Let  $G_E$  and  $G_I$  be the compositions of each transform in  $\bar{G}_E, \bar{G}_I$ , respectively. Thus,  $G_E = \bar{G}_{E|\bar{G}_E|} \circ \dots \circ \bar{G}_{E1}$  (similarly for  $G_I$ ).

As a design choice, we select all physics-driven, acquisition-related transforms to be in the family of invariant transforms. This choice is made to ensure reconstructions are invariant to plausible acquisition-related perturbations. Thus, the family of physics-driven transforms are synonymous with the family of invariant transforms for our purposes.

Because signal from physics-driven perturbations (noise and motion) is sampled at acquisition, these perturbations are applied after undersampling in the supervised augmentation methods, where fully-sampled data is available. Thus, the consistency loss for example  $i$  ( $\mathcal{L}_{cons,i}$ ) in Eq. 1 can be generalized to multiple composed transforms (Eq. 3).

$$\mathcal{L}_{cons,i} = \|G_E(f_\theta(y_i^u, A)) - f_\theta(G_I \circ G_E(y_i^u), A)\|_1 \quad (3)$$

## D Experimental Details

### D.1 Dataset

In this section, we provide details for the two datasets used in this study: the mridata 3D FSE knee dataset and the fastMRI 2D multi-coil brain dataset.

#### D.1.1 mridata 3D FSE Knee Dataset

**Dataset Splits.** The mridata 3D FSE knee dataset consists of 6080 fully-Cartesian-sampled knee slices (19 scans) from healthy participants. The dataset was randomly partitioned into 4480 slices (14 scans) for training, 640 slices (2 scans) for validation, and 960 slices (3 scans) for testing.

**Simulating Data-Limited and Label-Limited Settings.** In this study, we evaluate all methods in the data-limited and label-limited regimes, where supervised examples are scarce compared to unsupervised (undersampled) examples. To simulate this scenario, a subset of training scans are retrospectively undersampled using fixed undersampling masks, resulting in unsupervised training examples. To limit the total (supervised and unsupervised) amount of available training data, we train with only 6 of the 14 training scans, where 1 scan is supervised and 5 scans are unsupervised.

**K-space Hybridization and Sensitivity Maps.** 3D FSE scans were acquired in 3D, resulting in Fourier encoded signal along all dimensions ( $k_x \times k_y \times k_z$ ). Because the readout dimension  $k_x$  is fully-sampled in these scans, scans were decoded along the  $k_x$  dimension, resulting in a hybridized k-space as mentioned in Section 4. All sensitivity maps were estimated with JSENSE as implemented in SigPy [44], with a kernel width of 8 and a  $20 \times 20$  center k-space auto-calibration region.

**Table 4:** Data augmentation configuration for mridata 3D FSE knee dataset experiments.  $p$  is the effective probability of applying an augmentation. In MRAugment, this is equivalent to the base probability multiplied by the weighting factor. Acquisition-based augmentations were configured in separate experiments at both light and heavy settings.

Kind	Transform	Parameters	p
Image	H-Flip	N/A	0.275
	V-Flip	N/A	0.275
	$k \times 90^\circ$ rotation	$k \in \{2\}$	0.275
	Rotation	$[-180^\circ, 180^\circ]$	0.275
	Translation	$[-10\%, 10\%]$	0.55
	Scale	$[0.75, 1.25]$	0.55
	Shear	$[-15^\circ, 15^\circ]$	0.55
Acquisition	Gaussian Noise	$\sigma=[0.1,0.3]$ (light) $\sigma=[0.2,0.5]$ (heavy)	0.2
	Motion	$\alpha=[0.1,0.3]$ (light) $\alpha=[0.2,0.5]$ (heavy)	0.2

**Mask Generation.** Scans for training and evaluation were undersampled using 2D Poisson Disc undersampling, a compressed sensing-motivated pattern for 3D Cartesian imaging. Given an acceleration rate  $R$ , undersampling masks were generated in the  $k_y \times k_z$  dimensions for all scans such that the number of pixels sampled would be approximately  $\frac{|k_y||k_z|}{R}$ . To maintain consistency with generated sensitivity maps, a  $20 \times 20$  center k-space auto-calibration region was used when constructing undersampling masks for all examples. To simulate prospectively undersampled acquisitions, scans were retrospectively undersampled with a fixed 2D Poisson Disc undersampling pattern [9]. Following Cartesian undersampling convention, all  $k_y \times k_z$  slices for a single scan are undersampled with the identical 2D Poisson Disc mask. This procedure was used for both simulating prospectively undersampled scans during training (i.e. unsupervised examples) and evaluation. All undersampling masks for testing and simulating undersampled k-space are generated with an unique, fixed random seed for each scan to ensure reproducibility.

### D.1.2 fastMRI Brain Multi-coil Dataset

**Dataset Splits.** The distributed validation split of the fastMRI 2D brain multi-coil dataset was divided into 757 scans for training, 207 scans for validation, and 414 scans for testing. To control for confounding variables when comparing performance between reconstruction methods, all data splits were filtered to include only T2-weighted scans acquired at a 3T field strength, resulting in 266, 70, and 137 scans for training, validation, and testing, respectively. Data-limited and label-limited training settings were simulated by limiting training data to 18 supervised and 36 unsupervised scans and validation data to 50 scans.

**Sensitivity Maps.** Like for mridata, sensitivity maps were estimated using JSENSE with a kernel width of 8 and calibration region of  $12 \times 12$ . This calibration region corresponds to the 4% auto-calibration region used for 8x undersampling.

**Mask Generation.** Scans for training and evaluation were undersampled using 1D random undersampling, a compressed sensing-motivated pattern for 2D Cartesian imaging [39]. Given an acceleration rate  $R$ , undersampling masks were generated in the  $k_y$  phase-encode dimension for all scans such that the number of readout lines sampled would be approximately  $\frac{|k_y|}{R}$ . Training and evaluation was conducted at  $R = 8$  acceleration with a 4% auto-calibration region. Like in mridata, fixed undersampling masks were generated to simulate prospectively undersampled data and for the testing data to ensure reproducibility.

## D.2 Baselines

We compared VORTEX to state-of-the-art supervised, supervised augmentation, and self-supervised MRI reconstruction baselines. We provide an overview of these methods and their notation in the main text. Hyperparameters for all methods are provided in Appendix D.3.1.

**Supervised.** We compared VORTEX to standard supervised training without augmentations (termed *Supervised*). In supervised training, fully-sampled scans are retrospectively undersampled. The model is trained to reconstruct the fully-sampled scan from its undersampled counterpart. Note, in supervised settings, only fully-sampled scans can be used for training. Any prospectively undersampled (unsupervised) scans cannot be leveraged in this setup.

**Supervised+Augmentation (*Aug*) and MRAugment.** Supervised baselines with augmentation (termed *Aug*) were trained with image and/or physics-based augmentations, which are denoted by parentheses. Image-based augmentations were applied prior to the retrospective undersampling, following the MRAugment protocol. Physics-based acquisition augmentations were applied after this undersampling to model the MRI data acquisition process. For example *Aug (Motion)* indicates a supervised method trained with motion augmentations. Image-based augmentations were identical to those used in MRAugment. As such, *Aug (Image)* is equivalent to MRAugment, and is referred to as such for readability.

**SSDU.** We also compared VORTEX to the state-of-the-art self-supervised learning via data undersampling (SSDU) baseline [62]. This method was originally proposed for fully unsupervised learning, in which all training scans are prospectively undersampled. We propose a trivial extension to adapt it for the semi-supervised setting. In cases of prospectively undersampled (unsupervised) data, the training protocol proposed in SSDU was used. Fully-sampled (supervised) data was retrospectively undersampled using the undersampling method and acceleration for the specified experiment. These simulated undersampled scans were used as inputs to the SSDU protocol. Because the retrospective undersampling is done dynamically (i.e. each time a supervised example is sampled), it may serve as a method of augmenting supervised scans.

**Compressed Sensing (CS).** As a non-DL baseline, we used  $\ell_1$ -wavelet regularized CS [39]. Note, because of the time-intensive nature of CS-based reconstructions, CS is difficult to scale for reconstructing large datasets like the fastMRI brain dataset. Thus, we only evaluate CS on the mridata 3D FSE knee dataset.

## D.3 Training Details

All code is written in Python with PyTorch 1.6 and is available at <https://github.com/ad12/meddlr>.

### D.3.1 Hyperparameters

**Architecture and Optimization.** All models used a 2D U-Net architecture with [52] with 4 pooling layers. Convolutional block at depth  $d$  consisted of two convolutional layers with  $32^d$  channels for  $d = \{1, \dots, 5\}$ . All models were trained with the Adam optimizer with default parameters ( $\beta_1=0.9$ ,  $\beta_2=0.999$ ) and weight decay of  $1e-4$  for 200 epochs [31, 37]. Training was conducted with an effective training batch size of 24 and learning rate  $\eta=1e-3$ . All models used VORTEX methods used 1:1 balanced sampling between supervised and unsupervised examples [18]. In the VORTEX consistency pathway, gradients from the consistency loss ( $L_{cons}$ ) only flowed through reconstructions of the augmented input.

**Aug Baselines and MRAugment.** Supervised augmentation baselines were trained with image-based and acquisition-based augmentations. Image-based augmentations for each dataset followed the augmentation configuration provided in the MRAugment. With the mridata 3D FSE knee dataset, integer rotations could only be conducted at 180 degrees due

**Table 5:** Comparison of different scheduling methods and warmup periods on the mridata knee multi-coil dataset with heavy motion augmentations. All scheduling methods outperform non-scheduled training (base). There is no advantage of a specific scheduling protocol, suggesting that some curriculum is better than none.

Perturbation	None		Motion (light)		Motion (heavy)	
Curricula	SSIM	cPSNR (dB)	SSIM	cPSNR (dB)	SSIM	cPSNR (dB)
None	0.861	36.4	0.855	35.8	0.819	33.2
Linear (20e)	0.866	36.4	0.862	35.8	<b>0.828</b>	33.3
Linear (100e)	0.877	36.3	<b>0.871</b>	35.8	0.822	32.6
Linear (200e)	0.869	36.4	0.865	35.8	0.817	32.7
Exp (20e, $\gamma = 5$ )	0.865	<b>36.4</b>	0.857	<b>35.9</b>	0.822	<b>33.4</b>
Exp (100e, $\gamma = 5$ )	0.864	36.3	0.857	35.8	0.812	33.2
Exp (200e, $\gamma = 5$ )	<b>0.877</b>	36.4	0.867	35.8	0.812	32.3

to the anisotropic matrix shape of the  $ky \times kz$  slice. *Aug* baselines using physics-driven acquisition-based augmentations used a maximum probability of  $p = 0.2$  as recommended by Desai et al. [18], and use the same range of  $\sigma$  for noise and  $\alpha$  for motion that are used in the corresponding VORTEX experiments. Augmentations, their parameters, and their effective probabilities used for the mridata 3D FSE knee dataset are listed in Table 4. All augmentation methods were trained with the exponential augmentation probability scheduler with  $\gamma = 5$  and a scheduling period equivalent to the training length as proposed by Fabian et al. [21].

**SSDU.** SSDU is sensitive to the loss function and masking extent ( $\rho$ ). Thus, these hyperparameters that should be optimized for different datasets. We swept through loss functions k-space  $\ell_1$ , k-space  $\ell_1$ - $\ell_2$ , and image  $\ell_1$  and masking extent  $\rho = 0.2, 0.4, 0.6$ . Models with the highest validation cPSNR were selected for all SSDU experiments. For the mridata 3D FSE knee dataset, the configuration with loss function k-space  $\ell_1$  and  $\rho = 0.4$  was used. For the fastMRI multi-coil brain dataset, the configuration with normalized  $\ell_1$ - $\ell_2$  loss in k-space and  $\rho = 0.2$  was used.

**Compressed Sensing (CS).** Because CS is sensitive to the regularization parameter  $\lambda$ , this parameter must be carefully tuned for different motion and noise settings. To hand-tune  $\lambda$  for both light and heavy motion levels, we swept through  $\lambda$  values within the range  $[0, 0.5]$  and selected the optimal  $\lambda$  that balances reconstruction quality and artifact mitigation. For the 3D FSE mridata knee dataset,  $\lambda = 0.1$  was used for light motion, and  $\lambda = 0.15$  was used for heavy motion at 12x and 16x acceleration. For in-distribution (no noise, no motion), light noise, and heavy noise settings, we use the regularization parameters suggested by Desai et al. [18].

**Consistency Augmentations in VORTEX.** Like *Aug* baselines, VORTEX was trained with combinations of image and physics-based augmentations. We use the same parenthetical nomenclature to indicate the augmentation type used in the consistency branch (e.g. *VORTEX (Motion)* for motion consistency). The family of image augmentations used for consistency in VORTEX were identical to those used in MRAugment. Physics-based consistency augmentations were sampled from either the light ( $\mathcal{R}(\cdot)=[0.1, 0.3]$ ) or heavy ( $\mathcal{R}(\cdot)=[0.2, 0.5]$ ) range during training.

## D.4 Evaluation

### D.4.1 Evaluation Settings

We perform evaluation in both in-distribution and clinically-relevant, simulated OOD settings. In-distribution evaluation consisted of evaluation on the test set described in D.1.

For OOD evaluation, we considered two critical settings that have been shown to affect image quality: (1) decrease in SNR and (2) motion corruption. The extent of the distribution shift is synonymous with the difficulty level for each perturbation ( $\sigma$  for noise,  $\alpha$  for motion), where larger difficulty levels indicate correspond to larger shifts. Thus, we

**Table 6:** Impact of training duration on cPSNR of supervised methods without augmentations (*Supervised*), supervised methods with motion augmentations (*Aug (Motion)*), MRAugment, and VORTEX with motion consistency (*VORTEX (Motion)*). Training duration are percentages of the full training duration (200 epochs). Asterisk (\*) indicates the default training configuration. Both supervised augmentation methods and MRAugment are more sensitive to training time than Supervised or VORTEX methods. Supervised underperforms Aug, MRAugment, and VORTEX. VORTEX achieves highest performance and is insensitive to training duration relative to the other methods.

Model	Perturbation		
	None	Motion (light)	Motion (heavy)
Supervised (10%)	35.0	33.3	27.4
Supervised (25%)	35.3	32.3	27.1
Supervised (50%)	35.5	32.0	26.4
Supervised (100%)*	35.8	33.6	27.0
Supervised (200%)	36.0	33.9	27.6
Supervised (300%)	36.0	33.9	27.6
<hr/>			
MRAugment (10%)	35.4	32.3	26.0
MRAugment (25%)	35.8	31.5	25.1
MRAugment (50%)	36.0	31.5	24.3
MRAugment (100%)	36.2	31.8	24.0
MRAugment (200%)	36.3	32.2	24.3
MRAugment (300%)	36.4	33.4	25.0
<hr/>			
Aug (Motion) (10%)	34.8	33.9	30.8
Aug (Motion) (25%)	35.3	34.5	31.4
Aug (Motion) (50%)	35.4	34.6	31.1
Aug (Motion) (100%)*	35.9	35.1	31.5
Aug (Motion) (200%)	36.0	35.1	30.8
Aug (Motion) (300%)	36.0	35.2	32.1
<hr/>			
VORTEX (Motion) (10%)	36.2	35.5	32.4
VORTEX (Motion) (25%)	36.3	35.7	33.1
VORTEX (Motion) (50%)	36.4	35.8	33.2
VORTEX (Motion) (100%)*	36.4	35.8	33.2
VORTEX (Motion) (200%)	36.3	35.7	33.0
VORTEX (Motion) (300%)	36.3	35.7	33.0

define *low* and *heavy* noise and motion difficulty levels for evaluation – low noise  $\sigma=0.2$ , heavy noise  $\sigma=0.4$ , low motion  $\alpha=0.2$ , heavy motion  $\alpha=0.4$ . These values are selected based on visual inspection of clinical scans (see 4.1.1). Note, by definition ( $\sigma=0$ ,  $\alpha=0$ ) corresponds to the in-distribution evaluation.

Given difficulty levels for motion and noise, each scan was perturbed by a noise or phase error (motion) maps generated with a set difficulty level. These perturbations were fixed for each testing scan to ensure reproducibility and identical perturbations in the test set across different experiments.

In the text, we refer to different evaluation configurations as *perturbations*. *None* indicates the in-distribution setting. *LN*, *HN*, *LM*, *HM* correspond to light noise, heavy noise, light motion, and heavy motion, respectively.

#### D.4.2 Metric Selection

Conventional computational imaging uses magnitude metrics for quantifying image quality. However, MRI images contain both magnitude and phase information (i.e. real and imaginary components). Because phase-related errors may not be captured by magnitude metrics, we use a combination of complex and magnitude metrics – complex PSNR (cPSNR) and magnitude SSIM – to quantify image quality. Equation 4 defines the cPSNR formulation for complex-valued ground truth  $x_{ref}$  and prediction  $x_{pred}$ .  $\|\cdot\|_2$  corresponds to the complex- $\ell_2$  norm and  $|\cdot|$  denotes the magnitude of complex-valued input. Additionally, SSIM has shown to be a better corollary for MRI reconstruction quality compared to pSNR on magnitude images [32]. Thus, we use SSIM to quantify magnitude image quality.

$$\text{cPSNR (dB)} = 20 \log_{10} \frac{\max |x_{ref}|}{\|x_{pred} - x_{ref}\|_2} \quad (4)$$

By default, metrics were computed over the full 3D scan. An additional set of metrics were also computed per reconstructed slice (termed *slice metrics*). Because different slices have different extents of relevant anatomy, per-slice



**Table 7:** Ablation for acceleration – 12x vs 16x. Similar to results at 16x acceleration, *VORTEX (Motion)* outperformed supervised methods, and MRAugment at 12x acceleration. This may suggest that VORTEX is broadly applicable to different acceleration levels.

Aug Range	Perturbation	None		Motion (light)		Motion (heavy)	
	Model	SSIM	cPSNR (dB)	SSIM	cPSNR (dB)	SSIM	cPSNR (dB)
N/A	Supervised 12x	0.814	36.2	0.814	32.4	0.689	25.4
	Supervised 16x	0.798	35.8	0.809	33.6	0.706	27.0
N/A	MRAugment 12x	0.828	36.5	0.814	31.9	0.637	23.6
	MRAugment 16x	0.811	36.2	0.793	31.8	0.660	24.0
N/A	SSDU 12x	0.819	34.9	0.816	34.5	0.762	30.9
	SSDU 16x	0.787	34.9	0.783	34.7	0.734	31.9
light	Aug (Motion) 12x	0.811	36.1	0.807	35.3	0.765	31.3
	Aug (Motion) 16x	0.802	35.6	0.793	34.7	0.739	30.4
heavy	Aug (Motion) 12x	0.818	36.1	0.811	35.2	0.758	31.2
	Aug (Motion) 16x	0.793	35.9	0.793	35.1	0.751	31.5
light	VORTEX Motion 12x	0.881	36.8	0.875	36.1	0.815	32.1
	VORTEX Motion 16x	0.882	36.4	0.875	35.7	0.813	31.5
heavy	VORTEX Motion 12x	0.888	36.7	0.883	36.1	0.846	33.5
	VORTEX Motion 16x	0.861	36.4	0.855	35.8	0.819	33.2

**Table 8:** Pixel-level vs. latent space consistency. **LM:** light motion; **HM:** heavy motion.

Model	cPSNR (dB)	SSIM	cPSNR (dB) (LM)	SSIM (LM)	cPSNR (dB) (HM)	SSIM (HM)
Supervised	35.8	0.798	33.6	0.809	27.1	0.706
Pixel-Level	36.4	0.873	35.9	0.866	33.2	0.828
$R_4$	36.4	0.877	34.7	0.865	29.8	0.778
$R_3, R_4$	36.4	0.873	34.0	0.852	30.1	0.781
$R_2, R_3, R_4$	36.3	0.873	34.4	0.854	29.5	0.769
$R_1, R_2, R_3, R_4$	36.3	0.875	34.7	0.864	30.3	0.775

metrics can provide a more nuanced comparison of 2D slice reconstructions among different methods. Statistical comparisons were conducted using Kruskal-Wallis tests and corresponding Dunn posthoc tests with Bonferroni correction ( $\alpha=0.05$ ). All statistical analyses were performed using the SciPy library.

## E Ablations


We perform ablations to understand two design questions for key components in our framework: (1) Can consistency be enforced at different points in the network; (2) How should example difficulty be specified during training. All methods use the default configurations specified in Appendix D.3. To evaluate each piece thoroughly, we consider augmentation and VORTEX approaches trained with heavy motion perturbations.

### E.1 Latent Space vs Pixel-level Consistency

We compare enforcing consistency at the pixel-level output image versus learned latent representations at varying U-Net resolution levels. Let  $R_k$  be  $k^{th}$  resolution level at which consistency is enforced, where  $k \in \{1, 2, 3, 4\}$  since our U-Net architecture had 4 pooling layers. For the  $k^{th}$  resolution level, we enforce consistency after the final convolution in the encoder, and after the transpose convolution in the decoder. For  $k = 4$ , consistency is enforced at the bottleneck layer, after the convolution in the encoder. To control for the impact of loss weighting, we normalize  $\lambda$  by the number of consistency losses that are computed in latent space when consistency is enforced at multiple resolution levels  $R_k$ .

We compare these approaches in the case of light and heavy motion.


We find that latent space consistency performed similarly across all resolution levels, outperforming the *Supervised* baseline on both in- and out-of-distribution data (Table 8). For in-distribution data, latent space consistency at any resolution level performed on par with pixel-level consistency. However, for OOD data, it performed considerably worse than pixel-level consistency, by at least 0.047 SSIM and 3.2dB cPSNR under heavy motion. Although not common in the consistency training literature, we find that pixel-level consistency was a better technique for capturing the semantics of global distribution shifts such as motion for accelerated MRI reconstruction, which might occur at the pixel-level.



**Figure 4:** Ablation for balanced sampling (*SB*) and augmentation curricula (*curr*) in *VORTEX (Motion)*.

## E.2 Augmentation Scheduling.

We seek to quantify the utility of scheduling augmentation difficulty in *VORTEX*’s consistency branch (see §4.2). We evaluate linear and exponential scheduling functions with different warm up schedules – 10%, 50%, and 100% of the training period. We show that curricula methods outperformed non-curricula methods for both in-distribution and OOD evaluation (see Table 5). However, no one curricula configuration outperformed others, which may indicate that all curricula methods are feasible ways to schedule augmentations. Curriculum learning is also compatible with the balanced sampling protocol proposed by Desai et al. [18], where supervised and unsupervised examples are sampled at a fixed ratio during training. Incorporating balanced sampling (*SB*) into training led to an increase in SSIM for both in-distribution and OOD light motion and light noise evaluation configurations (Fig. 4). Increase in SSIM may indicate that curricula can help the network gradually learn useful representations without a mode collapse into the trivial solution (i.e. image blurring), which is common for pixel-level losses.



**Figure 5:** Augmentation difficulty curricula over training period of 200 epochs using linear and exponential (*exp*) schedulers defined in Section 4.2. Time constant for exponential scheduling  $\tau = \frac{M}{\gamma}$  where  $\gamma=5$ .

## E.3 Training Time

We ablate the sensitivity of the performance of supervised, augmentation, and *VORTEX* methods to training duration. To compute the performance at different training duration, we select best checkpoints (quantified by validation cPSNR) up to a given duration and run evaluation using these weights. As all methods were trained for 200 epochs, we compare the performance at training times of 10% (20 epochs), 25% (50 epochs), 50% (100 epochs), and 100% (200 epochs). Supervised methods were insensitive to training time, but considerably underperformed both supervised augmentation (Aug) and *VORTEX* (Table 6). Augmentation based methods were sensitive to training time, with changes in cPSNR of  $>1$ dB. *VORTEX* achieved the highest performance across all metrics and evaluation setups and was relatively insensitive to training duration.

## E.4 Sensitivity to Acceleration Factors

We evaluated the performance of *VORTEX* at different acceleration factors in Table 7. At 12x acceleration, *VORTEX* trained with heavy motion recovered +0.061 SSIM and +0.8dB cPSNR compared to the *Supervised* baseline in the in-distribution setting. At the same acceleration, *VORTEX* also outperformed the *Supervised* baseline by +0.157 SSIM and +8.1dB cPSNR. The stability of *VORTEX* at different accelerations may indicate that *VORTEX* is generalizable across different acceleration extents.

**Table 9:** Average performance on mridata 3D FSE knee dataset with light motion ( $\alpha=0.2$ ) and noise ( $\sigma=0.2$ ) perturbations. Physics augmentations are compositions of noise and motion in their *heavy* ( $\mathcal{R}(\alpha) = \mathcal{R}(\sigma) = [0.2, 0.5]$ ) training difficulty configurations. Models are identical to those reported in Table 1.

Perturbation	Motion (light)		Noise (light)	
	SSIM	cPSNR (dB)	SSIM	cPSNR (dB)
Compressed Sensing	0.810 (0.024)	29.8 (1.7)	0.828 (0.002)	32.9 (0.2)
Supervised	0.809 (0.028)	33.6 (0.2)	0.830 (0.024)	33.8 (0.3)
MRAugment	0.793 (0.021)	31.8 (1.8)	0.793 (0.024)	33.0 (0.4)
SSDU	0.783 (0.025)	34.7 (0.2)	0.752 (0.024)	33.7 (0.3)
Aug (Physics)	0.785 (0.036)	35.0 (0.4)	0.761 (0.049)	34.2 (0.3)
Aug (Image+Physics)	0.782 (0.045)	<b>35.5 (0.4)</b>	0.750 (0.051)	34.6 (0.5)
VORTEX (Image)	0.831 (0.036)	32.5 (0.8)	<b>0.859 (0.004)</b>	33.6 (0.1)
VORTEX (Physics)	0.846 (0.025)	35.2 (0.4)	0.841 (0.035)	<b>34.7 (0.2)</b>
VORTEX (Image+Physics)	<b>0.849 (0.024)</b>	35.0 (0.5)	0.850 (0.028)	34.3 (0.1)

**Table 10:** Slice metrics (mean [standard deviation]) on the mridata knee dataset. Asterisk (\*) indicates significant performance of VORTEX over *all* baselines ( $p < 0.05$ ).

Perturbation Model	None		Motion (heavy)		Noise (heavy)	
	SSIM	cPSNR (dB)	SSIM	cPSNR (dB)	SSIM	cPSNR (dB)
Compressed Sensing	0.682 (0.122)	29.4 (3.2)	0.559 (0.135)	18.9 (2.4)	0.534 (0.102)	24.8 (2.5)
Supervised	0.635 (0.133)	29.7 (3.7)	0.545 (0.117)	21.4 (2.5)	0.591 (0.139)	25.8 (2.9)
MRAugment	0.653 (0.130)	30.1 (3.5)	0.505 (0.106)	18.6 (2.3)	0.563 (0.128)	25.0 (2.8)
SSDU	0.621 (0.147)	28.9 (3.4)	0.564 (0.146)	25.9 (3.5)	0.528 (0.142)	26.7 (2.8)
Aug (Physics)	0.623 (0.144)	29.6 (3.6)	0.566 (0.136)	26.0 (3.8)	0.557 (0.144)	27.6 (3.1)
Aug (Image+Physics)	0.618 (0.136)	30.1 (3.4)	0.565 (0.134)	<b>26.9 (3.8)</b>	0.540 (0.134)	27.9 (2.8)
VORTEX (Image)	0.718 (0.125)*	<b>30.4 (3.4)</b>	0.499 (0.110)	20.6 (2.3)	0.584 (0.104)	25.8 (2.5)
VORTEX (Physics)	<b>0.729 (0.138)*</b>	30.3 (3.4)	<b>0.628 (0.137)*</b>	26.0 (3.8)	0.653 (0.143)*	<b>28.1 (2.8)</b>
VORTEX (Image+Physics)	0.716 (0.131)*	30.3 (3.4)	0.616 (0.130)*	25.3 (3.7)	<b>0.658 (0.132)*</b>	27.5 (2.7)


## F Extended Results

In this section, we provide extended results for the mridata dataset using slice metrics and for the fastMRI multi-coil brain dataset [65].

### F.1 Extended mridata Results

**Light Perturbations.** Table 9 shows performance of baselines and VORTEX on the mridata knee dataset with light motion ( $\alpha=0.2$ ) and noise ( $\sigma=0.2$ ) perturbations. Like in heavy settings (Table 1), VORTEX methods consistently achieve higher SSIM than and comparable cPSNR to baselines. Improved SSIM values may indicate VORTEX can recover structural features in the image in cases of varying extents of perturbations. This trait may be useful for generalizing to data with unknown perturbation extents and may reduce the need for extensive hyperparameter search.

**Slice metrics.** Table 10 shows slice metrics of baselines and VORTEX on the mridata knee dataset. Among slice metrics, VORTEX also outperforms all baselines in both in-distribution and OOD settings. In particular, VORTEX significantly outperformed all baselines in SSIM in all evaluation settings ( $p < 0.05$ ). This may indicate that VORTEX has higher fidelity in recovering image structure even in OOD settings where perturbations can result in a considerable degradation in SSIM. To understand the efficacy of VORTEX in anatomically-dense regions of the image, slice metrics were also computed on the center 50% of axial slices (Fig. 6). Not only does VORTEX significantly outperform all baselines ( $p < 0.05$ ), it also had the least variance among all reconstruction methods. This may indicate VORTEX




**Figure 6:** SSIM at different perturbation levels on center 50% of slices. Asterisk (\*) indicates significant performance ( $p < 0.05$ ) over *all* baselines (Compressed Sensing, *Supervised*, MRAugment, SSDU, *Aug*). VORTEX methods significantly outperformed baselines in both in-distribution and OOD perturbation settings. VORTEX also had less variance across slices, which may indicate more consistent per-slice reconstruction than baseline methods.

**Table 11:** Test performance (mean [standard deviation]) on the fastMRI multi-coil brain dataset at 8x acceleration. Results are shown on both in-distribution data and different motion levels of  $\alpha = 0.4, 0.6, 0.8, 1.0$  for Supervised, SSDU, MRAugment, augmentation baselines, and VORTEX.

Perturbation	None		Motion ( $\alpha = 0.4$ )		Motion ( $\alpha = 0.6$ )		Motion ( $\alpha = 0.8$ )		Motion ( $\alpha = 1.0$ )	
Model	SSIM	cPSNR (dB)	SSIM	cPSNR (dB)	SSIM	cPSNR (dB)	SSIM	cPSNR (dB)	SSIM	cPSNR (dB)
Supervised	0.811 (0.052)	27.4 (2.6)	0.681 (0.108)	21.6 (3.2)	0.594 (0.161)	18.8 (3.7)	0.537 (0.172)	16.7 (3.9)	0.522 (0.160)	15.2 (3.8)
MRAugment	<b>0.852 (0.037)</b>	29.2 (1.4)	0.731 (0.097)	21.9 (3.6)	0.641 (0.164)	18.7 (4.1)	0.578 (0.182)	16.3 (4.3)	0.564 (0.171)	14.7 (4.0)
SSDU	0.844 (0.042)	27.9 (1.3)	0.798 (0.081)	22.8 (3.4)	0.704 (0.178)	19.6 (4.3)	0.622 (0.208)	17.0 (4.6)	0.592 (0.197)	15.2 (4.4)
Noise2Recon ( $\mathcal{R}(\sigma) = [0.5, 0.7]$ )	0.842 (0.044)	<b>29.7 (1.5)</b>	0.736 (0.102)	22.2 (3.7)	0.642 (0.168)	18.9 (4.2)	0.577 (0.181)	16.4 (4.4)	0.558 (0.168)	14.7 (4.1)
Aug (Motion, $\mathcal{R}(\alpha) = [0.2, 0.5]$ )	0.816 (0.048)	28.1 (1.5)	0.782 (0.084)	24.2 (2.8)	0.687 (0.189)	21.0 (4.2)	0.607 (0.224)	18.2 (5.0)	0.589 (0.207)	16.1 (5.2)
Aug (Motion, $\mathcal{R}(\alpha) = [0.5, 0.7]$ )	0.809 (0.048)	27.8 (1.4)	0.767 (0.079)	24.2 (2.5)	0.678 (0.175)	22.2 (3.8)	0.605 (0.203)	19.7 (4.5)	0.582 (0.193)	17.8 (4.5)
VORTEX (Motion, $\mathcal{R}(\alpha) = [0.2, 0.5]$ )	0.840 (0.045)	29.4 (1.5)	<b>0.823 (0.050)</b>	<b>25.9 (2.2)</b>	<b>0.749 (0.148)</b>	23.2 (3.7)	0.688 (0.181)	20.1 (5.2)	0.676 (0.170)	17.8 (5.8)
VORTEX (Motion, $\mathcal{R}(\alpha) = [0.5, 0.7]$ )	0.833 (0.045)	29.2 (1.5)	0.781 (0.059)	25.8 (1.9)	0.741 (0.084)	<b>24.3 (2.7)</b>	<b>0.702 (0.100)</b>	<b>22.8 (3.3)</b>	<b>0.683 (0.097)</b>	<b>21.7 (3.2)</b>
VORTEX (Image+Motion, ( $\alpha) = [0.5, 0.7]$ )	0.840 (0.043)	29.0 (1.5)	0.783 (0.064)	24.1 (2.7)	0.711 (0.133)	21.9 (3.3)	0.661 (0.146)	19.6 (4.1)	0.652 (0.137)	17.7 (4.4)
VORTEX (Image+Physics, ( $\alpha) = [0.5, 0.7]$ )	0.834 (0.043)	29.3 (1.6)	0.740 (0.097)	22.2 (3.7)	0.645 (0.168)	18.9 (4.2)	0.581 (0.180)	16.4 (4.4)	0.565 (0.168)	14.7 (4.1)

can *consistently* reconstruct higher quality images compared to state-of-the-art supervised, augmentation-based, and self-supervised methods.

**Scan reformatting.** Reconstructed scans are often reformatted to enable anatomical inspection from multiple views. In Fig. 7, we show an example mridata 3D FSE scan, which has been reformatted to the sagittal plane. Artifacts in reconstructions from baseline compressed sensing and deep learning-based methods are acutely visible in the reformatted slice. Motion ghosting artifacts seen in the axial plane (Fig. 3) appeared as coherent streaks in the sagittal reformat. Noise artifacts were amplified by DL-based baselines and lead to blurring among compressed sensing reconstructions. In contrast, VORTEX-based reconstructions sufficiently suppressed these artifacts, resulting in high-quality reformatted images.




**Figure 7:** Sample scan reconstruction (mridata) with different extents of motion ( $\alpha$ ) and noise ( $\sigma$ ) perturbations. Scans were reconstructed along axial slices, but are reformatted along the sagittal direction to illustrate through-plane artifacts. Unresolved motion artifacts can result in coherent ghosting and streaking artifacts along the through-plane direction. Noise artifacts are also more prominent in fat-suppressed regions (e.g. bone) and near articular (femoral, tibial, patellar) cartilage. VORTEX suppresses both noise and motion artifacts, producing higher quality images even along reformatted directions.

## F.2 fastMRI Results

We compare VORTEX to Supervised, SSDU, Aug (Motion), and MRAugment baselines for in distribution and OOD motion settings of different motion levels on the fastMRI multi-coil brain dataset in Table 11 [65]. Data preparation and experimental details follow the description in Appendix D, and all experiments are conducted at 8x acceleration.

We demonstrate that VORTEX has comparable performance to baselines for in distribution, and outperforms SSDU by +0.025 SSIM and +3.1dB cPSNR, and MRAugment by +0.092 SSIM and 4.0dB cPSNR on motion level  $\alpha = 0.4$ ; SSDU by +0.045 SSIM and +4.7dB cPSNR, and MRAugment by +0.108 SSIM and +5.6dB cPSNR on motion level  $\alpha = 0.6$ ; SSDU by +0.08 SSIM and +5.8dB cPSNR, and MRAugment by +0.124 SSIM and +6.5dB cPSNR on motion level  $\alpha = 0.8$ ; SSDU by +0.091 SSIM and +6.5dB cPSNR, and MRAugment by +0.119 SSIM and +7.0dB cPSNR on motion level  $\alpha = 1.0$ . This demonstrates that the effectiveness of VORTEX for both in distribution and OOD data



**Figure 8:** Sample fastMRI brain reconstructions with heavy motion ( $\alpha=0.4$ ) perturbation. All baseline methods suffer from coherent ghosting artifacts. SSDU can suppress the coherence of these artifacts, but results in extensive blurring of vasculature. VORTEX can minimize this blurring while suppressing coherent ghosting artifacts.

generalizes to 2D MRI sequences which implies broader clinical utility. Sample reconstructions are shown in Fig. 8.

While augmentations are routinely used in supervised training (e.g. MRAugment), we find that supervised augmentation baselines that leverage physics-driven augmentations (e.g. motion) resulted in considerable degradation in both SSIM and cPSNR on in-distribution data, while consistently underperforming VORTEX on motion-corrupted data ( $\alpha \geq 0.4$ ). This may indicate that using physics-driven MRI augmentations in supervised training may not be well-posed for mitigating label paucity. Like existing supervised and self-supervised methods, augmentation-based methods also suffer from a performance tradeoff between in-distribution and out-of-distribution settings. Thus, designing training protocols with certain augmentations may require careful tuning to find a balance between in-distribution and OOD performance.

In contrast, VORTEX performs well among *both* in-distribution and OOD data, mitigating the performance tradeoff between these two settings. This generalizability to both settings can be attributed to VORTEX’s use of consistency, where *weak references* (i.e. pseudo-labels) are used to supervise a unified objective for both reconstruction and artifact correction. This may also help eliminate careful hand-tuning of augmentation parameters that is often needed to balance this tradeoff.