# Radon Concentration Measurement with a High-Sensitivity Radon Detector at the Yemilab

Kyung Min Seo,<sup>*a*</sup> Hyunsoo Kim,<sup>*a*</sup> Yeong Duk Kim,<sup>*b*,*c*</sup> Hye Young Lee,<sup>*b*</sup> Jaison Lee,<sup>*b*</sup> Moo Hyun Lee,<sup>*b*,*c*,1</sup> Jungho So,<sup>*b*</sup> Sang Cheol Yoon,<sup>*b*</sup> and Young Soo Yoon<sup>*d*</sup>

<sup>a</sup>Department of Physics and Astronomy, Sejong University, Seoul, Republic of Korea

<sup>b</sup>Center for Underground Physics, Institute for Basic Science (IBS), Daejeon, Republic of Korea

<sup>c</sup>IBS School, University of Science and Technology (UST), Daejeon, Republic of Korea

<sup>d</sup>Korea Research Institute of Standards and Science (KRISS), Daejeon, Republic of Korea

*E-mail:* mhlee@ibs.re.kr

ABSTRACT: The radiation emitted from radon is a critical background in rare event search experiments conducted at the Yemi Underground Laboratory (Yemilab) in Jeongseon, Korea. A Radon Reduction System (RRS) has been developed and installed in Yemilab to reduce radon concentration in the air. The RRS primarily provides a purified air of 50 m<sup>3</sup>/h to the cleanroom used to assemble crystal detectors in the AMoRE, a neutrinoless double beta decay search experiment. RRS can reduce the radon level by a factor of 300, so a high-sensitivity radon detector was required. A highly sensitive radon detector was constructed using a 70 L chamber with a large PIN photodiode to measure radon concentration in the purified air. The radon detector shows an excellent resolution of 72 keV (FWHM) for 6.003 MeV alphas from <sup>218</sup>Po decay and a sensitivity down to 23.8 ± 2.1 mBq/m<sup>3</sup> with a boil-off N<sub>2</sub> gas sample. The radon concentration level from the RRS measured by the radon detector was below 0.29 Bq/m<sup>3</sup> with a reduction factor of about 300.

KEYWORDS: Radiation monitoring, Gaseous detectors, Heavy-ion detectors; Analogue electronic circuits

<sup>&</sup>lt;sup>1</sup>Corresponding author.

# Contents

1	Introduction	1
2	Radon Reduction System	4
3	Radon Detector	4
4	Radon Concentration Measurements	11
5	Summary	12

# 1 Introduction

The underground laboratory, developed to reduce the background caused by cosmic rays, must consider the other background created by the Earth's surface, radon. Radon is derived from uranium and thorium, commonly present in the Earth's crust, and can be represented by the radium series from <sup>238</sup>U and <sup>232</sup>Th decay chains, as shown in Fig. 1. Radon is a radioactive noble gas with very low chemical reactivity in general conditions. Because of these characteristics, the concentration of radon cannot be measured directly. However, the radon concentration can be inferred through the daughter nuclide, poloniums. Subsequent stages of radon decay ultimately result in lead production, accompanied by the emission of significant radiation. Radon contamination in the experimental environment presents a substantial background source in underground experiments. Therefore, it is crucial to measure the radon concentration.

Radon emanates from rocks, and as underground experimental facilities are encased within rock formations, there is a potential for radon contamination within the underground experimental facility. In such facilities, radon is removed by radon reduction systems (RRS), which reduce the radon concentration in the air by more than a hundred folds. Commercially available radon detectors cannot measure this low radon concentration. Therefore, a highly sensitive PIN photodiode-based radon detector has been designed and built for use at Yemilab.

Yemilab [1-3] is the new underground laboratory in Korea that has been operating since the end of 2022, which has 16 research spaces with a total area of 3000 m<sup>2</sup> and a 1000-meter overburden. The schematic of Yemilab is shown in Fig. 2. Yemilab has a cosmic muon background that is four times lower than the previous underground laboratory, Yangyang Underground Laboratory (Y2L) in Yangyang, Korea [2].

The AMoRE [4, 5] and COSINE [6] experiments, two major physics experiments searching for neutrinoless double  $\beta$  decay and weakly interacting massive particles, are being constructed. Both experiments require extremely low background levels to observe very rare processes. It requires efforts to reduce the detector's intrinsic background and understand and reduce the background from the environment [7].



**Figure 1**. A schematic of Radium series from the  $^{238}$ U(top) and  $^{232}$ Th(bottom) decay chain. For  $\beta$  decay, only the half-life information was included.

Noble gases like radon bind weakly to materials. To effectively adsorb radon in the air, a material with a large surface area is advantageous, and charcoal plays this role. A radon reduction system (RRS) is a device designed to reduce radon in the air by drawing radon-rich air and adsorbing the radon into the charcoal tank. As the charcoal tank's temperature decreases, the charcoal's capacity to adsorb the radon increases. Several RRS were developed in several underground laboratories and are still in operation today [8–10].

The radon detector is designed to measure the concentration of radon that decays to airborne polonium in trapped air in a sampling chamber. Polonium becomes ionized and acquires a positive charge due to the recoil energy of its previous decay [11]. Ionized polonium drifts into the



Figure 2. A schematic of the Yemilab experimental areas [1].

photodiode following the electric field created by the voltage applied. To achieve this, a sufficient volume is needed to contain sampled air and form an electric field to collect polonium in that air. From this perspective, a radon detector with a large air sample holding chamber was constructed, similar to radon detectors currently being used in other underground laboratories [12–23].

This paper discusses the radon concentration measurements in the air from the RRS with a highly sensitive radon detector at the Yemilab.



**Figure 3**. The layout of the custom-made Radon Reduction System. 1: Compressor, 2: Receiver tank, 3: Absorption dryer, 4: Air chiller, 5: Heater and HEPA filter units, 6: Active charcoal tanks, and 7: Control panel.

# 2 Radon Reduction System

ATEKO [24], the most popular RRS manufacturer in the world, has provided the system at many underground facilities [8], including the one at Y2L. However, replacing the most high-pressure part of the system requires an additional certification by the Korea Occupational Safety and Health Agency (KOSHA). To avoid this vital regulation issue in Korea, we purchased an RRS manufactured by a Korean company specializing in ultra-pure water filtration systems.

Figure 3 shows the layout of the RRS for the AMoRE-II at Yemilab. Air from the environment is supplied by the compressor and pressurized in the receiver tank. The absorption dryer can remove the moisture in the air under the  $-70^{\circ}$ C of dew point. This moisture-free pressurized air is contained in active charcoal tanks up to 10 bars. The temperature of active charcoal in the tanks has been cooled below  $-70^{\circ}$ C by the air chiller to remove radon from the dried air effectively. A heater and HEPA filter units are installed at the last part of the system to warm up and for dust filtration. The control panel operates all these processes automatically, controlled by a Windows-based program for online operation and monitoring.

The RRS aims to reduce the radon concentration below 1/1000 that of the air in the Yemilab and supply the radon-free air with a capacity up to 60 m<sup>3</sup>/h. A pre-tuning of the system was performed in the factory to optimize the operation conditions for a period longer than six months. It was decided to supply the air at 50 m<sup>3</sup>/h to minimize the maintenance effort.

#### 3 Radon Detector

Figure 4 shows the radon detector developed for Yemilab. The detector consists of a 70 L air sampling chamber with a PIN photodiode connected to a ceramic feedthrough mounted on the top flange. The chamber is made of 3 mm thick stainless steel, and the inner surface has been electrically polished to reduce the radon emanating from the chamber wall. The chamber has an inlet and an outlet to measure the air sample, and an additional vacuum pump connection is designed to evacuate the air inside the chamber quickly. The corners of the chamber are curved to form a



**Figure 4.** The picture (left) and schematic (right) of the radon detector. The measurement volume is the rounded part of the chamber below the flange plate. The measurement volume comprises 3 mm thick stainless steel with an electropolished inner wall. The measurement volume is approximately 70 L. The detector has an air inlet, an air outlet, a port for the vacuum gauge, and a port connected to a vacuum pump to evacuate the internal air. There is a hole in the center of the flange for the feedthrough installation. Above the measurement volume is the front-end electronics board with voltage divider and preamplifier circuits. It is covered with a chamber lid. The chamber lid has three connector sockets, each with SHV, BNC, and power wires for the preamplifier. The outer wall of the chamber is connected to the ground.

uniform electric field, a different design from the radon chamber used in the radon monitoring for the KIMS experiment at Y2L [25, 26].

A windowless silicon PIN photodiode (S3204-09; Hamamatsu Photonics K.K. [27]) with dimensions of 18 mm  $\times$  18 mm is used. A negative voltage is applied to the photodiode to establish an electric potential difference between the photodiode and the grounded inner surface of the chamber. Ionized polonium, a positively charged daughter particle from the radon decay, travels to the photodiode in the chamber. Upon reaching the surface of the photodiode, polonium eventually decays and emits  $\alpha$  particles. The emitted  $\alpha$  particles lose energies in the photodiode's depletion region to create electron-hole pairs through which a signal is generated.

The air is introduced into the chamber via an oil-free vacuum pump (Rocker 400; Rocker Scientific Co. Ltd. [28]), with a flowing rate of ~ 20 L/min while keeping one atmospheric pressure. The dust in the sampled air is removed through a dust filter, which performs the same function as in the commercial radon detector (RAD7; Durridge Company, Inc. [29]). After filtration, the air undergoes dehumidification facilitated by a drying unit, a cylindrical acrylic container with approximately 500 g of desiccant (Drierite desiccant 23005 [30]). The air passing through the desiccant has a relative humidity of 2 to 3%RH. These components are seamlessly connected using silicon tubes, as shown in Fig. 5.

The signal generated by the photodiode is sent to the front-end electronics board via a feedthrough. Figure 6 shows the circuit diagrams of the front-end electronics. The front-end



**Figure 5.** A schematic for the radon measurement setup is shown on the left. The solid (dashed) line is the flow line for the air (electrical connection). Other lines are explained in the legends on the top left. The pictures on the right show each part of the measurement setup, while (h) is the radon detector in Fig. 4.

electronics board has a preamplifier (H4083; Hamamatsu Photonics K.K. [31]) and a high-voltage (HV) divider to supply HV between the photodiode and the ground. There are BNC and SHV sockets for signal and HV, respectively. The front-end electronics board is covered with a Faraday cage to shield the external electromagnetic fields. On the left, a socket connects to the SHV cable, and on the right, there is a socket through which signals are transmitted through the BNC cable. The two holes in the center connect via a feedthrough to the photodiodes in the chamber. On the left side of the BNC connector, a board is designed to mount a preamplifier. This preamplifier is Hamamatsu's H4083 and is installed under a Faraday cage after soldering to the front-end electronics board to shield the external electromagnetic noise. The power to the preamplifier is supplied by the 3-pin connector at the bottom. A HV supply (N470; CAEN Technologies, Inc.) is used to apply the HV for the radon detector and photodiode, and a low voltage (LV) supply (U8032A; Keysight Technologies, Inc. [32]) to the preamplifier.

Front-end electronics signals are shaped and amplified by a pulse-shaping amplifier (570; ORTEC Inc. [33]). The shaping time used is 2  $\mu$ s. The shaped signal is finally digitized by the 12-bit flash ADC with a sampling rate of 25 MS/s (NKFADC25; Notice Korea Co.). The pulse height trigger determines whether to save the waveform. Figure 7 shows a typical waveform of an  $\alpha$  signal. All of the analysis parameters are reconstructed from the waveform. Table 1 defines six parameters used in the analysis.

To correct the baseline of the waveform, the mean of the pedestal is calculated using the data in the 6  $\mu$ s window before the pulse, and each ADC value of the waveform is subtracted by the mean of the pedestal, as shown in Fig. 7. The parameters corresponding to the energy of an event are *H* and *A*. The *H* is the maximum height of the waveform, and *A* is the total ADC count



**Figure 6**. The schematic illustrates the front-end electronics, with the top circuit diagram depicting the HV divider section. This circuit is responsible for grounding the chamber and applying HV and bias to the photodiode. The schematics below detail the preamplifier circuit and its power supply. The photodiode signal goes into the preamplifier's first terminal for amplification, and the output is from terminal 9. This amplified signal is transmitted to the shaping amplifier through the BNC cable.



Figure 7. An example waveform of an  $\alpha$  event after the pedestal subtraction.

Table 1. The definition of parameters reconstructed from waveform.

Parameter	Description		
$P$ and $P_{RMS}$	pedestal mean and RMS		
$H$ and $H_x$	maximum pulse height and its position.		
A	integrated area $(H_X \pm 2 \ \mu s)$		
$T_{\rm trg}$	trigger time		

**Table 2.** The mean value  $(\bar{x})$  and  $\sigma$  after Crystall Ball function fitting in Fig. 8. The sigma resolution is calculated by  $\sigma$  and  $\bar{x}$ , and the FWHM is calculated by the histogram. E<sub> $\alpha$ </sub>, alpha energy, values are from NNDC [34].

Isotopes	$E_{\alpha}$ [MeV]	$\bar{x}$ [ADC]	$\sigma$ [ADC]	$\sigma/ar{x}$ [%]	FWHM [keV]
<sup>210</sup> Po	5.304	587.506(33)	4.214(37)	0.72	81
<sup>218</sup> Po	6.003	666.205(7)	3.275(5)	0.49	72
<sup>214</sup> Po	7.687	857.450(5)	3.376(4)	0.39	72

calculated by integrating the waveform in the range of  $H_x \pm 2 \mu s$ . All of these have characteristics proportional to the signal's size (energy). The parameter *H* is used as the corresponding energy value. The parameter *A* is used primarily for noise rejection. The *P* and *P*<sub>RMS</sub> are calculated at the first 6  $\mu s$  time window of the waveform. In the case of short-term electrical noise or instability, these parameters tend to deviate from the nominal case. Some electrical noise signals are filtered through these parameters.

As shown in Fig. 8, four peaks are observed. The first three peaks (<sup>212</sup>Po, <sup>218</sup>Po, and <sup>214</sup>Po) belong to the <sup>238</sup>U chain and the last peak is <sup>212</sup>Po from the <sup>232</sup>Th chain. Each peak, except for <sup>212</sup>Po, is fitted with the Crystal Ball function [Eq. 3.1], and the fitting results are summarized in Table 2.

$$f(x;\alpha,n,\bar{x},\sigma) = \begin{cases} \exp\left(-\frac{(x-\bar{x})^2}{2\sigma^2}\right), & \text{for } \frac{x-\bar{x}}{\sigma} > -\alpha\\ A \cdot \left(B - \frac{(x-\bar{x})}{\sigma}\right)^{-n}, & \text{for } \frac{x-\bar{x}}{\sigma} \le -\alpha \end{cases}$$
(3.1)

The resolutions of all peaks ( $\sigma/\bar{x}$ ) are smaller than 1%. <sup>218</sup>Po and <sup>214</sup>Po, which are from the same decay chain of <sup>222</sup>Rn, are used to estimate the concentration of radon. Due to its long half-life, 138 days, <sup>210</sup>Po is unsuitable for estimating the concentration. <sup>210</sup>Pb, which follows <sup>214</sup>Po, has a half-life of about 22.3 years. Because of this characteristic, <sup>210</sup>Pb accumulates in the chamber and generates a signal. The rate of <sup>210</sup>Po decay from the <sup>210</sup>Pb is proportional to the age of the chamber. To eliminate this, the inside of the chamber must be electropolished, and the silicon diode must be replaced. The <sup>212</sup>Po of the <sup>232</sup>Th decay chain has a different shape than the other polonium peaks. Due to the short half-life of 300 ns, <sup>212</sup>Po decays before reaching the photodiode and loses its recoil energy. Therefore, the locations of the three polonium peaks in the <sup>238</sup>U chain are used for the energy calibration through a linear function, as shown in Fig. 8.



**Figure 8**. The ADC distribution with their fitting results. The ADC value is the difference between height and pedestal (H - P) for each event. Three peaks are shown in the figure in the order of <sup>210</sup>Po, <sup>218</sup>Po, and <sup>214</sup>Po. Peaks in the <sup>238</sup>U chain are fitted with a Crystal Ball function, and the results are listed in Table 2. The upper part of the inset shows the logarithmic scale of the same distribution, indicating a relatively small <sup>212</sup>Po peak. The <sup>212</sup>Po has a short half-life of 299 ns and most decay before reaching the diode, making the distribution different from other peaks. The energy calibration plot with a linear function fit of three peaks is below.

To calibrate the HV and radon concentration of the radon detector, a closed loop system consisting of a radon detector, RAD7, and radium source is prepared as shown in Fig. 5. The efficiency of the radon detector depends on the HV applied. As the HV increases, the electric field for capturing polonium in the chamber intensifies. The efficiency is maximum when the voltage is high enough to capture all the polonium. RAD7 has a hemispherical detection volume of 0.7 L coated inside with an electrical conductor, and a silicon photodiode is installed at the center of the hemisphere. The radium source, purchased from the Korea Research Institute of Standards and Science (KRISS), has an activity of approximately 1.5 kBq and is sealed within a stainless steel container, as shown in Fig. 5(f). A filter is installed at the container outlet to prevent radium from being blown out directly. Measurements are taken at six points by increasing 250 V each time, from 1000 to 2250 V, and each measurement takes an hour. The activity ratios (*R*) for HV are calculated, and its definition is

$$R = A_c / A_R, (3.2)$$



Figure 9. The activity ratios obtained from each data set from 1000 to 2250 V with a step of 250 V. Each point is measured for about an hour. The location of the red diamond marker is where R begins to saturate. This point is set as the operating voltage.

where  $A_c$  is the number of <sup>218</sup>Po within  $3\sigma$  range of fitting result on an hour measurement,  $A_R$  is the activity measured by the RAD7. The *R* versus HV is shown in Fig. 9. The operating voltage is determined at the point where *R* starts to saturate, 1750 V.

Radon-rich air is injected into the closed-loop configuration to calibrate the detector's radon concentration. In the closed loop of the radon detector and RAD7, as shown in Fig. 5, the sampled air is circulated by the RAD7's pump (flowing rate of  $\sim 0.8$  L/min) with the source for approximately three hours to saturate it sufficiently. After confirming that the radon concentration is saturated, the source is isolated from the loop. Radon remaining in the closed-loop structure naturally decays. The radon detector and RAD7 measure the radon concentration in the closed loop simultaneously and continuously, which shows a reduction in radon levels. Figure 10 shows the correlation of radon concentration between RAD7 and the radon detector up to  $\sim 300$  Bq/m<sup>3</sup> with about two weeks of data. Each point corresponds to data collected for 2 hours. The correlation is fitted by a linear function without an offset, and the calibration function is obtained.

To understand the intrinsic background of the radon detector, about 15 L/min of pure nitrogen gas from boil-off in a liquid nitrogen gas bottle is flushed to the chamber for 2 hours. The amount of nitrogen gas is enough to refresh the chamber volume approximately 25 times. The humidity within the chamber is about 40 ppm when the chamber is sealed for measurement. Data are collected



**Figure 10**. The correlation between RAD7 and radon detector after saturation of radon concentration. Over two weeks, the naturally declining radon concentration is measured using two detectors (RAD7 and the radon detector). A linear calibration function (red) without an offset fits the data.

for approximately 12 days, and the intrinsic background of the radon detector is measured to be  $23.8 \pm 2.1 \text{ mBq/m}^3$ , as shown in Fig. 11.

#### 4 Radon Concentration Measurements

Environmental monitoring at Yemilab is in progress, including the concentration of radon. Radonreduced air by the RRS is piped to the AMoRE-II detector room and measured by the radon detector. Ambient air, before RRS, is measured by the RS9A, which is a compact pulsed-ionization type radon detector developed by RadonFTLab in South Korea [35, 36], and can measure the radon concentration below 3.7 kBq/m<sup>3</sup> with an accuracy of less than 15%. Figure 12 compares radon concentrations between the AMoRE experimental hall and the radon-reduced air from the RRS. The last nine days of data are used to estimate the reduction factor when the RRS is settled down. The average radon concentration levels before and after RRS are  $17.6 \pm 0.3$  and  $0.058 \pm 0.004$  Bq/m<sup>3</sup>, respectively, and the reduction factor is about 1/300. The requirement of radon concentration for AMoRE-II is 0.29 Bq/m<sup>3</sup> [37], aligning with the region of interest for the neutrinoless double  $\beta$ decay Q-value of <sup>100</sup>Mo. Figure 12 shows that the radon concentration satisfies the requirement after ten days.



Figure 11. The intrinsic background measurement result of the radon detector. The radon detector was filled with pure boil-off nitrogen gas, and the average data value measured over 12 days is displayed as the red line,  $23.8 \text{mBq/m}^3$ 

The calibration process of the radon detector performed in this paper is a relative calibration using a RAD7 in a closed loop. Even though there is a desiccant in the loop, the RAD7 shows the humidity level to be 2 to 3%RH, which is about  $0.58 \text{ g/m}^3$ . Since the humidity of the radon-reduced air from RRS is ~  $0.21 \text{ g/m}^3$ , humidity is slightly different between the two setups. However, it seems negligible in calibration constant values between the two when looking at the humidity dependence of calibration constant in Fig. 5 of [14]. The RRS output measurement with lower humidity conditions could be overestimated due to a possible under-correction.

More rigorous calibration works with different humidity environments and flow rates will be carried out in the near future.

# 5 Summary

The Yemilab, an underground laboratory in Jeongseon, Korea, was built for rare event search experiments to reduce cosmic-ray backgrounds. However, the Yemilab is still impacted by the radon background in the air.

An RRS using a cold and pressurized charcoal chamber was designed and installed at the Yemilab to reduce the background radon in the AMoRE experiment, which is looking for the



**Figure 12**. A comparison plot of radon concentration between the experiment hall air and the radon-reduced air from the RRS. Each data is measured by the radon detector (RS9A) for AMoRE's detector room air from the RRS (AMoRE experimental hall air to the RRS). It is assumed that the uncertainty of the RS9A data set is 15% in precision. The blue dashed line is the AMoRE-II radon concentration requirement.

neutrinoless double beta decay from <sup>100</sup>Mo isotopes in lithium molybdate crystal detectors at cryogenic temperatures. It reduces the radon level in the air at the experimental area to the level where the radon background becomes negligible to the experiment's sensitivity with a capacity of  $50 \text{ m}^3/\text{h}$ .

A highly sensitive radon detector with a 70 L chamber volume and a large silicon PIN photodiode sensor readout was designed and manufactured to quantify the amount of radon reduced by the RRS. A calibration function for different HV values up to 2250 V was obtained using a closed loop connected with a calibrated commercial radon detector RAD7 using a radium source. Also, a radon activity calibration using the same setup was made up to  $\sim 300$  Bq/m<sup>3</sup> at 1750 V. From the calibration, the radon detector can measure alpha particles from <sup>218</sup>Po (6.003 MeV) and <sup>214</sup>Po (7.687 MeV), daughter isotopes of <sup>222</sup>Rn, with an FWHM resolution of 72 keV. A boil-off N<sub>2</sub> gas test showed 23.8 ± 2.1 mBq/m<sup>3</sup>, the lowest background level indicating an intrinsic background.

To confirm the reduction in radon by RRS, air from the RRS was measured using the radon detector, while the air going into the RRS was measured by another commercial radon detector, RS9A, for 19 days. From the measurement, a radon reduction factor of  $\sim 300$  by the RRS was measured. Also, the RRS output level reached below 0.29 Bq/m<sup>3</sup>, which is the radon concentration

#### requirement of AMoRE-II.

#### Acknowledgments

This work was supported by the Institute for Basic Science (IBS) funded by the Ministry of Science and ICT, Korea (Grant: IBS-R016-D1).

# References

- [1] K. S. Park et al., *The new underground facility in Korea, Yemilab, Journal of Physics: Conference Series* **2156** 012171 (2021).
- [2] K. S. Park et al., Construction of Yemilab, Front. Phys. 12 1323991 (2024).
- [3] M. H. Lee Radioassay and Purification for Experiments at Y2L and Yemilab in Korea, Journal of Physics: Conference Series 1468 012249 (2020).
- [4] V. Alenkov et al. (AMoRE collaboration), *Technical Design Report for the AMoRE* 0νββ Decay Search Experiment, arxiv:1512.05957 (2015).
- [5] M. H. Lee (AMoRE collaboration), AMoRE: a search for neutrinoless double-beta decay of <sup>100</sup>Mo using low-temperature molybdenum-containing crystal detectors, JINST 15(08) C08010 (2020).
- [6] G. Adhikari et al. (COSINE collaboration), *Initial performance of the COSINE-100 experiment, The European Physical Journal C* **78** 107 (2018).
- [7] C. Ha et al., Radon concentration variations at the Yangyang underground laboratory, Front. Phys. 10 1030024 (2022).
- [8] J. Pérez-Pérez et al., *Radon Mitigation Applications at the Laboratorio Subterráneo de Canfranc* (*LSC*), *Universe* **8(2)** 112 (2022).
- [9] R. Hodák et al., *Characterization and long-term performance of the Radon Trapping Facility operating at the Modane Underground Laboratory, Journal of Physics G: Nuclear and Particle Physics* **46(11)** 115105 (2019).
- [10] M. Murra et al., *Design, construction and commissioning of a high-flow radon removal system for XENONnT, The European Physical Journal C* 82(12) 1104 (2022).
- [11] E. Albrecht and A. Kaul, Continuous registration of <sup>222</sup>Rn concentration in air varying with time (SM-95/39), Assessment of Airborne Radioactivity, Proceeding of IAEA Symposium (Vienna, 3-7 July, 1967) 643 (1967).
- Y. Takeuchi et al., Development of high sensitivity radon detectors, Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment 421 334 (1999).
- [13] J.-X. Wang et al., An electrostatic radon detector designed for water radioactivity measurements, Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment 421 601 (1999).
- [14] K. Hosokawa et al., Development of a high-sensitivity 80 L radon detector for purified gases, Progress of Theoretical and Experimental Physics **2015** 033H01 (2015).
- [15] C. Mitsuda et al., Development of super-high sensitivity radon detector for the Super-Kamiokande detector, Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment 497 414 (2003).

- [16] S. Folea et al., WI-FI portable solution for distributed radon measurements, Romanian Journal of Physics 58 S126 (2013).
- [17] J.L. Gutiérrez et al., *Radon emanation measurements using silicon photodiode detectors*, *Applied Radiation and Isotopes* **60** 583 (2004).
- [18] K. Okamoto et al., *Improvement of radon detector performance by using a large-sized PIN-photodiode*, arxiv:2112.06614 (2021).
- [19] S. Elisio and L. Peralta, Development of a low-cost monitor for radon detection in air, Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment 969 164033 (2020).
- [20] E.A. Yakushev and A.V. Lubashevskii, *Radon-induced background and methods of its elimination in the EDELWEISS-II experiment, Russian Physics Journal* **53** 616 (2010).
- [21] E. Choi et al., Highly sensitive radon monitor and radon emanation rates for detector components, Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment 459 177 (2001).
- [22] J. Kiko, Detector for <sup>222</sup>Rn measurements in air at the 1 mBq/m<sup>3</sup> level, Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment 460 272 (2001).
- [23] F. Mamedov et al., *Development of an ultra-sensitive radon detector for the SuperNEMO experiment*, *JINST* 6(12) C12008 (2011).
- [24] ATEKO a.s., https://www.ateko.cz/en, (Accessed February 15, 2024).
- [25] M. Lee et al., Radon Environment in the Korea Invisible Mass Search Experiment and Its Measurement, Journal of the Korean Physical Society 58 713 (2011).
- [26] K. M. Seo et al., Silicon PIN photodiode-based radon detectors for an underground experiment environment, Proceedings of The 39th International Conference on High Energy Physics — PoS(ICHEP2018) 340 803 (2019).
- [27] Hamamatsu silicon photodiode S3204-09, https://www.hamamatsu.com/us/en/product/opticalsensors/photodiodes/si-photodiodes/S3204-09.html, (Accessed February 16, 2024).
- [28] Rocker 400 vacuum pump, https://www.rocker.com.tw/en/product/lab-pumps/oil-free-vacuum-pump/rocker-400-oil-free-vacuum-pump/, (Accessed February 16, 2024).
- [29] RAD7 RADON DETECTOR, https://durridge.com/products/rad7-radon-detector/, (Accessed February 16, 2024).
- [30] DRIERITE 23005, https://secure.drierite.com/catalog3/product.cfm?item\_num=23005/, (Accessed February 16, 2024).
- [31] Hamamatsu charge amplifier H4083, https://www.hamamatsu.com/content/dam/hamamatsuphotonics/sites/documents/99\_SALES\_LIBRARY/ssd/charge\_amp\_kacc9001e.pdf, (Accessed February 16, 2024).
- [32] Keysight Triple Output DC Power Supply U8032A, https://www.keysight.com/us/en/product/U8032A/triple-output-dc-power-supply-60v-3a-2x-5v-3a-375w.html, (Accessed February 16, 2024).
- [33] Ortec amplifier 570,

https://www.ortec-online.com/-/media/ametekortec/brochures/5/570.pdf?la=en&revision=53f9a77cd8f8-4459-80b8-a93822b2aab2, (Accessed February 16, 2024).

- [34] National Nuclear Data Center, https://www.nndc.bnl.gov/, (Accessed February 15, 2024).
- [35] IoT radon sensor: RS9A, http://radonftlab.com/kr/iot-radon-sensor-rs9a/, (Accessed February 15, 2024).
- [36] I. Dimitrova et al., *Study of the performance and time response of the RadonEye Plus2 continuous radon monitor*, *Measurement* **207** 112409 (2023).
- [37] A. Agrawal et al. (AMoRE collaboration), *Radioassay of the materials for AMoRE-II experiment*, *Submitted to Front. Phys.* (2024).