

Measurements of Protein-Protein Interactions by Size Exclusion Chromatography

J. Blustine, V. Berejnov, S. Fraden

Complex Fluids Group,

The Martin J. Fisher School of Physics,

Brandeis University, Waltham, MA 02454

April 14, 2024

Abstract

A method is presented for determining second virial coefficients (B_2) of protein solutions from retention time measurements in size exclusion chromatography (SEC). We determine B_2 by analyzing the concentration dependence of the chromatographic partition coefficient. We show the ability of this method to track the evolution of B_2 from positive to negative values in lysozyme and bovine serum albumin solutions. Our SEC results agree quantitatively with data obtained by light scattering.

Introduction

It is well known in size exclusion liquid chromatography (SEC) that the solute retention time depends sensitively on the solute's size, although no universal calibration for SEC has yet been achieved. It has also been realized that thermodynamic

Present Address: Physics Department, Cornell University, Ithaca, NY 14853

non-ideality leads to concentration dependent retention times (Nichol et al., 1978). Such dependence can be utilized to quantify the second osmotic virial coefficient, B_2 .

For a non-ideal solution the osmotic pressure can be written as a power series expansion in the solute number density (Hill, 1960).

$$\frac{\pi}{k_B T} = c + B_2(T) c^2 + \dots \quad (1)$$

In Eq.1 T is the absolute temperature and k_B is Boltzmann's constant. All terms higher than first order in density represent non-ideality.

The second virial coefficients of protein solutions have generated a great deal of interest since (George and Wilson, 1994) showed a correlation between protein crystallisability and B_2 . Their work demonstrated that many proteins crystallize in conditions where the second osmotic virial coefficient becomes slightly negative, indicating net attractive interactions between protein molecules. The most prevalent experimental procedure for measuring B_2 is light scattering. Additionally, sedimentation equilibrium (Behlke and Ristau, 1999), osmometry (Moon et al., 2000), neutron (Velev et al., 1998) and x-ray scattering (Bonnet et al., 1999), and self-interaction chromatography (Tessier et al., 2002) have been employed to quantify protein solution non-ideality.

(Nichol et al., 1978) showed the possibility of measuring B_2 with frontal elution liquid chromatography. Although frontal chromatography (Nichol et al., 1978, Wills et al., 1980) allows one to fix the solute concentration in the column directly, it requires a large amount of protein (> 0.5 g) and long experiment times (about three hours per column run). In this study we extend their method to pulse size exclusion HPLC, where a small amount of protein is injected into and subsequently flows down the column. This adaptation drastically reduces the amount of protein (< 25 mg) and time needed (about 15 minutes per column run) to measure B_2 by SEC. We show that our results for B_2 obtained with size exclusion chromatography agree well with those from frontal chromatography and from light scattering measurements. We also demonstrate that SEC can track the evolution of B_2 from positive to negative values.

Theory

For the reader's convenience we reproduce the theory of (Nichol et al., 1978). We assume a balance of the solute, i.e. protein, chemical potentials (μ_p and μ_i) between the stationary and mobile phases as the solute is transported through the column. The pore volume (i.e. stationary phase) is labelled with the subscript p, and the inter-pore volume (i.e. mobile phase) with the subscript i. Equilibrium requires $\mu_p = \mu_i$. We write these chemical potentials by including the standard part μ° , the ideal term, and a term accounting for thermodynamic non-ideality through the activity coefficient:

$$\begin{aligned}\mu_p &= \mu_p^\circ + RT \ln(C_p \gamma_p(C_p)) \\ \mu_i &= \mu_i^\circ + RT \ln(C_i \gamma_i(C_i))\end{aligned}$$

where C_{ip} are the local solute weight concentrations, R is the universal gas constant and $\gamma_p(C_p)$; $\gamma_i(C_i)$ are the thermodynamic activity coefficients of solute molecules in the pore and inter-pore volumes respectively. Rearrangement of these equations yields:

$$\ln \frac{\gamma_i}{\gamma_p} = \ln \frac{C_p}{C_i} - \ln(K_0) \quad (2)$$

where K_0 is the partition coefficient of solute molecules between chromatographic phases in the limit of infinite dilution. The relation between weight concentration, C , and number density, ρ , is $\rho = C \frac{N_A}{M_w}$. N_A is Avogadro's number and M_w is the solute molecular mass. (Nichol et al., 1978) made a virial expansion of the activity coefficients

$$\ln(\gamma) = 2B_2(N_A/M_w)C + \text{higher terms} \quad (3)$$

We note that this consideration assumes no difference in the solute-solute interactions in the mobile and stationary phases. The local solute distribution coefficient is $K_D = \frac{C_p}{C_i}$. If K_D is independent of concentration, as is the case for pulse chromatography with $B_2 = 0$, or if the concentration is constant as in frontal elution

chromatography, then (Nichol et al., 1978, Yau et al., 1979)

$$K_D = \frac{C_p}{C_i} = \frac{t_r - t_0}{t_T - t_0} = \frac{V_r - V_0}{V_T - V_0} \quad (4)$$

where t_r and V_r are the solute retention time and volume, t_0 and V_0 are the retention time and volume of completely excluded molecules (i.e. the "dead" volume), and t_T and V_T the retention time and volume of completely included molecules (i.e. the "total" volume). Inserting the definition of K_D (Eq. 4) and Eq. 3 into Eq. 2 and keeping only the first order terms in concentration one obtains a relation between K_D , B_2 , and C_i valid for frontal chromatography where the concentration C_i is the plateau value of the solute concentration in the mobile phase:

$$\ln \frac{K_D}{K_0} = 2B_2 \frac{N_A}{M_w} C_i (1 - K_D) \quad (5)$$

To adapt this to pulse chromatography we replace the plateau value with the average concentration $\langle C_i \rangle$ of the mobile phase in the pulse:

$$\ln \frac{K_D}{K_0} = 2B_2 \frac{N_A}{M_w} \langle C_i \rangle (1 - K_D) \quad (6)$$

Since $\langle C_i \rangle$ is not directly accessible in a HPLC experiment one must relate it to measurable parameters. One determines the mass of solute molecules in the pulse, or migration zone, (m_{zone}) by integrating the concentration as a function of time over the zone volume, i.e. the peak (V_z). For our columns, in which there is no irreversible binding of protein molecules to the column, all the injected molecules are accounted for by integrating the peak. Therefore the total injected mass is the same as the total mass in the zone, $m_{inj} = C_{inj}V_{inj} = m_{zone}$, but the concentration of solute in the migration zone is much lower than the injected concentration because the pulse spreads as it is transported through the column. The condition for the conservation of mass of solute molecules in the migration zone (subscript z) is

$$m_i + m_p = m_{inj} \quad (7)$$

$$\langle C_i \rangle V_i + \langle C_p \rangle V_p = C_{inj}V_{inj} = m_{inj}$$

Here V_i and V_p are the mobile (inter-pore) and stationary (pore) portions of the zone volume V_z , with

$$V_p = (V_T - V_0) \frac{V_z}{V_T} ; V_i = (V_0) \frac{V_z}{V_T} \quad (8)$$

We measure the solute zone volume V_z from the full width at half maximum of the chromatogram peak using $V_z = t$, where t is the average flow rate. After substituting the definition of the partition coefficient given in Eq. 4 and definitions Eq. 8 into Eq. 7, one obtains:

$$\langle C_i \rangle = \frac{m_{inj}}{V_z \left(\frac{V_R}{V_T} \right)} \quad (9)$$

A simple way to understand Eq. 9 is to note that the numerator is the total mass in the zone and the denominator is the volume of the zone accessible to the protein. Thus, the concentration $\langle C_i \rangle$ is the ratio of these terms. In this derivation we have assumed Eq. 4 holds, which is no longer the case when both $B_2 \neq 0$ and the concentration is changing during transport down the column. However, as we will show below, the changes in K_D with concentration are small, which may justify our approximation. This relation allows us to extend the method of (Nichol et al., 1978), originally developed using frontal elution chromatography, to pulse HPLC. Alternatively, one could use the maximum concentration C_{max} of eluted solute instead of $\langle C_i \rangle$ in Eq. 6. As shown in Fig. 1 C_{max} and $\langle C_i \rangle$ are almost equal. Our procedure is then to inject different volumes of samples at various concentrations, measure K_D from the retention times as given in Eq. 4 and then plot $\ln K_D$ as a function of either $\langle C_i \rangle (1 - K_D)$ or $C_{max} (1 - K_D)$. The slope of that plot is then $2B_2 N_A / M_w$.

Experimental

Materials

We obtained lysozyme (6x crystallized hen egg white), from Seikagaku America. Our studies, along with others' (Muschol and Rosenberger, 1997), of the purity of lysozyme preparations from Sigma and Seikagaku showed the Seikagaku to be purer and it was used without further purification. We obtained bovine serum albumin (BSA), from Sigma, and it was used without further purification. All buffer components were obtained from Fisher Scientific. A Millipore Elix system purified

water for all the experiments. We prepared potassium phosphate buffers by mixing 50 mM solutions of K_2HPO_4 and KH_2PO_4 , at various NaCl concentrations to adjust the ionic strength, to reach the desired $pH = 6.2$ as measured by an Orion SA 520 pH meter. The $pH = 4.7$ of sodium acetate buffers was adjusted by adding concentrated acetic acid to solutions of sodium acetate and NaCl. Additionally all buffers were passed through 0.45 μm nylon filters, also obtained from Millipore, prior to use. Protein concentrations were measured using a Varian instruments Cary 50B biospectrophotometer at a wavelength of 278 nm. The extinction coefficient used for lysozyme was $\epsilon_{278nm} = 2.64 \text{ ml (mg cm)}^{-1}$, and $\epsilon_{278nm} = 0.667 \text{ ml (mg cm)}^{-1}$ for BSA.

Chromatography

An 1100 series liquid chromatography (HPLC) system from Agilent Technologies (Wilmington, DE) was used for all chromatographic measurements. Protein retention times were determined using an Agilent differential refractive index detector (RID) and an Agilent diode-array-detector (DAD) by absorbance at 278 nm. A TSK-G2000SW (30 cm \times 0.75 cm ID.) column from Tosoh Biosep and a YMC-Diol-200AMP (30 cm \times 0.60 cm ID.) column from YMC were used in the chromatographic measurements. We used a flow rate of 1 ml/min for all measurements. These columns contain a packing of porous silica beads whose surfaces have been hydrophilically modified. From the manufacturer's specifications the diameter of a single bead is about 5 μm for both columns. The average pore diameter is 125 Å for the TSK-G2000SW, and 200 Å for the YMC-Diol-200AMP. We determined the SEC calibration curve for these columns by using poly-ethylene-glycol (PEG) samples with molecular weights $200 < M_w < 10^5 \text{ g/mol}$, obtained from Sigma and Fluka. For every run the eluent was the same as the sample buffer. The random run-to-run difference in retention times for our system was $< 0.1\%$. Any dependence of the dimensionless distribution coefficient K_D for protein molecules between the stationary and mobile phases on the average flow rate would indicate non-equilibrium effects. We found K_D to be totally independent of flow rate for

the experimentally accessible values: 0.1 m l/m in 1.3 m l/m in.

Methods

For each solvent condition we performed a series of HPLC experiments varying solute (protein) injected concentration C_{inj} and using two injection volumes, $V_{inj} = 20$ and $100 \text{ } \mu\text{L}$. We identify the protein retention time t_r as the time of the maximum in the RID signal (Fig. 1), where the injection time is $t = 0$. We plot t_r as a function of C_{inj} , and find that t_r depends on V_{inj} as shown in Fig. 2. In order to apply our modification of (Nicholet al., 1978)'s method to HPLC, we recalculate the average solute concentration in the peak zone, $\langle C_i \rangle$, as described in Eq. 9, and find that this reassuringly collapses the multiple t_r vs. C_{inj} curves from Fig. 2 to a single curve as shown in the insert of Fig. 3. The slope of this collapsed curve is proportional to the second virial coefficient according to Eq. 5.

In order to calculate K_D according to Eq. 4 we must measure the total (t_T) and dead (t_0) times. We have measured the total time for each run using the solvent peak (these are maximums of the second peaks (t_T) in Fig. 1). In order to measure the dead time, we used PEG with a molecular weight of 10^5 g/mol , which is totally excluded from the TSK and the YMC columns. We have measured the dead times for all solvent conditions and injection volumes. It is important to measure t_T and t_0 separately for all injection volumes to avoid any instrumental errors associated with precisely identifying the injection time.

We have performed light scattering measurements to determine B_2 independently for a condition where results were not found in the literature. We employed the same method as in (George and Wilson, 1994) to measure the Rayleigh ratio of protein solutions using toluene as a standard at a scattering angle of 90 degrees. In Eq. 1 B_2 has the units of volume, but virial coefficients are often reported in units of m l mol/g^2 , which is denoted by A_2 (George and Wilson, 1994). Then $B_2 = A_2 M_w^2 / N_A$, where N_A is Avogadro's number. Our results are shown in Table 1.

Results

We have measured the dependence of the retention factor K_D on C_{inj} and V_{inj} for lysozyme and BSA in the above mentioned buffers and columns. These buffer conditions were chosen to investigate the cross-over from positive to negative B_2 values and to compare with data available in the literature.

Fig. 1 shows the RID signal measuring the concentration of the eluted protein versus time for representative lysozyme chromatograms with $V_{inj} = 20$ μ L. One can see the retention time increase with increasing protein concentration, while t_r remains constant.

In the size exclusion mode, the direction of the shift in the retention time with concentration depends on the sign of B_2 . For conditions where $B_2 > 0$, t_r increases with increasing protein concentration and where $B_2 < 0$, t_r decreases with increasing concentration. If $B_2 = 0$, t_r is independent of concentration. Previous studies (Velev et al., 1998, Muschol and Rosenberger, 1995, Gripon et al., 1997, Kulkarni, 1999) have shown that B_2 for protein solutions depends on the ionic strength of the solution.

Fig. 2 shows the dependence of lysozyme retention times on the injected concentration C_{inj} . The two sets of data correspond to different injection volumes (V_{inj}): 20 μ L and 100 μ L. Following the procedure introduced above for determining the average solute concentration in the mobile phase of the migration zone $\langle C_i \rangle$, we plot the dimensionless retention parameter, K_D , versus $\langle C_i \rangle$ in the insert of Fig. 3. This procedure collapses the data from Fig. 2 onto a single curve from which V_{inj} has been removed as an independent parameter. At the smallest concentrations in the insert of Fig. 3, some non-linear dependence of K_D on $\langle C_i \rangle$ can be observed. We attribute this behavior to errors introduced at the smallest signal to noise ratios. We have not included these points in our fits.

In order to extract B_2 from chromatographic data, one calculates $\langle C_i \rangle$ by Eq. 9 and then plots $\ln K_D$ versus $\langle C_i \rangle (1 - K_D)$. Following (Nichol et al., 1978), the slope of a linear fit to such a plot is then $2B_2 N_A / M_w$, as in Fig. 3.

The protein concentration range typically used to measure B_2 by light scat-

tering is approximately $0 < C_i < 30 \text{ mg/ml}$, (Velev et al., 1998, Muschol and Rosenberger, 1995). In our SEC measurements the protein concentrations $< C_i >$ eluting from the column, correspond to precisely the same range, although the injected concentrations are much higher as shown in Fig. 2. Even with these high concentrations, we never saturated our column. Such high injected concentrations may not be accessible for other protein systems, and may in fact be avoided by employing larger injection volumes, as shown by the $V_{inj} = 100 \text{ l}$ data in Fig. 2.

In Fig. 4 we compare our B_2 results for BSA from pulse SEC and those obtained by (Nichol et al., 1978) using frontal chromatography. Our results show the same slope for $\ln K_D$ as a function of $< C_i > (1 - K_D)$ as those obtained by (Shearwin and Winzor, 1990), which means the B_2 values are the same. The solution conditions for the two data sets differ, but other studies (George and Wilson, 1994, Moon et al., 2000) have shown that B_2 for BSA is insensitive to many changes in solution conditions until crystallizing conditions are approached. Therefore we expect to measure a similar value of B_2 . We measured different values of K_D than those in (Shearwin and Winzor, 1990) simply because we used a different column.

In order to further validate the extraction of B_2 from size exclusion chromatography (SEC), we compare our results to those obtained by light scattering in Fig. 5, and in Table 1. Fig. 5 shows the dependence of the second osmotic virial coefficient on solution ionic strength (added NaCl concentration) for lysozyme. Our data agrees quantitatively with those previously obtained over a wide range of ionic strengths. Table 1 compares A_2 values obtained in a different buffer, Potassium Phosphate 50mM pH 6.2. For this buffer our SEC measurements of A_2 also agree with those from light scattering in their sign. The differences in magnitude can be attributed to systematic errors associated with light scattering and SEC measurements of A_2 , not statistical variation. Note that previously published results for A_2 from various groups, as shown in Fig. 5, differ by as much or more than the values shown in Table 1. These results illustrate the ability of SEC to track the evolution of protein interactions from net repulsive $A_2 > 0$ to attractive $A_2 < 0$.

Conclusion

We have adapted the idea of (Nichol et al., 1978) and present measurements of protein second virial coefficients using the standard practice of size exclusion liquid chromatography, thereby reducing the cost in time and material of performing B_2 measurements for protein solutions. After calculating the protein concentration in the solute zone, our results agree with those previously obtained using an independent method, light scattering, in a number of other studies.

Acknowledgement

Research was funded by NASA Office of Biological & Physical Research, Fundamental Microgravity Research in Physical Sciences (Fluids Physics) Grant # NAG 3-2386.

NaCl (mM)	A_2 (10^{-4} ml mol/g ²)	
	From SEC	From LS
0	2.4	1.8
50	1.6	–
150	–1.0	–1.4

Table 1: Comparison of size exclusion chromatography (SEC) and light scattering (LS) measurements of the second virial coefficients (10^{-4} ml mol/g²) for lysozyme in Potassium Phosphate Buffer 50 mM, pH = 6.2, at various added NaCl concentrations.




Figure 1: Lysozyme chromatograms for $V_{inj} = 20$ l and different injected concentrations (C_{inj}) as indicated next to each curve. The average concentrations $\langle C_i \rangle$ used in the analysis in Fig. 3 are dash – 0.74 mg/ml, points – 1.72 mg/ml, dash and points – 2.20 mg/ml, long dash – 2.68 mg/ml, long dash and points – 3.08 mg/ml. The vertical line marks the retention time for the most dilute sample (not shown). The retention time t_r is the time corresponding to the peak of the concentration profile (C_{max}) and increases with increasing concentration. The retention time of completely included molecules (the "total" volume) is marked as t_T and is caused by the buffer. The retention time of completely excluded molecules (the "dead" volume) was $t_0 = 6.07$ min (not shown in figure). Note that C_{max} and $\langle C_i \rangle$ are similar. The buffer is Sodium Acetate, 50mM, pH 4.7.




Figure 2: Lysozyme retention times versus injected concentrations for two injection volumes. Set a : $V_{inj} = 20$ l and set b : $V_{inj} = 100$ l. Buffer : Sodium Acetate 50 mM pH 4.7.




Figure 3: $\ln K_D$ vs. $\langle C_i \rangle (1 - K_D)$ (mg/ml) for lysozyme as in Eq. 5. The insert is a plot of K_D vs. $\langle C_i \rangle$ (mg/ml), where multiple curves from Fig. 2 with different injected volumes collapse after recalculating the solute concentration in the mobile phase of the migration zone as in Eq. 9. Buffer: Sodium Acetate 50 mM pH 4.7, black dots: $V_{inj} = 100$ l, open circles: $V_{inj} = 20$ l.




Figure 4: Comparison of size exclusion chromatography (SEC) and frontal chromatography measurements for BSA. Open diamonds: BSA in Sodium Acetate 20mM, NaCl 0.18M, pH=4.6, $A_2 = 1.9 \cdot 10^{-4} \text{ ml mol/g}^2$ (Shearwin and Winzor, 1990). Black points are for BSA in Potassium Phosphate 50mM, pH=6.2, $A_2 = 2.0 \cdot 10^{-4} \text{ ml mol/g}^2$. Injected concentrations are 1.14, 4.85, 10.05, 15.0, 20.7, 25.27, 30.44, 40.72, 50.99 mg/ml. Injection volumes are for squares: 2 μl ; stars: 10 μl ; triangles: 40 μl ; diamonds: 100 μl .




Figure 5: Comparison of size exclusion chromatography (SEC) and light scattering measurements for lysozyme. A_2 (10^{-4} ml mol/g²) versus NaCl concentration. Buffer: Sodium Acetate 50 mM pH 4.7. The SEC measurements are denoted by open pentagons. The data for 0, 100, 200, and 300 mM NaCl were taken on a TSK column, and the data for 400 mM NaCl was with a YMC column. The remaining data comes from published light scattering data. Black diamonds: (Gripson et al., 1997), black triangles: (Velez et al., 1998), black stars: (Kulkarni, 1999), black rectangles: (Muscholet al., 1995)

References

- [1] Behlke, J., and O. Ristau. 1999. Analysis of the thermodynamic non-ideality of proteins by sedimentation equilibrium experiments. *Biophys. Chem.*, 76:13{23
- [2] Bonnet, F., S. Finet, and A. Tardieu. 1999. Second virial coefficient: variations with lysozyme crystallization conditions. *J. Crystal Growth*, 196:403{414
- [3] George, A., and W.W. Wilson. 1994. Predicting protein crystallization from a dilute solution property. *Acta. Cryst., D* 50:361{365
- [4] Gripon, C., L. Legrand, I. Rosenman, O. Vidal, M.C. Robert, and F. Bou. 1997. Lysozyme-lysozyme interactions in under- and super-saturated solutions: a simple relation between the second virial coefficients in H_2O and D_2O . *J. Cryst. Gr.*, 178:575{584
- [5] Hill, T., 1960. An introduction to statistical thermodynamics. Addison-Wesley, MA
- [6] Kulkarni, A.M. 1999. Depletion interactions in the protein limit. Master's thesis, U. Illinois
- [7] Nichol, L.W., R.J. Siezen, and D.J. Winzor. 1978. The study of multiple polymerization equilibria by glass bead exclusion chromatography with allowance for thermodynamic non-ideality effects. *Biophys. Chem.*, 9:47{55
- [8] Moon, Y.J., R.A. Curtis, C.O. Anderson, H.W. Blanch, and J.M. Prausnitz. 2000. Protein-protein interactions in aqueous Ammonium Sulfate solutions. Lysozyme and Bovine Serum Albumin (BSA). *J. Sol. Chem.*, 29(8):699{717
- [9] Muschol, M., and F. Rosenberger. 1997. Liquid-liquid phase separation in supersaturated lysozyme solutions and associated precipitate formation/crystallization. *J. Chem. Phys.*, 107(6):1953{1962
- [10] Muschol, M., and F. Rosenberger. 1995. Interactions in undersaturated and supersaturated lysozyme solutions: Static and dynamic light scattering results. *J. Chem. Phys.*, 103(24):10424{10432

- [1] Shearwin, K. E., and D. J. Winzor. 1990. Thermodynamic nonideality in macromolecular solutions. *Eur. J. Biochem.*, 190:523{529
- [2] Tessier, P. M., A. M. Lenhoff, and S. I. Sandler. 2002. Rapid measurement of protein osmotic second virial coefficients by self-interaction chromatography. *Biophys. J.*, 82:1620{1631
- [3] Velev, O. D., E. W. Kaler, and A. M. Lenhoff. 1998. Protein interactions in solution characterized by light and neutron scattering: comparison of lysozyme and chymotrypsinogen. *Biophys. J.*, 75:2682{2697
- [4] Wills, P. R., L. W. Nichol, and R. J. Siezen. 1980. The indefinite self-association of lysozyme: consideration of composition-dependent activity coefficients. *Biophys. Chem.*, 11:71{82
- [5] Yau, W. W., J. J. Kirkland, and D. D. Bly. 1979. Modern size-exclusion liquid chromatography. *J. Wiley*, N. Y.