

This work has been submitted to the IEEE for possible publication. Copyright may be transferred without notice, after which this version may no longer be accessible.

Personal Dosimetry in Direct Pulsed Photon Fields with the Dosepix Detector

Dennis Haag^{1,*}, Sebastian Schmidt¹, Patrick Hufschmidt¹, Gisela Anton¹, Rafael Ballabriga³, Rolf Behrens², Michael Campbell³, Franziska Eberle¹, Christian Fuhg², Oliver Hupe², Xavier Llopart³, Jürgen Roth², Lukas Tlustos³, Winnie Wong⁴, Hayo Zutz², and Thilo Michel¹

¹ Erlangen Centre for Astroparticle Physics, Friedrich-Alexander Universität Erlangen-Nürnberg, 91058 Erlangen, Germany

² Physikalisch-Technische Bundesanstalt (PTB), 38116 Braunschweig, Germany

³ CERN, 1211 Geneva, Switzerland

⁴ was with CERN, 1211 Geneva, Switzerland. She is now with Mercury Systems, 1212 Geneva, Switzerland

* E-mail: dennis.den.haag@fau.de

Abstract

First investigations regarding dosimetric properties of the hybrid, pixelated, photon-counting Dosepix detector in the direct beam of a pulsed photon field (RQR8) for the personal dose equivalent $H_p(10)$ are presented. The influence quantities such as pulse duration and dose rate were varied, and their responses were compared to the legal limits provided in PTB-A 23.2. The variation of pulse duration at a nearly constant dose rate of about 3.7 Sv/h shows a flat response around 1.0 from 3.6 s down to 2 ms. A response close to 1.0 is achieved for dose rates from 0.07 Sv/h to 35 Sv/h for both pixel sizes. Above this dose rate, the large pixels (220 μm edge length) are below the lower limit. The small pixels (55 μm edge length) stay within limits up to 704 Sv/h. The count rate linearity is compared to previous results, confirming the saturating count rate for high dose rates.

Index Terms— Active personal dosimetry, hybrid pixel detector, pulsed photon fields, Dosepix

1 Introduction

Dosimetry in pulsed photon fields with active electronic personal dosimeters (APDs) is an important topic of the last decade. The advantages of APDs compared to their passive counterparts are direct readability of the dose and an active warning of the wearer if pre-defined dose/dose rate thresholds are surpassed. An active warning is, for example, an alarm sound, a flashing light, or vibration of the dosimeter. They can help increase the wearers' awareness of unintentional high exposures and allow them to react to the situation directly. A pulsed radiation field is defined as ionizing radiation with a constant dose rate for pulse durations shorter than 10 s [1]. With this definition, most X-ray tubes in medical applications are classified as pulsed radiation emitters [2]. The staff of interventional radiology, interventional cardiology (IR/IC), or veterinary medicine is not exposed to the direct radiation field during X-ray examinations. They stand next to the patient, i.e., in the scattered radiation field where dose rates of 0.1 to 3 % of the dose rate in the direct radiation field can be expected [3]. Nonetheless, it is essential to investigate the performance of APDs for dose rates of several hundreds of Sv/h. These dose rates occur in the case of situations in which the wearer of the dosimeter accidentally gets exposed by the direct beam of a pulsed radiation field. For example, the relevant radiation field characteristics in IR/IC are a pulse duration of 1-20 ms, a dose rate in the direct radiation field of 2-360 Sv/h, and of 5 mSv/h to 10 Sv/h in the scattered radiation field [4]. Latter dose rate is stated for a dosimeter worn above the lead apron of the operator. The energy range of the spectra in the scattered beam of the radiation field for interventional procedures is between 20-100 keV for peak high voltages of up to 120 kV [4]. In this work, Dosepix [5] is utilized in a system consisting of three Dosepix detectors to investigate its dose rate and pulse duration dependence for the personal dose equivalent $H_p(10)$ in the direct beam of a pulsed photon field. It is tested whether a dosimeter demonstrator consisting of three Dosepix detectors fulfills the requirements in the direct beam of a pulsed photon field, in particular, for high peak pulse dose rates and short radiation pulse durations. When speaking of dose rate and pulse duration in the following, the dose rate in the pulse and the pulse width of a single pulse are meant.

2 State of the art APDs

A survey regarding the use of active personal dosimeters in hospitals showed that the predominantly used detector type are silicon diodes in IR/IC and Geiger-Muller tubes in nuclear medicine [6]. Two of the most commonly used silicon diode APDs were tested in an RQR8 reference radiation field [7]. It was found that they perform well for dose rates of a few Sv/h but become increasingly insensitive for higher dose rates. Those types of dosimeters need correction factors for dead-time loss. These correction factors become inefficient for pulse durations below 100 ms [7]. Further tests in pulsed photon fields were performed in [8] investigating dosimeters from the questionnaire in [6]. The tests showed that only 3 of 10 APDs fulfilled the requirement of a dose rate above 1 Sv/h within the margin of $\pm 20\%$ for the dose indication relative to the reference dose for continuous radiation. The maximum achieved dose rate within limits was 4 Sv/h in an RQR8 reference photon field. Another study compares APDs to their passive counterparts in pulsed photon fields and hospitals [9]. APDs (silicon diodes) show an under-response due to pulsed radiation compared to passive dosimeters, which are not influenced by pulsed radiation. The tested dosimeters perform well in the scattered radiation field in the hospital but show a response below 20 % of the reference dose in the direct beam of the radiation field (dose rate of 74 Sv/h in the pulse). The same dosimeters show a response below 70 % relative to a continuous field in the direct beam of an RQR8 reference field at 5 Sv/h and 10 Sv/h. One of the three dosimeters shows a 50 % response at already 1 Sv/h. All in all, the main issues of APDs in pulsed photon fields are high peak pulse dose rates, as they occur in the direct beam of the radiation field, and short radiation pulse durations, where dead-time correction factors become inefficient. Worst case scenario is that the detector is exposed to high dose rates in the direct beam of a pulsed radiation field but does not respond at all. It gives the wearer a false sense of security, particularly that no noticeable radiation is present. A comprehensive overview of radiation dosimeter types is given in [10]. Alternative active personal dosimeter systems in IR consist of CMOS image sensors as presented in [11, 12]. In this work, the detector type of interest are hybrid photon-counting pixel detectors. Their principle of dosimetry is elaborated in [13].

3 The Dosepix detector

Dosepix is a hybrid, pixelated, photon-counting X-ray detector. The hybrid design consists of an application-specific integrated circuit (ASIC) and a semiconductor sensor layer connected pixelwise to the ASIC. A fully depleted 300 μm thick silicon sensor is used with an applied bias voltage of 100 V. The pixel layout comprises 16×16 square pixels with 220 μm pixel pitch with a p-in-n doping profile. The upper two and lower two rows of the sensor pixel matrix consist of small pixels with an edge length of 55 μm , while the remaining 12 rows consist of larger pixels with an edge length of 220 μm . The smaller pixels detect fewer events than the larger pixels and therefore have a lower tendency for pile-up, which allows applications at high-flux conditions. The Dosepix can be operated in 3 different programmable modes: the photon-counting mode, the integration mode, and the energy-binning mode. Here, the Dosepix detector is used in the latter one, which is specifically designed for dosimetry applications. The energy bin edges are individually programmed for each pixel. The Dosepix operates dead-time-free using the rolling-shutter principle. A single column is read out at a time while the rest of the matrix continues to process signals. This is a significant advantage in practical applications in pulsed photon fields, where radiation pulses are random in time. A single column consists of 12 large and 4 small pixels. The energy-binning mode pixelwise counts events in one of 16 histogram bins according to the deposited energy of the event. The total number of registered events is determined by summing the number of events per bin for each pixel type individually, resulting in two total energy histograms. Each energy histogram is multiplied by a dose equivalent and pixel size specific set of conversion factors to yield the personal dose equivalent H_{DPX}^x . The dose for a single detector is calculated via [13]

$$H_{\text{DPX}}^x = \sum_{i=1}^{16} k_i^x N_i^x \quad (1)$$

where N_i^x denotes the number of events in energy bin i for the pixel type x , and k_i^x its corresponding conversion factor. The total dose is the sum of the dose contributions of the three detectors. Two separate dose values are calculated, one for the small and one for the large pixels. The statistical uncertainty of H_{DPX}^x is calculated following [14] via a propagation of uncertainty assuming the uncertainty of N_i^x as $u(N_i^x) = \sqrt{N_i^x}$:

$$u(H_{\text{DPX}}^x) = \sqrt{\sum_{i=1}^{16} (k_i^x \sqrt{N_i^x})^2}. \quad (2)$$

Information regarding the characterizations of Dosepix with X-rays and analog test-pulses can be found in [15], and measurements of the count rate linearity in dependence of the dose rate can be found in [16]. A dosimetry system consisting of 3 Dosepix detectors is utilized as described in [17], where the energy and angular dependence in continuous photon fields were already presented.

4 Methods

The tests in pulsed photon fields were performed in collaboration with and at the PTB (Physikalisch-Technische Bundesanstalt Braunschweig) using its X-ray unit for pulsed radiation GESA (GEpulste Strahlungs Anlage) presented in [18]. The chosen reference radiation field is the medical radiation quality RQR8 with a tube voltage of 100 kV filtered with 3.36 mm aluminum and a mean energy (fluence) of 51 keV [19]. Each measurement was repeated 2 or 3 times for statistical purposes. According to ISO standards [20], the dosimeter was irradiated on the center front of an ISO water slab phantom. The phantom is necessary because the operational quantity for individual monitoring of strongly penetrating radiation, $H_p(10)$, is measured [21]. It is defined at a depth of 10 mm in the body, and is used for whole body dosimeters. The phantom serves as substitute of the human body in the radiation field and considers its backscattering. However, to achieve very high dose rates up to 1080 Sv/h, the dosimeter had to be irradiated relatively close to the X-ray tube resulting in small field diameters down to 8.5 cm. In these cases the dosimetry system was irradiated without the ISO water slab phantom which has a 30 cm × 30 cm cross-section and a thickness of 15 cm. Correction factors were determined for measurements with and without the phantom and measurements with a small (8.5 cm) and a large (42.0 cm) field diameter. The correction factors are stated in Table 1 and were used to correct all measurements to the equivalent of the dosimetry system being placed on the ISO water slab phantom that is completely irradiated to guarantee $H_p(10)$ conditions. The quantity of interest is the change of the response R relative to the response at reference conditions R_0 , i.e., the normalized response. The response is defined by the ratio of the indication (calculated dose with Dosepix) and the reference dose determined by monitor ionization chambers which are practically independent of the dose rate and pulse duration. The change of the normalized response has to fulfill the following condition according to [22]

$$1 + f_{\min} \leq R_{\text{Norm}}^i = \frac{R_i}{R_0} = \frac{H_{\text{DPX}}^i H_{\text{ref}}^0}{H_{\text{ref}}^i H_{\text{DPX}}^0} \leq 1 + f_{\max} \quad (3)$$

with R_i being the response at measurement i , H_{DPX} the dose measured by the Dosepix dosimetry system, H_{ref} the reference dose, and f_{\min} and f_{\max} depending on the influence quantity (see Table 2). The statistical uncertainty of the normalized response is calculated via:

$$u(R_{\text{Norm}}^i) = R_{\text{Norm}}^i \sqrt{\left(\frac{u(H_{\text{DPX}}^i)}{H_{\text{DPX}}^i}\right)^2 + \left(\frac{u(H_{\text{DPX}}^0)}{H_{\text{DPX}}^0}\right)^2}. \quad (4)$$

The minimum requirements for the dose rate and pulse duration for conformity assessment are shown in Table 2. The limits of the statistical uncertainty of the measured dose for $H_p(10)$ dosimeters regarding X-radiation are between 5-15% depending on the applied reference dose [23]. Test conditions with respect to pulsed radiation are stated in [24].

5 Results and Discussion

5.1 Dependence on the pulse duration

The pulse duration was varied between 2 ms and 3.6 s, while the dose rate was held nearly constant with an average of about 3.7 Sv/h. The usual pulse duration is within the interval of several milliseconds in medical X-ray facilities. Due to the dead-time-free measurement of the Dosepix, no dependency on pulse duration is expected. Figure 1 shows the normalized response for both pixel sizes. The response value at 3.6 s was chosen as reference point R_0 . The normalized response is flat within the margin of the uncertainty. The uncertainty increases for lower pulse durations due to fewer registered events in the energy bins. The number of registered events decreases because the reference dose is reduced with the pulse duration to hold the dose rate nearly constant for all irradiations. The reference dose values range from about 2.3 μSv at 2 ms to about 3.5 mSv at 3.6 s. For 2.3 μSv an uncertainty limit of 15% is tolerated [23]. Here, the

Table 1: Correction factors for the phantom influence and the field-diameter influence

Pixel size [μm]	Presence of Phantom correction	Field diameter correction
55	1.148±0.004	1.033±0.007
220	1.168±0.002	1.024±0.003




Figure 1: Normalized response for $H_p(10)$ in an RQR8 photon reference field. Influence quantity is the pulse duration at nearly constant dose rate. The uncertainty bars are calculated via (4).

large and small pixels show their highest relative statistical uncertainty of 1.0% and 5.5%, respectively. Overall, all data points are within limits. The data points below 10 ms correspond to a potential use case in IR/IC. No issues with the Dosepix are expected for these pulse durations.

5.2 Dependence on the dose rate

The Dosepix dosimetry system was irradiated in the direct radiation field for dose rates in the range of 0.07 Sv/h to 1080 Sv/h. To achieve these dose rates, both the reference dose and the pulse duration were varied. The response for both the small and large pixels was evaluated and is shown in Figure 2. Both pixel sizes have a nearly flat response up to about 35 Sv/h. The normalized response of the large pixels falls below the lower limit at slightly less than 100 Sv/h, whereas for the small pixels, a dose rate up to about 704 Sv/h is achievable in the used reference field. The explanation for the decreasing normalized response with increasing dose rate is pile-up and its corresponding decrease of the count rate. This results in a decreased dose and subsequently a decreased response. The small pixels show a small increase of the normalized response prior to the decrease. The small pixels are less prone to pile up due to their smaller detection volume. An overestimation is observed in the case of several low-energy photons being registered simultaneously, which yields an event with a higher energy. Such an event is sorted into an energy bin a higher conversion factor. Such an overestimation is not observed in this radiation field for the large pixels due a stronger influence of pile up, which results in an overflow of the last bin, and sets an overflow flag. If the flag is set, the bin is not utilized in the dose calculation because it would lead to a significant overestimation of the dose. The normalized response is constant for the dose rates of scattered radiation fields in IR/IC. The spectrum of the scattered radiation

Table 2: Minimum requirements for conformity assessment according to PTB-A 23.2 [22] for $H_p(10)$ dosimeters

Quantity	Minimum rated range of use	Reference value	f_{\min}, f_{\max}
Dose rate	0.1 μ Sv/h to 1 Sv/h	1 mSv/h	-0.13, 0.18
Radiation pulse duration	1 ms to 10 s	Response at continuous radiation	-0.2, 0.2




Figure 2: Normalized response for $H_p(10)$ in an RQR8 photon reference field. Influence quantity is the dose rate. The uncertainty bars are calculated via (4).

field is softer than the spectrum of the primary radiation field. The effectively smaller energies do not pose a problem to the response of Dosepix, which operates above 12 keV. The small pixels would allow an active warning in accident situations, e.g., if the person is exposed to the direct X-ray radiation field of a medical diagnostic X-ray tube where dose rates up to 360 Sv/h can occur [4]. The maximum relative statistical uncertainties of the large and small pixels are 0.24% and 0.49%, respectively. The applied reference dose values are 1 mSv and higher, and the corresponding uncertainty limits between 5-6% [23]. The observed uncertainties stay within limits.

5.3 Count rate linearity

The measurements with the RQR8 spectrum when varying the dose rate are compared to previous measurements performed in [16]. For comparability, the abscissa is first re-scaled to Sv/s and then divided by the $H_p(10)/K_{\text{air}}$ conversion coefficient for the RQR8 spectrum (1.438 Sv/Gy). The results are shown in Figure 3 for each pixel size of the three detectors and are additionally labeled by their filter cap, i.e., aluminum cylinder with a thin aluminum foil on its top, which has a hole above the sensor (free), aluminum half-sphere (Al), and tin half-sphere (Sn). Similar behavior is observed as presented by Zang et al. [16], which means that the count rate saturates with high dose rates for unfiltered or weakly filtered detectors. The Dosepix filtered with tin shows for both pixel types no saturation and overall a low count rate. The explanation for the saturation is stated by Zang et al. in arguing that analog pile-up is increasing with an increasing dose rate which is equal to an increase of the flux. Therefore, an increase in the dose rate flattens the deposition spectrum in the detector, namely by converting several low-energy photons into a single high-energy event. For dosimetry, this implies that the dose determination is impacted. A correlation between the normalized response and count rate is observed. From the turning point at 35 Sv/h onward, the response falls below the limit. The compensation of the larger values of the conversion factor in higher energy bins is not strong enough to counteract the loss of separate events. Even due to the count rate saturation and analog pile-up, a good normalized response is achieved. The reason for this is that one of the three detectors - namely the Dosepix filtered with the tin cap - is only in the beginning stage of its saturation. The latter statement implies that its 16 energy bins appropriately represent the energy deposition spectrum and that its partial dose is still correctly determined.

6 Conclusion

The Dosepix detector's dependence of the normalized response in an RQR8 pulsed photon field (with a mean energy of 51 keV) was shown for the variation of the pulse duration and the variation of the dose rate. Both tests show promising results for applying the Dosepix detector as an active personal dosimeter measuring the personal dose equivalent $H_p(10)$ in pulsed photon fields. The pulse duration independence of the dose measured by the prototype of a Dosepix dosimetry system is a direct result of the dead-time-free readout principle of the Dosepix acting as a camera-like radiation detector. Even the shortest pulse durations will not pose any problems to the Dosepix detector, provided that the dose rate during the pulse does not exceed certain limits. As demonstrated here, these limits concerning dose rate are substantial - i.e., in the order of 100 Sv/h and higher - compared to other commercial electronic dosimeters that saturate in the region of a few Sievert per hour. Therefore, it is concluded that the Dosepix detector is a viable




Figure 3: Count rate linearity in an RQR8 photon reference field. Influence quantity is the dose rate. The individual count rates of the three differently filtered detectors and their different pixel sizes are shown.

detector for dosimetry of pulsed photon fields. The pulsed fields occurring in IR/IC with their parameter space of several milliseconds of pulse duration and a dose rate of up to 10 Sv/h above the lead apron of the operator do not pose any problem to Dosepix. Even for dose rates in the direct beam of the radiation field, a correct indication of the response within limits is expected up to at least 35 Sv/h, and approximately to 70 Sv/h with the large pixels when interpolating between the data points. The small pixels allow a correct indication of the dose up to a factor 10 of the dose rate obtained from large pixels. The large pixels are intended for use in dosimetry, while the small pixels act as a control mechanism. If the dose deviation between large and small pixels exceeds a threshold value, it can be assumed that the dose rate is too high, which would switch the dose measurement to the small pixels until the deviations are below the threshold. The large pixels allow a correct indication of the dose of at least one order of magnitude higher and the small pixels at least two orders of magnitude higher than dosimeters tested in [7–9]. It has to be pointed out, that the demonstrated large dose rate range is guaranteed only for the used radiation quality RQR8. In the relevant X-ray energy regime, the detection efficiency is larger for lower energies. On the other hand, the signal pulse length decreases towards lower energies and so does the pile-up probability. The resulting energy dependence of the dose rate limit of our system is therefore not trivial and depends on the spectral shape. Further tests in different energy ranges need to be performed to identify the largest possible dose rate within the legal limits for the normalized response. An application in radiotherapy, e.g., in the scattered radiation field of medical Linac, where photon energies of several hundreds of kilo electron volts up to mega electron volts are expected needs to be investigated separately. If the dose rates in the scattered beam are within the ranges investigated in this publication, then an application of Dosepix in Linac is feasible.

Acknowledgements

This project is funded by the Deutsche Forschungsgemeinschaft (DFG, German Research Foundation) - 394324524.

References

- [1] International Organization for Standardization (ISO), *Radiological protection - Characteristics of reference pulsed radiation - Part 1: Photon radiation*. ISO/TS 18090-1:2015.
- [2] U. Ankerhold, O. Hupe, and P. Ambrosi, “Deficiencies of active electronic radiation protection dosimeters in pulsed fields,” *Radiation Protection Dosimetry*, vol. 135, no. 3, pp. 149–153, Jul. 2009, doi:10.1093/rpd/ncp099.
- [3] O. Hupe and U. Ankerhold, “Determination of the dose to persons assisting when x-radiation is used in medicine, dentistry and veterinary medicine,” *Radiation Protection Dosimetry*, vol. 144, no. 1-4, pp. 478–481, Nov. 2010, doi:10.1093/rpd/ncq351.
- [4] I. Clairand, J.-M. Bordy, J. Daures, J. Debroas, M. Denoziere, L. Donadille, M. Ginjaume, C. Itie, C. Koukorava, S. Krim, A.-L. Lebacqz, P. Martin, L. Struelens, M. Sans-Merce, M. Tosic, and F. Vanhavere, “Active personal

- dosemeters in interventional radiology: tests in laboratory conditions and in hospitals,” *Radiation Protection Dosimetry*, vol. 144, no. 1-4, pp. 453–458, Mar. 2011, doi:10.1093/rpd/ncq556.
- [5] W. Wong, G. Anton, R. Ballabriga, M. Böhnel, M. Campbell, E. Heijne, X. Llopart, T. Michel, I. Münster, R. Plackett, P. Sievers, P. Takoukam, L. Tlustos, and P. Valerio, “A pixel detector ASIC for dosimetry using time-over-threshold energy measurements,” *Radiation Measurements*, vol. 46, no. 12, pp. 1619–1623, Dec. 2011, doi:10.1016/j.radmeas.2011.06.061.
- [6] O. Ciraj-Bjelac, E. Carinou, and F. Vanhavere, “Use of active personal dosimeters in hospitals: EURADOS survey,” *Journal of Radiological Protection*, vol. 38, no. 2, pp. 702–715, May 2018, doi:10.1088/1361-6498/aabce1.
- [7] O. Hupe, H. Zutz, and J. Klammer, “Radiation protection dosimetry in pulsed radiation fields,” presented at IRPA 13 Glasgow. [Online]. Available: "https://www.irpa.net/members/TS2f.3.pdf".
- [8] O. Hupe, S. Friedrich, F. Vanhavere, and M. Brodecki, “Determining the dose rate dependence of different active personal dosimeters in standardized pulsed and continuous radiation fields,” *Radiation Protection Dosimetry*, vol. 187, no. 3, pp. 345–352, Sep. 2019, doi:10.1093/rpd/ncz173.
- [9] F. Vanhavere, E. Carinou, I. Clairand, O. Ciraj-Bjelac, F. De Monte, J. Domienik-Andrzejewska, P. Ferrari, M. Ginjaume, H. Hršak, O. Hupe, Z. Knezevic, U. O’Connor, M. S. Merce, S. Sarmiento, A. Savary, and T. Siskoonen, “The use of active personal dosimeters in interventional workplaces in hospitals: comparison between active and passive dosimeters worn simultaneously by medical staff,” *Radiation Protection Dosimetry*, vol. 188, no. 1, pp. 22–29, Dec. 2019, doi:10.1093/rpd/ncz253.
- [10] F. Ravotti, “Dosimetry techniques and radiation test facilities for total ionizing dose testing,” *IEEE Trans. Nucl. Sci.*, vol. 65, no. 8, pp. 1440–1464, Aug. 2018, doi:10.1109/TNS.2018.2829864.
- [11] E. Conti, P. Placidi, M. Biasini, L. Bissi, A. Calandra, B. Checcucci, S. Chiochini, R. Cicioni, R. Di Lorenzo, A. C. Dipilato, A. Esposito, M. Paolucci, D. Passeri, A. Pentiricci, A. Scorzoni, and L. Servoli, “Use of a CMOS image sensor for an active personal dosimeter in interventional radiology,” *IEEE Trans. Nucl. Sci.*, vol. 62, no. May, pp. 1065–1072, 2013, doi:10.1109/TIM.2012.2223331.
- [12] A. Omar, N. Kadesjö, C. Palmgren, M. Marteinsdottir, T. Segerdahl, and A. Fransson, “Assessment of the occupational eye lens dose for clinical staff in interventional radiology, cardiology and neuroradiology,” *Journal of Radiological Protection*, vol. 37, no. 1, pp. 145–159, Jan. 2017, doi:10.1088/1361-6498/aa559c.
- [13] T. Michel, M. Böhnel, J. Durst, P. Sievers, and G. Anton, “Low energy dosimetry with photon counting pixel detectors such as medipix,” *IEEE Trans. Nucl. Sci.*, vol. 56, no. 2, pp. 417–423, Apr. 2009, doi:10.1109/TNS.2009.2015229.
- [14] Joint Committee for Guides in Metrology (JCGM), *Evaluation of Measurement Data - Guide to the Expression of Uncertainty in Measurement*, 2008.
- [15] I. Ritter, G. Anton, R. B. Sune, F. Bisello, M. Campbell, T. Gabor, X. L. Cudie, S. Wölfel, W. S. Wong, and T. Michel, “Characterization of the dosepix detector with xrf and analog testpulses,” *Journal of Instrumentation*, vol. 9, no. 05, p. C05069, May 2014, doi:10.1088/1748-0221/9/05/C05069.
- [16] A. Zang, G. Anton, R. Ballabriga, F. Bisello, M. Campbell, J. Celi, A. Fauler, M. Fiederle, M. Jensch, N. Kochanski, X. Llopart, N. Michel, U. Mollenhauer, I. Ritter, F. Tennert, S. Wölfel, W. Wong, and T. Michel, “The dosepix detector—an energy-resolving photon-counting pixel detector for spectrometric measurements,” *Journal of Instrumentation*, vol. 10, no. 04, p. C04015, Apr. 2015, doi:10.1088/1748-0221/10/04/C04015.
- [17] D. Haag, S. Schmidt, P. Hufschmidt, F. Eberle, T. Michel, G. Anton, O. Hupe, J. Roth, C. Fuhg, H. Zutz, R. Behrens, M. Campbell, X. Llopart, R. Ballabriga, L. Tlustos, and W. Wong, “Personal dosimetry in continuous photon radiation fields with the dosepix detector,” *IEEE Trans. Nucl. Sci.*, vol. 68, no. 5, pp. 1129–1134, May 2021, doi:10.1109/TNS.2021.3068832.
- [18] J. Klammer, J. Roth, and O. Hupe, “Novel reference radiation fields for pulsed photon radiation installed at PTB,” *Radiation Protection Dosimetry*, vol. 151, no. 3, pp. 478–482, 04 2012, doi:10.1093/rpd/ncs043.
- [19] International Electrotechnical Commission (IEC), *Medical diagnostic X-ray equipment - Radiation conditions for use in the determination of characteristics*. IEC 61267:2005.
- [20] International Organization for Standardization (ISO), *Radiological protection - X and Gamma Reference Radiation for Calibrating Dosimeters and Doserate Meters and for Determining their Response as a Function of Photon Energy — Part 3: Calibration of Area and Personal Dosimeters and the Measurement of their Response as a Function of Energy and Angle of Incidence*. ISO 4037-3:2019.
- [21] W. Alberts, P. Ambrosi, J. Böhm, G. Dietze, K. Hohlfeld, and W. Will, “New dose quantities in radiation protection,” *Physikalisch-Technische Bundesanstalt (PTB)*, Aug. 1995, doi:10.7795/110.20200417EN.

- [22] Physikalisch-Technische Bundesanstalt (PTB), “PTB-Anforderungen PTB-A 23.2: Strahlenschutzmessgeräte; Personendosimeter zur Messung der Tiefen- und Oberflächen-Personendosis. Ausgabe November 2013,” Nov. 2013, doi:10.7795/510.20151109M.
- [23] Physikalisch-Technische Bundesanstalt (PTB), “Ergänzung der PTB-Anforderungen PTB-A 23.2: Strahlenschutzmessgeräte; Personendosimeter zur Messung der Tiefen- und Oberflächen-Personendosis,” Physikalisch-Technische Bundesanstalt, Braunschweig und Berlin, Nov. 2018, doi:10.7795/510.20181129A.
- [24] International Electrotechnical Commission (IEC), *Radiation protection instrumentation - Dosimeters for pulsed fields of ionizing radiation*. IEC TS 63050:2019.